A-A-51315
27 December 1985

COMMERCIAL ITEM DESCRIPTION

MANIKIN, CARDIOPULMONARY RESUSCITATION TRAINING (Head and Torso)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a manikin head and torso for use in cardiopulmonary resuscitation (CPR) training.

Salient characteristics:

The manikin shall be an adult size head and torso for teaching cardiopulmonary resuscitation (CPR). It shall be capable of accurate simulation of both Expired Air Resuscitation (EAR), and External Cardiac Compression (ECC).

The manikin shall include a moveable head section connected to a thorax section. The xiphoid shall be palpable beneath the skin covering the torso section. The neck and jaw shall be moveable. A collapsible lung bag shall be included, covered via an elastic chest cover that simulates an average adult lung compliance of 45 ml/cm H_2O. A stomach bag to simulate gastric distention which inflates at a pressure of 30 cm H_2O shall be included. A removeable thorax section shall be included which automatically and physiologically simulates the characteristics of an adult chest. The thorax section shall contain a xiphoid clicker which sounds if the xiphoid is inadvertently depressed. A pulse variant bulb to simulate the cardiac pulse shall be included. A reuseable mouth/nose piece and disposable head bags shall be included.

The manikin shall include at least the following:

a. Ten individual mouth/nose pieces
b. One hundred head bags.
c. One each lung, stomach, thorax, abdomen, xiphoid and pulse variator.

The manikin shall include service data. The contractor shall enclose, with each unit, two copies each, of one or more service data booklets. The service data shall contain complete instructions for installation, operation, maintenance, and a list of component parts. Wiring diagrams shall be included for electrical items or components. Each part of the unit, other than common fasteners, shall be identified in the service data and parts list by the manufacturer's part number.

The manikin, with all its formentioned components, shall be packaged in a suitable suitcase container for transport and storage.

In addition, the manikin shall include the following accessories:

FSC 6910

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

AMSC N/A

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Vent Meter.

An analog type vent meter which indicates the range of ventilation volumes from 0 to 2.0 liters shall be included. The vent meter face shall be color coded green and red to indicate correct volume and incorrect volume respectively. The vent meter shall be connected to the simulator via pneumatic tube permanently attached to the manikin.

Compression Meter.

A compression meter shall be included to indicate the depth of compression obtained when pressure is applied to the correct section of the sternum. The dial shall be calibrated in millimeters and shall have a color coded green segment indicating correct depth range of 38 to 51 mm (1-1/2 to 2 inches). The compression meter shall be connected to the simulator via a pneumatic tube permanently attached to the manikin.

Recording Device.

A recording device to record EAR and ECC performance on pre-printed pressure sensitive strip paper by two separate pens located inside the recorder shall be included. The paper transport shall be operated by a synchronously running motor/120 VAC, 60 Hz. The writing device shall record the volume, frequency, and duration of each ventilation, the depth and frequency of each compression; stomach inflation; wrong hand position; and the ratio between compression and relaxation. The recorder shall be connected to the simulator via pneumatic tubing permanently attached to the manikin.

The manikin accessories shall be packaged in a suitable carrying case container for storage and transport.

All of the above shall be over-packed in a commercially acceptable container

Toxicity. All elements shall be certified to consist of nontoxic materials which will not degrade to a toxic state.

Workmanship. The manikin shall be free from defects which detract from its appearance or impair its serviceability.

Unit Each (EA). One manikin, as specified, constitutes one unit

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Sampling.

For examination and testing. Sampling for examination and testing shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be one complete unit as specified.

TABLE I. Sampling for examination and testing.

	Inspection level	AQL (defects per hundred units)
For visual examination	I	1.0
Major defects		
Total defects (Major and Minor Combined)	I	4.0
For dimensional defects	S-1	2.5
All tests	S-2	0.0

For tests. Sampling for tests shall be conducted as follows:

a. First article. Where a first article sample is required, all tests shall be performed on that unit.

b Acceptance testing When a first article sample has not been required, one unit shall be subjected to all tests specified herein. If a first article sample has successfully passed all tests, performance testing shall not be required at acceptance

Examination. The complete unit shall be examined to determine compliance with the requirements of this specification in accordance with the classification of defects in table II.

TABLE II Classification of defects. *

Categories	Defects
Major	
101	Any component missing.
102	Any accessory part missing.
103	Any part does not fit properly.
104	Any movable part does not operate smoothly, and/or requires excessive force to operate.
105	Tubing not free of any kink, crimp, or unspecified flatness.
106	Any exposed or accessible surface not free of burrs or sharpness.
Minor	
201	External surface not free of scratches or otherwise marred.
202	Unit not clean, not free of dirt, oil, and grease.
203	Service data missing or incomplete.
204	Identification marking missing, incomplete, or illegible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

*Examination shall not be limited to the classified possible defects listed herein.

Tests.

a. Test that the vent meter operates as specified.
b. Test that the compression meter operates as specified.
c. Test that the recording device performs as specified.
d. Test that the lung simulates compliance as specified.
e. Test that the stomach bag inflates as specified.
f. Test that thorax section is removable and automatically and physiologically simulates the characteristics of an adult chest.
g. Test that the clicker in the thorax section sounds as specified.
h Test that the pulse variant bulb pulses as specified.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number are covered by this document:

6910-01-129-3195

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities and labeling and markingmust be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

PGC

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

Project No. 6910-0434
TXARED "C-3195"

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:54Z
Check the source to verify that this is the current version before use.