MILITARY SPECIFICATION

FORCEPS, DRESSING, EYE, STRAIGHT, 4 INCH

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers one type of eye dressing forceps, suitable for use in eye surgery.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-585	- Boxes, Wood, Wirebound.
PPP-B-601	- Boxes, Wood, Cleated-Plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

MILITARY

MIL-P-116	- Preservation-Packaging, Methods of.
MIL-L-10547	- Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, ATTN: DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

20627 Forceps, Dressing, Eye, Straight, 4 inch.

(Copies of specifications, standards, drawings and publications, required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM E 18 - Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials, Tests for.
ASTM E 92 - Vickers Hardness of Metallic Materials, Test for.
ASTM E 140 - Hardness Conversion Tables for Metals (Relationship Between Brinell Hardness, Vickers Hardness, Rockwell Hardness, Rockwell Superficial Hardness and Knoop Hardness).

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

(Technical society and technical association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

3. REQUIREMENTS

3.1 General requirements. The forceps shall be in accordance with the applicable requirements of GG-I-526 and as specified herein.

3.2 Material. Shall be fabricated from class 4, type 410, 420 or 420F material. As an alternate, type 420 material shall have a minimum carbon content of 0.12 percent and a maximum phosphorous content of 0.045 percent.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

3.3 Hardness. After proper stress relieving, the forceps shall have a Rockwell hardness of C40 to C48 when tested as specified in 4.4.1. The requirements of GG-I-526 shall apply.

3.4 Style, design and dimensions. Shall be a straight, regular dressing forceps suitable for use during surgical procedures involving the human eye. Forceps shall have serrated jaws that are consistent in depth and well defined. The action of the forceps shall be easy during opening and closure of the instrument. Forceps possessing guide pin shall not bind or stick in clearance hole on the opposite shank. The forceps shall be in accordance with DPSC Drawing 20627.

3.5 Finish. Finish shall be mirror.

3.5.1 Passivation. The entire instrument shall be properly passivated.

3.6 Identification marking. In addition to GG-I-526 requirements, forceps of foreign manufacture (other than U.S.) shall be marked "stainless".

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection shall be in accordance with GG-I-526 and as specified herein.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three (3) years after delivery of the supplies to which such records relate.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the corrosion-resistant steel, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:17Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for examination purposes shall be one forceps.

TABLE I. Sampling for examination.

Type of defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for test purposes shall be one forceps. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Forceps	Hardness	3.3 & GG-I-526	4.4.1	S-2	1.5
	Dimensions	3.4	----	S-2	2.5
	Copper Sulfate test	GG-I-526	GG-I-526	S-2	1.0
	Boil test	GG-I-526	GG-I-526	S-2	1.0

4.2.3 Raw materials. A sufficient quantity of each lot of metal used in the manufacture of the forceps shall be selected for chemical analysis.

4.3 Examination.

4.3.1 Examination of forceps. The forceps shall be examined to determine compliance with all requirements of this specification. Examination shall be conducted as specified in tables I and II. Examination, however, shall not be restricted to defects listed in table III. Any forceps in the sample containing one or more defects shall be rejected, and if the number of defective forceps in any sample exceeds the acceptance number for the sample, the lot represented by the sample shall be rejected.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:17Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects.

Categories	Defects
Major	
101	Forceps not of the design specified.
102	Jaw serrations not of uniform depth or not well defined.
103	Action of forceps not easy.
104	Guide pin, if provided, binds or sticks in corresponding clearance hole.
105	Jaw tips and jaw tip serrations not in proper alignment during opening and closing of forceps.
106	Finish not as specified.
107	Forceps not free from burrs, cracks, pits, rough areas, scale or corrosion.
Minor	
201	Identification markings not present, not correct, not complete, or not legible.

4.3.2 Packaging inspection. The inspection of the preservation-packaging shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Hardness test. Hardness shall be determined in accordance with ASTM E 18 or E 92. Conversion of Vickers hardness values to Rockwell values shall be in accordance with ASTM Hardness Conversion Tables E 140.

5. PACKAGING

5.1 Preservation-packaging. Preservation shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning and drying. Each forceps shall be cleaned and dried in accordance with the applicable methods as outlined in MIL-P-116.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

5.1.1.2 Preservative. Preservatives shall not be utilized.

5.1.1.3 Unit package. Each forceps shall be wrapped in neutral tissue or other suitable anticorrosive material. Forceps shall be further packaged in accordance with MIL-P-116, Method 1C-1.

5.1.1.4 Intermediate package. Six (6) forceps shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-566, PPP-B-676 or PPP-B-636, type CF, class domestic. Closure of box shall be as specified in the appendix of the applicable box specification.

5.1.2 Level C. Forceps shall be clean and individually packaged in a manner that will afford adequate protection against corrosion, deterioration and physical damage during shipment from supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Two hundred and eighty-eight (288) forceps shall be packed in an exterior container designed for a type 2 load, constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601 overseas type; PPP-B-621, class 2 or PPP-B-636, class weather-resistant, grade V3C. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraph of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof, as specified in the appendix of the applicable box specification.

5.2.2 Level B. Two hundred and eighty-eight (288) forceps shall be packed in an exterior container designed for a type 2 load, constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; PPP-B-636, type CF, class domestic. Closure of box shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The packaged forceps shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from supply source to the first receiving activity. These packs shall conform to the applicable carrier's rules and regulations.

5.3 Marking. Each unit, intermediate package and exterior container shall be marked as specified in MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Procurement documents shall specify the following:

(a) Number, title and date of this specification.
(b) Selection of applicable levels of packaging and packing required (see 5.1 and 5.2).
(c) National Stock Number.

6.2 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6515-00-333-8800	FORCEPS, DRESSING, Eye, Straight, 4 inch.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD - MB

Preparing acitivity:
DLA-DM

Project No. 6515-2080

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-F-37684 FORCEPS, DRESSING, EYE, STRAIGHT, 4 INCH

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO: 1974—603-680/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:17Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.