MILITARY SPECIFICATION

SCRAPER, UTERINE, CERVICAL, 7 INCH.

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a wooden scraper for obtaining a cell biopsy from the squamo-columnar junction of the cervix.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bids or request for proposals form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-601 - Boxes, Wood, Cleated-plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-corner.
PPP-B-636 - Boxes, Fiberboard.
PPP-B-676 - Boxes, Set-up.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Headquarters, Defense Personnel Support Center, ATTN: Directorate of Medical Materiel, Code ATT, 2800 South 20th Street, Philadelphia, Pa. 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116 - Preservation-packaging, Methods of.
MIL-L-10547 - Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-1363 - Measurement of Wood Moisture Content.

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. The scraper shall be fabricated from a durable, flexible, well-seasoned wood, of the highest quality, such as white birch, suitable for the purpose intended.

3.2 Style, design, and dimensions. The scraper shall be a cell biopsy scraper with two formed ends. One end shall have an elongated point, formed to obtain cell biopsy of the squamo-columnar junction of the cervix. The other end of the scraper shall be spatula shaped for the scraping of eroded or multiporous cervix. The scraper shall be shaped and conform to the dimensions of figure 1. All radii shall be smoothly blended and no sharp radii shall be accepted.

3.3 Finish. The scraper shall be smoothly finished, clean, free from foreign matter and fungus. The scraper shall be free from splinters, burrs, cracks, holes , pits, dry rot, warp, rough edges, sharp edges, fissures, tool marks and pitch pockets.

3.4 Moisture content. The moisture content of the finished scrapers shall range from 6 to 12 percent when tested as specified in 4.4.1.

3.5 Workmanship. The scrapers shall have first class workmanship throughout. Scrapers shall be free from defects which detract from their appearance or impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 20261O-21T11:17Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in table I. The unit of product for examination purposes shall be one scraper.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	None defined	

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one scraper. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Scraper	Dimensions	3.2	----	S-1	2.5
	Moisture content	3.4	4.4.1	S-1	1.5

4.3 Examination.

4.3.1 Scraper. The scrapers shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in tables I and III. Examination, however, shall not be restricted to defects listed in table III. Any scrapers in the sample containing one or more defects shall be rejected and if the number of defective scrapers in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

TABLE III. Classification of defects.

Categories	Defects
Critical	None defined.
Major	
101	Scrapers not of the design specified.
102	Scrapers not fabricated from wood.
103	Scrapers not smoothly finished including all edges.
104	Scrapers not free from splinters, burrs, foreign matter, cracks, holes, pits, rot, sharp edges, fissures, tool marks, fungus, or pitch pockets.
105	Scrapers warped.
106	All radii not smoothly finished.

4.3.2 Packaging inspection. The inspection of the preservation-packaging shall be in accordance with the group A quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Moisture content. The finished scrapers shall be used as the test specimen. Determination of moisture content shall be accomplished by the oven drying method specified in MIL-STD-1363.

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A, B, or C as specified (see 6.2).

5.1.1 Unit of issue. One package containing 100 scrapers, as specified, constitutes one unit of issue.

5.1.1.1 Level A.

5.1.1.1.1 Unit package. One hundred scrapers shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676. Closure of box shall be as specified in the appendix of the applicable box specification. Method III packaging, as specified in MIL-P-116 shall apply.

5.1.1.2 Level C. One hundred scrapers shall be packaged in a manner that will afford adequate protection against deterioration and physical damage during shipment from supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Twenty four (24) units of issue shall be packed in an exterior container designed for a type 2 load constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather resistant, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Fiberboard box shall be waterproof as specified in the appendix of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Twenty four (24) units of issue shall be packed in an exterior-container designed for a type 2 load constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure of box shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The packaged scrapers shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the applicable carrier rules and regulations.

5.3 Marking. Each unit package and exterior-container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The scraper covered by this specification is intended for use in obtaining cell biopsy of the cervix.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
(c) The National Stock Numbers (NSN).

6.3 This specification is intended to cover the following item listed in the Federal Supply Catalog.

National Stock Number	Item Identification
6515-01-004-1036	Scraper, Uterine, Cervical, 7 Inch.


Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activity:
DOD-MB

Preparing activity:
DLA-DM

Project No. 6515-2083

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:17Z
Check the source to verify that this is the currentversion before use.



UNLESS OTHERWISE SPECIFIED
ALL DIMENSIONS ARE IN INCHES.
TOLERANCES ON FRACTIONS ± 1/32.

FIGURE 1. Scraper, uterine, cervical, 7 inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-S-37688 SCRAPER, UTERINE, CERVICAL, 7 INCH.

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

GPO 1974—603-680/6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:17Z
Check the source to verify that this is the current version before use.