MILITARY SPECIFICATION

SILVER NITRATE APPLICATORS

This specification is approved for use by the Defense Personnel Support Center, Defense Supply Agency, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers Silver Nitrate Applicators (See 6.2).

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

GG-A-616	- Applicators, Disposable.
PPP-C-186	- Containers, Packaging and Packing for Drugs, Chemicals and Pharmaceuticals. (See 6.3)

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

FSC 6505

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:46Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U. S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

U. S. PHARMACOPEIAL CONVENTION, INC.

The United States Pharmacopeia.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

National Formulary.

(Application for copies should be addressed to the Mack Publishing Company, Easton, PA 18042.)

3. REQUIREMENTS

3.1 Material. Shall be wood applicators, 6 inches in length, minus 1/8, plus 1/4 inch, having a bead consisting of a fused mixture of 75 percent ± 2.5 percent silver nitrate and 25 percent potassium nitrate on one end. The beads shall be formed on the applicators (sticks) by dipping one end of the sticks approximately 1/4 of an inch into the fused mixture and allowing the beads to form by vertically withdrawing the sticks from the fused mixture and allowing the adhered fused mixture to cool and set into the form of a bead.

3.1.1 Sticks. The applicators (sticks) shall conform to GG-A-616, and shall meet the applicable requirements of sections 3 and 4 of that specification, except that the length shall be as specified in the document.

3.2 Ingredients.

3.2.1 Silver nitrate. The silver nitrate used in the manufacture of the product shall comply with the tests, standards, and requirements of the U.S.P., including any supplements or revisions thereto.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:46Z
Check the source to verify that this is the current version before use.

3.2.2 Potassium nitrate. The potassium nitrate used in the manufacture of the product shall be of a high quality and purity and suitable for use in the applicators.

3.3 Delivery. Not more than 6 months shall have elapsed from the date of manufacture of the product to the date of delivery to the Government.

3.4 Workmanship. The material and its containers shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period (potency period) specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the item to the Government.

4.1.2 No company supplying any ingredient(s) to the contractor will be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of product produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final containers.

4.3 Sampling. Sampling shall be conducted in accordance with MIL-STD-105.

TABLE I. Sampling.

For visual examination	Inspection level	AQL (percent defective)	Unit of product
Major	S-2	1.0	Filled immediate container
Minor	II	2.5	Filled immediate container

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:46Z
Check the source to verify that this is the current version before use.

4.4 Examination. The finished product shall be examined to determine compliance with all requirements of this specification. Nonconformance will be permitted to the extent indicated in 4.3.

4.4.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects:

TABLE II. Classification of defects

Categories	Defects 1/
Major	
101	Applicators not free from foreign matter.
102	Fused tip missing.
103	Applicators not homogeneous.
Minor	
201	Exterior of immediate container (unit package) is not free from contents or foreign matter.
202	Label not free from defacing marks and tears.

1/ Inspection is not restricted to the classified possible defects listed above.

4.5 Packaging inspection. The inspection of the intermediate package, packing and marking for shipment and storage shall be in accordance with quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4. Labeling inspection. Sampling shall be in accordance with MIL-STD-105 and the unit of product shall be the unit of issue. The inspection lot shall be the number of units of issue of each batch, lot, control number, etc., offered for inspection at one time. The inspection level for labeling shall be S-2 and the acceptance number shall be zero.

5. PACKAGING

5.1 Packaging requirements. The requirements for packaging shall be in accordance with all applicable requirements of PPP-C-186 (See 6.3) and as specified herein.

5.2 Packaging. Packaging shall be level A or B, as specified (See 6.1).

5.2.1 Unit of issue. One bottle containing one hundred applicators, as specified, constitutes one unit of issue.

5.2.2 Level A.

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:46Z
Check the source to verify that this is the current version before use.

5.2.2.1 Immediate container. The immediate container (unit package) shall be a suitable container considered acceptable by the Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other.

5.2.2.2 Intermediate package. Twelve units of issue shall be packaged as specified in 5.2.7.1 of PPP-C-186.

5.2.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.3 Packing. Packing shall be level A, B or C, as specified (see 6.1).

5.3.1 Level A. Twenty-four intermediate packages (288 units) shall be packed in an exterior container as specified in 5.4.4 of PPP-C-186.

5.3.2 Level B. Twenty-four intermediate packages (288 units) shall be packed in an exterior container as specified in 5.4.3.1.1 of PPP-C-186.

5.3.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.4 Marking.

5.4.1 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug and Cosmetic Act and shall include the information required below:

5.4.1.1 Immediate container. Each immediate container (unit package) label shall bear the following information:

(a) label information in accordance with commercial practice.
(b) the date of manufacture.
(c) the National Stock Number.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:46Z
Check the source to verify that this is the current version before use.

NOTES: (1) The National Stock Number shall not be preceded by any prefix.
(2) Strip labeling may be used.
(3) Label adhesion test shall not apply.

5.4.2 Intermediate package. Each intermediate package shall be marked as specified in 5.5.3 of PPP-C-186 except that the date of manufacture shall be shown in lieu of the date packed.

5.4.3 Exterior container. Exterior container shall be marked as specified in 5.5.4 of PPP-C-186 except that the date of manufacture shall be shown in lieu of the date packed.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) National Stock Number.
(c) Selection of applicable levels of packaging and packing (See 5.2 and 5.3).

6.2 This specification covers the following item(s):

National Stock Number	Item Identification
6505-00-299-9672	SILVER NITRATE APPLICATORS, 6 inch, 100s

6.3 All references to PPP-C-186 refer to Interim Federal Specification PPP-C-00186B which is mandatory for use until superseded by a revision.

Preparing activity:
Defense Supply Agency - DM

Project No. 6505-1615

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:46Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-S-37465(DSA-DM) - SILVER NITRATE APPLICATORS

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

☆ GPO 1974—603-680 6126 2-1

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T10:46Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, PA 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T10:46Z
Check the source to verify that this is the current version before use.