

FED. SUP CLASS
6640

REVIEW ACTIVITIES: ARMY, MD, ME
NAVY, AS, OS, YD



MS36329-1



MS36329-2

PART NUMBER	APPROXIMATE OVERALL DIMENSIONS inches (mm)				NUMBER OF STATIONS	CLOCK CYCLE HOURS
	HEIGHT	WIDTH	DEPTH	DIAMETER		
MS36329-1	26.5 (673)	-	-	20 (508)	12	24
MS36329-2	25 (635)	35 (889)	20 (508)	-	12	2 to 48

MATERIAL: The materials shall meet the design, finish and performance requirements of this standard.

DESIGN: The tissue processor and accessories shall be designed to suitably perform the required functions.

FINISH: All ferrous metal parts shall be treated with a rust preventing compound.

PERFORMANCE: The tissue processor shall automatically perform the necessary stages of tissue processing including fixation, dehydration, washing, impregnation and staining. Operations shall be accomplished to a predetermined time cycle programmed to follow a prescribed technique. The processor shall be capable of precisely duplicating a process with repreated uniform results.

ACCESSORIES: MS 36329-1 shall have the following accessories: glass beaker, plastic beaker, tissue container, tissue basket, wax bath, timing disc, notching pliers, beaker cover and beaker cover carrier.

MS 36329-2 shall have the following accessories: stainless steel tissue capsule with cover, white plastic tissue capsule (disposable) with presslock cover (case of 1000), white plastic cassetts (disposable) without cover (case of 1500) and cassetts covers (package of 25).

NOTE:
1. Overall height of MS36329-1 is with superstructure raised.
2. Items diffcring from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance.

APPROVED 3 Sep 1963 REVISED (A) 31 MAR 1981

(A) ENTIRE STANDARD REVISED

This standard has been approved by the Department of the Army and is mandatory for use by that activity. All other military activities are required to employ this standard where suitable.

P.A.	EA	TITLE	MILITARY STANDARD
Other Cust	MS 99	TISSUE PROCESSOR, AUTOMATIC, LABORATORY	MS36329
Procurement Specification None		SUPERSEDES:	SHEET 1 OF 1

Source: https://assist.dla.mil -- Downloaded: 2020-10-23T17:38Z
Check the source to verify that this is the current version before use.

DD FORM 1 SEP 63 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED.

Project Number 6640-0713

DATE