A-A-50720
30 December 1988

COMMERCIAL ITEM DESCRIPTION

VACUUM GAGE, PORTABLE

Scope. This commercial item description covers the basic requirements for a portable vacuum gage that is sealed in a carrying case. This unit is to be used in monitoring or checking the vacuum of any cryogenic tanks and should have the capability of rapid and easy calibration to any thermocouple with an operating range of 0 to 25 milliamperes.

APPLICABLE DOCUMENTS

MIL-P-116	Preservation Method of.
MIL-STD-129	Marking for Shipment and Storage.
PPP-B-636	Box Shipping Fiberboard.

(Copies of specifications, standards, publications and other documents required by contractors in connection with specific acquisitions should be obtained from the contracting activity or as directed by the contracting Officer.)

Salient characteristics. The vacuum indicator shall have the following characteristics.

1. Power supply. Power supply for the indicator shall be 1.5 volt "D" size battery.

2. Pressure range. Pressure range shall be 0-1000 millitorr.

3. Connection type. Connecting cable shall be thermocouple sensing tube with an octal type plug to interface with thermocouple tube, NSN 6685-00-877-9593.

4. Carrying case. Carrying case construction shall be rugged enough to protect contents. Case shall have a handle for carrying purpose and be resistant to weather conditions.

5. Gage unit size. Size shall be compact and enhance portability and storage.

6. Dial. The dial shall be calibrated in units of pressure (millitorr) and be sufficiently large to be readable without the use of visual aids.

7. Compatibility. Gage shall be compatible to the thermocouple sensing tube NSN 6685-00-877-9593.

AMSC FSC 6685

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

8. Instructions and charts. An instruction sheet containing pertinent information on adjustment and use of the unit shall be furnished with each instrume

Preservation, packaging, labeling and marking. The items shall be packaged in accordance with method IC-1 of MIL-P-116 and placed in a PPP-B-636 container. Shipping containers shall conform to the Uniform Freight-Rules Classification or regulations of other carriers as applicable to the mode of transportation. Markings shall be in accordance with MIL-STD-129.

Certification. The contractor shall certify by letter that the product offered meets the salient characteristics of this description and that the product conforms to the producer's own drawings, specifications, standards and quality assurance practices and is the same product offered for sale in the commercial marketplace. Small businesses that cannot offer products with established commercial market acceptability, may compete by submission of bid samples. (The Government reserves the rights to require proof of such conformance before first delivery and thereafter as may be otherwise provided for under the provisions of the contract).

Custodian:
Air Force - 99

Review activities:
Army - ER
Navy - AS
DLA - GS

Preparing activity:
Air Force - 82

(Project 6685-0813)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.