COMMERCIAL ITEM DESCRIPTION

HYGROTHERMOGRAPH

The General Service Administration has authorized the use of this Commercial Item Description

Scope. This document covers the basic requirements for a portable hygrothermograph that can simultaneously measure and record temperature and humidity variations in the air. The instrument shall be complete and easy to operate when used in accordance with its intended application.

Salient characteristics. The hygrothermograph shall have at least the following characteristics:

1. Self-contained. The instrument shall be self-contained with no external connections for operation or use.

2. Relative humidity. Relative humidity shall be measured directly by using a sensitive hygroscopic measuring element mechanically linked to a recording pen. The instrument shall record relative humidity from 0% to 100% with accuracy of 3%.

3. Temperature. The instrument shall measure and record temperature from 0° F to 100° F (-18° C to +38° C). Accuracy shall be ±2% of full scale.

4. Weight. Maximum weight for the hygrothermograph shall not exceed 18 pounds (lb) (8.16 Kilogram (Kg)).

5. Case. The hygrothermograph case shall be made of an impact and corrosion resistant material. If a light weight metal is used, it shall have a corrosion resistant surface coating. A handle shall be provided on the top of the unit for carrying purpose.

6. Chart drive. The instrument shall have a spring wound chart drive and it shall operate for at least 26 hours between windings.

7. Chart. Charts provided for the instrument shall be of circular type with 8 inch (in.) (203.20 millimeter (mm)) in diameter. Days and hours shall be labelled uniformly around the chart; temperature and humidity scale shall be labelled radially with increments of 10 from 0 to 100. Between each 10th increment there shall be 5 smaller increments. A pack of 100 charts shall be supplied with each unit.

8. Recording time. The hygrothermograph shall have a 24 hour chart drive.

9. Connecting tube. Connecting tube shall be stainless steel and equivalent to type 0 of Weksler Instrument Corp.

AMSC N/A FSC 6680

DISTRIBUTION STATEMENT A. Approved for public release, distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.

10. Inking. The inking pens for the hygrothermograph shall be reliable and easy to replace. The ink colors shall be blue and red and shall last for at least 3 months of continuous use.

11. Dimensions. Case size for the instrument shall not exceed 10x10x5 in. (254x254x127 mm).

12. Glass window. A shatterproof glass window shall be provided in the hinged cover of the instrument to make chart reading easy.

13. Calibration adjustment. Calibration adjustments for the inking pens shall be easy and accurate when performed by the operator.

14. Information supplied with each unit. Installation, operation, and maintenance procedures shall be provided with each instrument.

Certification. The contractor shall certify by letter that the product offered meets the salient characteristics of the description, and that the product conforms to the producer's own drawings, specifications, standards, and quality assurance practices, and is the same product offered for sale in the commercial marketplace. The contractor shall submit a bid sample to the procuring activity for functional evaluation.

Packaging, packing, and marking. The items shall be packaged in accordance with normal commercial practice and packed to assure acceptance by common carrier and provide product protection against loss and damage during multiple shipments, handling, and storage. The shipping container shall be in compliance with national motor freight classification and uniform freight classification. Shipping container shall be marked in accordance with commercial practice.

Custodian:
Air Force - 82
Army - ME

Preparing activity:
Air Force - 82

Review activity:
DLA - GS

Project 6680-0208

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:46Z
Check the source to verify that this is the current version before use.