MILITARY SPECIFICATION

FREQUENCY METERS, COAXIAL: FR-153/USM-129 FR-154/USM-129, FR-155/USM-129

This amendment forms a part of Military Specification MIL-F-46349(Ord), dated 26 June 1962.

Paragraph 3.4.5: Delete and substitute:

"3.4.5 Accuracy.- The accuracy within the temperature range of 32°F to 110°F shall be 0.2%."

Custodian:

Army - MU

Preparing activity:

Army - MU

Project No. 1290-A167Y

Source: https://assist.dla.mil -- Downloaded: 2026-10-28T05:40Z
Check the source to verify that this is the current version before use.