MIL-M-29198A(NAVY)
5 June 1985
SUPERSEDING
MIL-M-29198(NAVY)
10 December 1979

MILITARY SPECIFICATION

MONITOR, CARBON MONOXIDE, CONTINUOUS AND ALARM (FOR BREATHING AIR SYSTEMS)

This specification is approved for use by the Department of the Navy, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a continuous carbon monoxide (CO) monitor system with both visible and audible alarms.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. Unless otherwise specified (see 6.2), the following specifications and standards of the issue listed in that issue of the Department of Defense Index of Specifications and Standards (DoDISS) specified in the solicitation, form a part of this specification to the extent specified herein.

SPECIFICATIONS

MILITARY

MIL-V-173 - Varnish, Moisture- and-Fungus-Resistant (for Treatment of Communications, Electronic, and Associated Equipment).
MIL-E-17555 - Electronic and Electrical Equipment, Accessories, and Repair Parts, Packaging and Packing of.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Commanding Officer (Code 156), Naval Construction Battalion Center, Port Hueneme, CA 93043, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6665

Source: https://assist.dla.mil -- Downloaded: 20261O-20T22:11Z
Check the source to verify that this is the current version before use.

STANDARD

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.

(Copies of specifications, standards, handbooks, drawings, and publications required by manufacturers in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. The issues of the documents which are indicated as Department of Defense adopted shall be the issue listed in the current DoDISS and the supplement thereto, if applicable.

AMERICAN NATIONAL STANDARDS INSTITUTE, INC. (ANSI)

C39.1 - Electrical Analog Indicating Instruments, Requirements for.
C39.6 - Electrical Instrumentation, Digital Measuring Instruments for.

(Application for copies should be addresssed to the American National Standards Institute, Inc., 1430 Broadway, New York, NY 10018.)

NATIONAL ELECTRICAL MANUFACTURERS ASSOCIATION (NEMA)

ICS 6 - Enclosures for Industrial Controls and Systems

(Application for copies should be addressed to the National Electrical Manufacturers Association, 2101 L Street, N.W., Washington, DC 20037.)

(Industry association specifications and standards are generally available for reference from libraries. They are also distributed among technical groups and using Federal agencies.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence.

3. REQUIREMENTS

3.1 Description. The CO monitor shall be capable of continuous online sampling operations. The monitor shall be capable of detecting concentrations of CO as low as 2 parts per million and shall have a range of either 0 to 50 parts per million or 0 to 100 parts per million. If the monitor has a multiple range selection capacity, an additional range of 0 to 500 parts per million may be included. The monitor alarm system shall give a continuous audiovisual indication when the preset (adjustable) threshold limit is reached. The visual part of the monitor alarm system shall be integral with or attached to the monitor. The monitor shall have auxiliary relays for remote alarm indication and for external circuit control. The inlet port of the monitor shall contain a 0.5-micron or smaller particulate filter. The CO

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T22:11Z
Check the source to verify that this is the current version before use.

monitor shall provide a continuous indication of the CO concentration in parts per million on a panel-mounted meter or digital display. The CO concentration display, either meter or digital, shall be mounted integrally on the internal control panel, along with the other essential controls and indicators. The CO concentration display shall be visible through a suitable clear plastic viewing window in the front cover of the monitor case. The monitor shall have an auxiliary output circuit, 0- to 1-volt (V) or 4- to 20-milliamperes direct current range, which can be connected to a suitable recorder to provide a permanent chart of CO concentration (in parts per million) versus time.

3.2 First article. When specified (see 6.2), the contractor shall furnish one monitor for first article inspection and approval (see 4.2.1 and 6.5).

3.3 Standard commercial product. The CO monitor shall, as a minimum, be in accordance with the requirements of this specification and shall be the manufacturer's standard commercial product. Additional or better features which are not specifically prohibited by this specification but which are a part of the manufacturer's standard commercial product, shall be included in the CO monitor being furnished. A standard commercial product is a product which has been sold or is being currently offered for sale on the commercial market through advertisements or manufacturer's catalogs, or brochures, and represents the latest production model.

3.4 Materials. Materials used shall be free from defects which would adversely affect the performance or maintainability of individual components or of the overall assembly. Materials not specified herein shall be of the same quality used for the intended purpose in commercial practice. Unless otherwise specified herein, all equipment, material, and articles incorporated in the work covered by this specification are to be new and fabricated using materials produced from recovered materials to the maximum extent possible without jeopardizing the intended use. The term "recovered materials" means materials which have been collected or recovered from solid waste and reprocessed to become a source of raw materials, as opposed to virgin raw materials. None of the above shall be interpreted to mean that the use of used or rebuilt products are allowed under this specification unless otherwise specified.

3.4.1 CO concentration display. The display shall be either a graduated meter or digital at the option of the manufacturer unless only one type is specified (see 6.2). If a meter display conforming to ANSI C39.1 (see 2.2) is used, the scale shall be graduated linearly from zero to full scale in 0.5 or 1 parts per million increments. The size of meter provided shall have a distance between any two successive 1 parts per million divisions of not less than 0.04 inch measured along the arc. If a digital display conforming to ANSI C39.6 (see 2.2) is used, the numerals shall indicate the parts per million with the minimum increment being 0.5 parts per million. The numerals of a digital display shall be a minimum of 1/4 inches high and shall be distinguishable at a minimum distance of 6 feet.

3.4.2 Relay contacts. The relay contacts for controlling the remote alarm and other control circuits of the CO monitor shall be Single-Pole Double-Throw (SPDT), not less than 3 ampere, 120/240V, 50/60 Hertz (Hz), and hermetically sealed. There shall be not less than two such SPDT relays.

Source: https://assist.dla.mil -- Downloaded: 20261O-20T22:11Z
Check the source to verify that this is the current version before use.

3.5 Performance. The CO monitor shall exhibit the following operational performance.

3.5.1 Warm-up time. The CO monitor shall not require an initial warm-up time (time from connecting 120V, 50/60-Hz power source, and turning on the power switch) greater than 16 hours when tested as specified in 4.5.1.

3.5.2 Accuracy. The CO monitor shall be accurate within ±2 percent of full scale when tested as specified in 4.5.2.

3.5.3 Response time. The CO monitor shall have a response time of less than 60 seconds to attain 90 percent of the final CO concentration reading when tested as specified in 4.5.3.

3.5.4 Sensitivity. The CO monitor continuous indicating panel-mounted meter shall display a linear movement of the instrument pointer of not less than 0.04 inch per 1 parts per million CO concentration when tested as specified in 4.5.4.1, or one digit per 1 parts per million when tested as specified in 4.5.4.2.

3.5.5 Ambient temperature drift. The ambient temperature drift of the CO monitor shall not exceed ±2 percent of full scale per 20 degrees Fahrenheit (°F) when tested as specified in 4.5.5.

3.5.6 Power source limits. The CO monitor shall perform all of the required functions adequately, when working from a power source in the range of 95 to 130 volts and 60 Hz when tested as specified in 4.5.6. For portable monitors or as a backup for stationary monitors, a 12 to 24V direct current battery may be used as the power source.

3.6 Audible and visible alarm. The CO monitor alarm shall have an alarm set point adjustable from 2 to 50 parts per million. With the adjustable alarm set at 10 parts per million CO, both the audible alarm and the visual alarm shall be activated when tested as specified in 4.5.7.

3.6.1 Audible alarm. The CO monitor shall have an audible alarm (horn, bell, or noisemaker) capable of producing 100 decibels, but not to exceed 120 decibels, of pulsating sound at a 5-foot distance from the noisemaking device when tested as specified in 4.5.7.

3.6.2 Visible alarm. The CO monitor shall have a built-in visible alarm light (flashing or pulsating, red, amber, blue, or a high intensity strobe), not less than 1 inch in diameter and height, mounted on the case which is clearly visible at a 100-foot radius in a horizontal 150° field of view.

3.7 Additional requirements.

3.7.1 Auxiliary output circuit. The CO monitor shall have an auxiliary output circuit, adjustable from 0 to 1V or 4- to 20-milliamperes direct current range, which can be used with an appropriate recorder to permit the recovery of permanent records of air purity.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T22:11Z
Check the source to verify that this is the current version before use.

3.7.2 Fungus resistance. When specified (see 6.2), electrical components and circuit elements, including terminal and circuit connections, shall be coated with varnish conforming to MIL-V-173, except that:

a. Components and elements inherently inert to fungi or in hermetically sealed enclosures need not be coated.
b. Current-carrying contact surfaces, such as relay contact points, shall not be coated.

3.7.3 Bench handling. For handling during calibration or repair, the monitor must be capable of withstanding the shock loading applied during the bench handling test specified in 4.5.8.

3.7.4 Auxiliary accessories. To become operational each CO monitor must have certain test, calibration, and operational accessories. The following items shall be procured and packed with each monitor unless otherwise specified (see 6.2):

a. One 100-liter cylinder of CO calibrating gas at 20 parts per million concentration ±2 percent.
b. Ten 2-liter sampling bags.
c. One reusable cylinder regulator and pressure level indicator for use with the gas cylinders.
d. One sampling hose, 10 feet long, with adapter for remote monitoring and coupling from the air source to the monitor.
e. One air filtration panel which removes condensed water, oil, and solid particulate matter. Filtration panel must be designed for use with the specified CO monitor instrument (see 6.2).

3.8 CO monitor enclosure (case). The CO monitor unit shall be enclosed in a gasketed, keylocked, hinged front panel (with plastic viewing window), dust and vapor resistant case. The case shall be suitable for hand carrying, wall mounting, or free standing on a pedestal. The case shall conform to a NEMA type 12 or type 4 enclosure in accordance with NEMA ICS 6.

3.9 Identification marking. Identification and instructions shall be permanently and legibly marked directly on the monitor or on an aluminum, brass, or corrosion-resisting steel plate, firmly affixed to the item. The information shall include the manufacturer's name or trademark, model number, and serial number.

3.10 Commercial publications. Commercial publications, showing operation, calibration, maintenance, repair, and an applicable parts list, shall be furnished (see 6.3).

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The

Source: https://assist.dla.mil -- Downloaded: 20261-20T22:11Z
Check the source to verify that this is the current version before use.

Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Component and material inspection. Components and materials shall be inspected in accordance with all the requirements specified herein and in applicable referenced documents.

4.2 Classification of inspections. The inspection requirements specified herein are classified as follows:

a. First article inspection (see 4.2.1).
b. Quality conformance inspection (see 4.2.2).

4.2.1 First article inspection. The first article inspection shall be performed on one monitor when a first article is required (see 3.2 and 6.2). This inspection shall include the examination of 4.4, the tests of 4.5, and, when specified, the first article pack inspection of 4.6 (see 4.2 and 6.2). The first article may be either a first production item or a standard production item from the supplier's current inventory provided the item meets the requirements of the specification and is representative of the design, construction, and manufacturing technique applicable to the remaining items to be furnished under the contract.

4.2.2 Quality conformance inspection. The quality conformance inspection shall include the examination of 4.4 and the tests of 4.5. This inspection shall be performed on the samples selected in accordance with 4.3.

4.3 Sampling. Sampling and inspection procedures shall be in accordance with MIL-STD-105. The unit of product shall be one monitor. All monitors offered for delivery at one time shall be considered a lot for the purpose of inspection. If an inspection lot is rejected, the contractor may rework it to correct the defects, or screen out the defective units, and resubmit for a complete reinspection. Resubmitted lots shall be reinspected using tightened inspection. If the rejected lot was screened, reinspection shall be limited to the defect causing rejection. If the lot was reprocessed, reinspection shall be performed for all defects. Rejected lots shall be separate from new lots, and shall be clearly identified as reinspected lots.

4.3.1 Sampling for examination. Examination shall be based on inspection level II and an Acceptable Quality Level (AQL) of 2.5 percent defective.

4.3.2 Sampling for tests. Tests shall be based on inspection level S-3 and an AQL of 4.0 percent defective.

4.4 Examination. Each monitor selected shall be examined for compliance with the requirements specified in section 3 of this specification. Any redesign or modification of the contractor's standard product to comply with specified requirements, or any necessary redesign or modification following failure to meet specified requirements shall receive particular attention for adequacy and suitability. This element of inspection shall encompass all visual examinations and dimensional measurements. Noncompliance with any specified requirement shall constitute one defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T22:11Z
Check the source to verify that this is the current version before use.

4.5 Tests. The first article shall receive the tests of 4.5.1 through 4.5.8. Unless otherwise specified (see 6.2), each production unit selected shall receive the tests of 4.5.1 through 4.5.8. Failure to pass any test shall constitute one defect.

4.5.1 Initial warm-up test. After initially connecting the instrument to a 120-volt alternating current (ac) power source, allow the instrument to electronically burn-in for 16 hours. Zero the instrument with zero gas or CO-free air. Introduce a Primary Standard Gas of 20 parts per million CO in air and record the reading on the panel meter. Failure of the instrument to indicate a constant reading for ten minutes of 20 parts per million within ±2 parts per million shall be cause for rejection.

4.5.2 Repeatability. Calibrate the monitor using a Primary Standard Gas of 20 parts per million CO in the air. During the next day after calibration, at intervals of 2, 4, 8, and 24 hours, check the instrument zero reading with zero gas or CO-free air, then supply the monitor with the 20 parts per million CO Primary Standard Gas and record the readings. The monitor shall be accurate within ±2 parts per million for each of the readings.

4.5.3 Response time test. The response time is the time required for the instrument to react to a change in input CO concentration. After calibrating the instrument for 20 parts per million CO in air and purging it of all gas, introduce a Primary Standard Gas of 20 parts per million CO in air. Record the time required by the instrument to indicate 90 percent of the final 20 parts per million reading (18 parts per million). This test shall be conducted three times. To be acceptable, this recorded time shall be not greater than 60 seconds for each individual test.

4.5.4 Sensitivity test.

4.5.4.1 Indicating meter. The linearity of an instrument using a panel indicating meter is the ratio of the linear movement of the meter pointer on the instrument to the change in the measured variable causing this motion. Introduce a Primary Standard Gas consisting of 20 parts per million CO. Record the reading and measure the distance the pointer traveled to determine that the linear movement was not less than 0.04 inch per parts per million CO indicated. The test shall be conducted not less than three times.

4.5.4.2 Digital display. The linearity of an instrument using a digital display is the ratio of the numeric difference to the change in the measured variable causing this difference. Introduce a Primary Standard Gas consisting of 20 parts per million CO. Record the reading and determine the numeric difference to determine that the difference was not less than 20 parts per million CO indicated. The test shall be conducted not less than three times.

4.5.5 Ambient temperature drift test. Temperature drift is the ratio of induced bias in output reading to a change in ambient temperature. Place the monitor in a controlled temperature chamber and allow the monitor to thermally equilibrate for a minimum of 1 hour. Calibrate using a Primary Standard Gas of 20 parts per million CO in air at an ambient temperature of 73 ±3°F. Increase the temperature in the chamber by 20°F and allow the monitor to thermally equilibrate for a minimum of 1 hour. Read the parts per million CO

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T22:11Z
Check the source to verify that this is the current version before use.

level of the Primary Standard Gas at the elevated temperature. If the recorded temperature drift exceeds ±2 parts per million for a 20°F change, the instrument shall be rejected.

4.5.6 Power source limits test. Introduce a Primary Standard Gas of 20 parts per million CO in air and calibrate the instrument at the ambient power level of 120V ±2V ac at 60 Hz ±1 Hz. Record the parts per million CO level reading observed at this time. Retaining the preceding conditions, except for the ambient power level, record the parts per million CO level reading observed as follows:

a. With a power source of 90V ±2V at 60 Hz ±1 Hz.
b. With a power source of 130V ±2V at 60 Hz ±1 Hz.

The monitor shall not deviate greater than ±2 percent full scale for conditions (a) or (b) when compared to the initial calibration.

4.5.7 Audible and visual alarm test. Calibrate the CO monitor to a Primary Standard Gas consisting of 20 parts per million CO in air. Purge the system, set the adjustable alarm at the 10 parts per million level at which the alarm system responds. Failure of both audio and visual alarms to respond in the 8 to 12 parts per million CO range shall be cause for rejection. During the time when the alarm system responds, the audible part of the system shall be checked with an a calibrated type II sound level meter. The sound level meter shall be accurate to within ±5 percent of the full scale reading of the instrument. With the sound level meter on the A-weighting scale, take one reading at each of the four sides of the CO monitor under test. Each reading shall be taken at a distance of not less than 5.0 feet. Failure of the four readings taken to average at least 100 decibels shall be cause for rejection.

4.5.8 Bench handling test. The CO monitor, which has been calibrated to known Primary Standard Gas of 20 parts per million CO, shall be subjected to the test procedures to determine the ability of equipment to withstand the shock encountered during servicing. The monitor shall be placed on a horizontal, solid wooden bench top at least 1 5/8 inches thick. The test shall be performed, as follows, in a manner simulating shocks liable to occur during servicing.

Step 1 - Using one edge as a pivot, lift the opposite edge of the chassis until one of the following conditions occurs (whichever occurs first):

a. The chassis forms an angle of 45° with the horizontal bench top.
b. The lifted edge of the chassis has been raised 4 inches above the horizontal bench top.
c. The lifted edge of the chassis is just below the point of perfect balance.

Let the chassis drop back freely to the horizontal bench top. Repeat, using other practical edges of the same horizontal face as pivot points, for a total of four drops.

Source: https://assist.dla.mil -- Downloaded: 202610-20T22:11Z
Check the source to verify that this is the current version before use.

- Repeat step 1, with the test item resting on other faces until it has been dropped for a total of four times on each face on which the test item could be placed practicable during servicing. The test item shall not be operating during the test. After this procedure, subject the monitor to the test of 4.5.3 to verify that no damage has occurred.

Packaging inspection. The inspection of the preservation, packing, :ation, and marking shall be in accordance with section 4 of '555. The inspection shall consist of the quality conformance .on; and, when specified (see 6.2), a first article pack shall be :d for examination and test within the timeframe required (see 6.2).

'ACKAGING

Preservation, packing, palletization, and marking. Preservation, palletization, and marking shall be in accordance with the ients of MIL-E-17555 with the level of preservation and the level of as specified (see 6.2).

'OTES

Intended use. The CO monitor is intended to be used as an online azard safety system in all industrial compressed air control plants, e used throughout the Naval Shore (Field) Activities to supply ed air for breathing purposes.

Ordering data. Purchasers shall select the preferred options d herein and include the following information in procurement s:

- Title, number, and date of this specification
- When specification and standards are other than as specified (see 2.1.1).
- When first article is required for inspection and approval (see 3.2).
- When type of display is other than as specified (see 3.4.1).
- When treatment for fungus resistance is required (see 3.7.2).
- What type of air filtration panel to be compatible with specified monitor (see 3.7.4).
- When tests are other than as specified (see 4.5).
- When a first article pack is required and the timeframe required for submission (see 4.2.1 and 4.6).
- Level of preservation and level of packing required (see 5.1).

Data requirements. When this specification is used in an acquisition corporates a DD Form 1423, Contract Data Requirements List (CDRL), the uirements identified below shall be developed as specified by an DD Form 1664, Data Item Description (DID), and delivered in ce with the approved CDRL incorporated into the contract. When the ns of DoD Federal Acquisition Regulations Supplement 27.410-6 are ed the DD Form 1423 is not used, the data specified below shall be

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T22:11Z
Check the source to verify that this is the current version before use.

delivered by the contractor in accordance with the contract or purchase order requirements. Deliverable data required by this specification is cited in the following paragraphs:

Paragraph No.	Data requirements title	Applicable DID No.	Option
3.10	Commercial Manual	DI-M-4022C	

(DIDs related to this specification, and identified in section 6 will be approved and listed as such in DoD 5000.19L, Vol. II, Acquisition Management Systems and Data Requirements Control List. Copies of DIDs required by the contractors in connection with specific acquisition functions should be obtained from the Naval Publications and Forms Center or as directed by the contracting officer.)

(DIDs related to this specification, and identified in section 6 will be approved and listed as such in DoD 5000.19L, Vol. II, Acquisition Management Systems and Data Requirements Control List. Copies of DIDs required by the contractors in connection with specific acquisition functions should be obtained from the Naval Publications and Forms Center or as directed by the contracting officer.)

6.4 Training session requirements. The CO monitors are highly sophisticated electronic devices requiring a programmed test and calibration sequence both before and after being installed. It is essential that the firm providing the CO monitors also be contracted to teach designated civilian and military personnel the operational and maintenance requirements of these units.

6.5 First article. When a first article inspection is required, the item will be tested and should be a first production item or it may be a standard production item from the contractor's current inventory as specified in 4.2.1. The first article should consist of one complete monitor. The contracting officer should include specific instructions in acquisition documents regarding arrangements for examination, test, and approval of the first article.

6.6 Changes from previous issue. Asterisks are not used in this revision to identify changes with respect to the previous issue, due to the extensiveness of the changes.

Custodian:

Navy - YD

Review Activities:

Navy - MS, SH, AS

User Activity:

Navy - OS

Preparing Activity:

Navy - YD

(Project 6665-N448)

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T22:11Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-M-29198A(NAVY)	MONITOR, CARBON MONOXIDE, CONTINUOUS AND ALARM (FOR BREATHING

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
AIR SYSTEMS)	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*:
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T22:11Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEPARTMENT OF THE NAVY
Commanding Officer (156)
Naval Construction Battalion Center
Port Hueneme, CA 93043

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 12503 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEPARTMENT OF THE NAVY

Commanding Officer (156)
Naval Construction Battalion Center
Port Hueneme, CA 93043

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T22:11Z
Check the source to verify that this is the current version before use.