MIL-M-49158(EL)
1 May 1978

MILITARY SPECIFICATION

MEASURING SETS, RADIO INTERFERENCE, GENERAL SPECIFICATION FOR

This specification is approved for use by the Communications Research and Development Command, Department of the Army, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification defines solid-state (vacuum tubes permitted, see 3.5.4) RFI measuring sets capable of making quantitative measurements of radiated or conducted broadband or carrier radio frequency interference over the total range of 20Hz to 40GHz, hereinafter called the equipment.

1.2 Classification. Equipment defined by this specification shall be configured as Type II, Class 5, Style E, Color R, per MIL-T-28800 and as herein with the convertible/rackmountable capability.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on the date of invitation bids, or the request for proposal, form a part of this specification to the extent specified herein.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Commander, Communications Research and Development Command, Fort Monmouth, New Jersey 07703 by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 202610-22T19:53Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

MILITARY

MIL-T-28800	Test Equipment for use with Electrical and Electronic Equipment General Specification for
MIL-M-49158/1(EL)	Measuring Set, Radio Interference AN/URM-201
MIL-M-49158/2(EL)	Measuring Set, Radio Interference AN/URM-200
MIL-M-49158/3(EL)	Measuring Set, Radio Interference AN/URM-199

STANDARDS

MIL-STD-461	Electromagnetic Interference Characteristics, Requirements for Equipments
MIL-STD-462	Electromagnetic Interference Characteristics, Measurement of
MIL-STD-781	Reliability Tests, Exponential Distribution

(Copies of specifications, standards, drawings, and publications required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Specification sheets. The individual item requirements shall be as specified herein and in accordance with the applicable specification sheets. In the event of conflict between requirements of this specification and the specification sheet, the latter shall govern.

3.2 Classification of requirements. The requirements for the equipment are classified herein as follows:

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

Requirement	Paragraph
Safety	3.3
Parts, Materials and Processes	3.4
Design and Construction	3.5
Electrical Power Sources & Connections	3.6
Dimensions, Weight	3.7
Enclosure Requirements	3.8
Marking and Identification	3.9
Environmental Requirements	3.10
Reliability	3.11
Performance Requirements	3.12
Inspection Requirements	3.13

3.3 Safety. Unless otherwise specified herein, the equipment shall comply with the Type II safety requirements specified in MIL-T-28800.

3.4 Parts, materials, and processes. Unless otherwise specified herein, the equipment shall comply with the Type II parts, materials and processes requirements of MIL-T-28800.

3.4.1 Restricted materials. Equipment shall comply with the restricted material requirements of MIL-T-28800 including the requirements for mercury or radioactive materials and shall contain no combination of materials which cause deterioration of any material contained in the equipment due to the effects of outgassing.

3.5 Design and construction. Unless otherwise specified herein, the equipment shall comply with the Type II design and construction requirements of MIL-T-28800.

3.5.1 Mainframe plug-in concept. Equipment utilizing externally accessible, externally removable subassemblies (mainframe plug-in concept) do not meet the requirements of this specification and therefore are unacceptable. Items using this concept may be offered, provided the plug-in(s) shall not be removable by access through the front panel or rear panel with the unaided hand. A system part number shall be assigned to any such mainframe plug-in combination required to meet the requirements of this specification. The system part number shall be marked on the mainframe identification plate and any plug-in(s).

Source: https://assist.dla.mil -- Downloaded: 202640-22T19:53Z
Check the source to verify that this is the currentversion before use.

3. 5. 2 Controls. Unless otherwise specified herein, built-in adjustments and compensating devices shall not be externally accessible.

3. 5. 2. 1 Front panel controls. All controls which are required to operate the equipment throughout its specified performance characteristics, shall be located on the front panel.

3. 5. 3 Accessibility. The equipment shall be constructed so that:

3. 5. 3. 1 Subassemblies. Subassemblies and chassis components can be removed without removing any other hard wired subassembly, printed circuit card or component.

3. 5. 3. 2 Adjustments. Adjustments can be made without removing any component, printed circuit card or subassembly except the use of extender cards is permitted.

3. 5. 3. 3 Printed circuit cards. Printed circuit cards can be removed without the need to unsolder cables and interconnecting wiring (connections to all printed circuit cards shall be through pin and socket connectors) except for instances in which violation of this requirement is mandated by the need to maintain shielding effectiveness. Printed circuit cards (mother boards) designed primarily to distribute power and signals to other printed cards (daughter boards) are excluded from this requirement. When such mother boards are used, they shall be accessible from both sides to allow maintenance testing.

3. 5. 3. 4 Indicator lights. Unless approved by the procuring activity upon presentation of acceptable reliability data, indicator lights other than light emitting diodes (LEDs) shall be accessible from the operator's side of the front panel.

3. 5. 3. 5 Encapsulation and embedment. Encapsulation and embedment (potting) of subassemblies shall not be used (high-voltage circuits may be potted, if necessary).

3. 5. 4 Solid-state construction. Unless otherwise specified herein, the equipment shall be of solid-state, modular, miniaturized construction (vacuum tubes may be used where required by state-of-the-art circuitry).

Source: https://assist.dla.mil -- Downloaded: 202610-22T19:53Z
Check the source to verify that this is the current version before use.

3.5.5 Accessories. The equipment shall consist of an interference analyzer plus a group of accessories. The accessories for each specific equipment shall be as listed in the applicable specification sheet.

3.6 Electrical power sources and connections. Unless otherwise specified herein, the equipment shall comply with the Type II electrical power sources and connections requirements of MIL-T-28800. The equipment shall operate from a nominal 115/230 volts, single phase, 50, 60, and 400 Hz source.

3.6.1 Alternate source. The equipment shall be capable of operation from an external battery source at 24 Vdc, assuming use of a separate DC to AC inverter, suitably suppressed so as not to compromise intended operation of the measuring set. Measuring Sets intended for operation at frequencies up to 1 GHz shall also be operable from an internal battery for an 8-hour period without recharging.

3.6.2 Maximum power. The maximum power consumption of the equipment shall be 100 watts.

3.6.3 Input power selection device. An input power selection device shall be provided for selection of input power voltages of 115/VAC or 230VAC. Provision shall be incorporated to prevent accidental switching. When the equipment is delivered, the power selection device shall be in the 115VAC position.

3.6.4 Fuses and circuit breakers. Fuses or circuit breakers shall be in accordance with MIL-T-28800. (115VAC/230VAC) Either common or separate fuseholders may be provided. If only one fuseholder is used (common), the equipment shall be provided with the 115VAC fuse installed and the 230VAC fuse shall be stowed with the accessories.

3.6.5 Input power switch. A front panel mounted power switch shall be provided. The ON position shall have panel identification lights for AC operation. The switch shall break both sides of the power source.

3.7 Dimensions and weight.

3.7.1 Dimensions. The overall dimensions shall be 558 millimeters (22 inches) maximum depth and 457 millimeters (18 inches) maximum

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

height. The equipment width shall be as specified for Type II rack-mounted units in MIL-T-28800. A blank panel may be required to fit a standard vertical height as defined in MIL-T-28800 for rack-mounted Type II equipments.

3.7.2 Weight. A maximum weight of 34 kilograms (75 pounds) is allowable for any one major component of the equipment.

3.8 Enclosure requirements. Unless otherwise specified herein, the equipment shall comply with the Style E, enclosure requirements of MIL-T-28800, and be rack-mountable (see 6.2).

3.8.1 Accessory stowage. Various accessories are required to use this equipment. These are listed in the individual specification sheets. A separate suitable case shall be provided for storage of the accessories or spaces shall be provided within the equipment enclosure.

3.9 Marking and identification. Unless otherwise specified herein, the equipment shall comply with the Type II marking and identification requirements of MIL-T-28800.

3.9.1 Accessory case. The accessory case and each accessory stowed therein shall have identification plates. The case shall have within it a locator plate showing the proper location of each accessory within the container. The accessory case shall have individual compartments or nests for each accessory.

3.9.2 Supplemental identification plate. The supplemental identification plate specified in MIL-T-28800 shall contain the following data only:

a. Nomenclature
b. Procurement instrument identification number (PIIN).
c. Serial number
d. National stock number
e. US

3.10 Environmental requirements. Unless otherwise specified herein, the equipment shall comply with the Class 5, Style E, environmental requirements of MIL-T-28800.

3.10.1 Electromagnetic interference (EMI). The equipment shall comply with the following emission and susceptibility requirements of Notice 4, MIL-STD-461.

Source: https://assist.dla.mil -- Downloaded: 202640-22T19:53Z
Check the source to verify that this is the currentversion before use.

CE02	CS02	RE02	RS03
CE04	CS03	RE02.1	
CE06	CS04		
	CS06		

RE02 shall be performed from 10KHz to 1GHz. RE02.1 and RS03 shall be performed from 10KHz to 10GHz. RS03 susceptibility level shall be one volt per meter (1V/m).

3.10.2 Humidity. The equipment shall meet the humidity requirements in accordance with the test specified in paragraph 4.5.5.1.1.3 of MIL-T-28800.

3.10.3 Vibration. The equipment shall comply with the Class 5, Style E, vibration requirements of MIL-T-28800, except that the equipment need not be operating during vibration.

3.11 Reliability requirements.

3.11.1 Reliability burn-in. Each equipment delivered against this specification shall be submitted to a minimum 96 hour on-time burn-in procedure as specified in 4.4.5. The last 24 hours of burn-in shall be failure free.

3.11.2 Reliability. Reliability shall comply with requirements as specified herein. The specified MTBF shall be 2,500 hours when tested as specified in 4.4.4. A failure shall be defined in MIL-STD-781, para 3.1, and as any departure from the required performance or operation of the required accuracies (not correctable by normal use of the operating controls) after the test is initiated. Test Level B, MIL-STD-781 shall be the required test level.

3.11.3 Maintainability requirements. The equipment shall comply with the Type II maintainability requirements of MIL-T-28800.

3.12 Performance characteristics.

3.12.1 Frequency.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

3.12.1.1 Range. The frequency range of the equipment shall be as noted in the individual specification sheets.

3.12.1.2 Accuracy. The frequency indication accuracy shall be within ±2% of the indicated frequency. This accuracy shall not depend upon individually calibrated indicators, calibration charts, curves, or tables.

3.12.2 Impedance. Input impedance shall be 50 ohms.

3.12.3 Output meter. The output meter shall indicate in two parameters as noted below:

3.12.3.1 Decibels. The front panel meter shall be graduated into decibels (dB), with a total range of at least 40 dB and to an indication accuracy of ±2dB.

3.12.3.2 Voltage. The meter shall be graduated in microvolts, corresponding to the dB level of 3.12.3.1 relative to 1 microvolt.

3.12.4 Attenuation. Output attenuation shall be controlled from 0 to 80 db in 20 db steps with an indication accuracy of ±2db.

3.12.5 Input VSWR. The input voltage standing wave ratio shall be better than 2.0:1.

3.12.6 Signal sensitivity. The continuous wave (CW) signal sensitivity shall be as indicated in the appropriate specification sheet.

3.12.7 Undesired signals. Undesired signals shall be at least 45 down from the desired signals to provide full rejection of spurious signals, image signals, and IF signals.

3.12.8 Audio. An audio output will be provided through a front panel phone jack. The output will be at least 60 mw and have a frequency range from 20 to 2000 Hz.

3.13 Inspection requirements.

3.13.1 First article inspection. When specified in the contract or purchase order, the contractor shall furnish first article units in accordance with 4.3 (See 6.2).

Source: https://assist.dla.mil -- Downloaded: 202610-22T19:53Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.2 Classification of inspections. The inspections required herein are classified as in a and b:

a. First article inspection (see 4.3).
b. Quality conformance inspection (see 4.3).

4.3 First article or quality conformance inspection. Unless otherwise specified herein, the first article and quality conformance inspections shall be in accordance with MIL-T-28800.

4.3.1 Test plan. Unless otherwise stated in the contract, the contractor shall prepare a test plan in accordance with paragraphs 4.3(a) and 4.4(a) of MIL-T-28800, for use with both classes of inspection. The test plan shall include as a minimum the tests listed in TABLES I, II, & III, the tests noted in section 4.4, and a description of the satisfactory operation check and satisfactory operation test, as defined in para 4.5 of MIL-T-28800. Unless otherwise required, the tests and inspections to be performed shall be identical for both classes of inspection.

4.3.2 Inspection sampling plan. The inspection sampling to be performed during both classes of inspection shall be as specified in MIL-T-28800, except as indicated below:

4.3.2.1 Unless otherwise specified herein or in the contract, the Group C, D, E & F quality conformance inspections shall be performed on the first production lot only.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

4.3.2.2 The Group D EMI tests and Group F Reliability test shall be performed during quality conformance inspection only if not performed during first article.

4.4 Quality assurance tests. The following tests shall be performed in lieu of the corresponding tests specified in MIL-T-28800.

4.4.1 Electromagnetic interference. One equipment shall be subjected to an EMI test in accordance with MIL-STD-461 and the requirements of 3.10.1. All test set-ups and procedures shall comply with the measurement techniques of Notice 3, MIL-STD-462. All emission and susceptability tests shall be performed with the output cable connected to, and extended parallel with, the front of the equipment. The output cable shall be terminated to simulate a normal loading configuration.

4.4.2 Humidity. The equipment shall be subjected to the humidity test specified in paragraph 4.5.5.1.1.3 of MIL-T-28800.

4.4.3 Vibration. The equipment shall be subjected to the vibration tests specified in MIL-T-28800, for Class 5 equipment, except that the equipment shall not be operated during the test. The satisfactory operation test shall be performed prior to and following vibration testing.

4.4.4 Reliability test. The Group F reliability tests shall be performed as follows:

4.4.4.1 Reliability sampling plan.

4.4.4.1.1 First Article. The Group F reliability demonstration shall be conducted on at least five (5) samples. The use of ten (10) samples is encouraged, but not more than ten (10) samples shall be used.

4.4.4.1.2 Quality conformance inspection. The Group F reliability tests shall be performed on the first production lot, only if not performed during First Article testing. From those units of the first lot that have passed the required Group A and B tests, a random sample of ten (10) units shall be selected, for Group F testing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

TABLE I

Examination and Test Groups

Description	Requirement	Test Method
Group A Preoperational Inspection Leakage current Level A performance	4.5.3.1 * 3.3 TABLE II	4.3.1
Group B Level B performance	TABLE III	4.3.1
Group C Electrical power Environmental requirements Humidity Vibration	3.6 3.10 3.10.2 3.10.3	4.3.1
Group D Electromagnetic interference	3.10.1	4.4.1
Group E Dimensions Weight Front panel marking	3.7.1 3.7.2 3.9	4.3.1
Group F Reliability	3.11	4.3.1, 4.4.2

* Paragraph 4.5.3.1 of MIL-T-28800

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

TABLE II

Level A Performance Tests 1/

Description	Requirements
Power	3.6
Alternate Source	3.6.1
Power switches, input & selection	3.6.3, 3.6.4
Frequency range	3.12.1.1
Frequency accuracy	3.12.1.2
Output meter	3.12.3
Attenuation	3.12.4

1/ Level A testing is abbreviated testing (see 6.4.1).

TABLE III

Level B Performance Tests

Description	Requirements
Power input	3.6
Alternate source	3.6.1
Frequency range	3.12.1.1
Frequency accuracy	3.12.1.2
Input Impedance	3.12.2
Output meter	3.12.3
Attenuation	3.12.4
Input VSWR	3.12.5
Sensitivity	3.12.6
Undesired signals	3.12.7
Audio output	3.12.8

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

4.4.4.2 Reliability test plan. The reliability tests shall be conducted in accordance with the following test plan:

θ_o = 2,500 hours

Number of Failures	Total Test Time * Accept ** (Equal or More)	Reject *** (Equal or Less)
0	2,000	N/A
1	3,750	558
2	5,375	2063
3	6,875	3850
4	8,375	5725
5	9,875	7725
6	11,375	11375 ****

* Total test time is total unit hours of "equipment-on" time.
** Accept if test time is greater than or equal to that listed.
*** Reject if test time is less than or equal to that listed.
**** Reject if test time is strictly less than 11,375 with 6 failures.

4.4.4.3 Test length. Testing will continue until the total unit hours together with the total count of relevant equipment failures permit either an accept or reject decision in accordance with the specified test plan. Only equipment "ON" time may be used in MTBF or longevity determinations. Testing shall be monitored in such a manner that the times to failure may be estimated with reasonable accuracy. No single equipment "ON" time shall be less than one half the average operating time of all equipments on test.

4.4.4.4 Test conditions.

4.4.4.4.1 Operating mode (Duty Cycle). Prior to burn-in, a satisfactory operation test shall be conducted. Daily satisfactory operation checks shall be conducted. For the last 24 hour failure free period, a complete satisfactory operation test shall be conducted prior to and after the period.

4.4.4.4.2 Test level. While under test, the equipment shall be subjected to the conditions specified by Test Level B, MIL-STD-781.

Source: https://assist.dla.mil -- Downloaded: 2026-04-22T19:53Z
Check the source to verify that this is the current version before use.

4.4.4.4.3 Input voltage cycling. When so directed by the procuring activity, voltage cycling shall be accomplished as follows: The input voltage shall be maintained at one hundred ten percent (110%) nominal value for one-third of the equipment "ON" cycle, at the nominal value for the second one-third of the equipment "ON" cycle, and at ninety percent (90%) for the final one-third of the equipment "ON" cycle. This cycling procedure is to be repeated continuously throughout the reliability test.

4.4.4.5 Reliability examination and test method. The following inspections shall be used to verify equipment operation during reliability testing. The satisfactory operation test of 4.3.1 shall be used to confirm proper equipment operation prior to and following reliability testing. The satisfactory operation test shall also be performed weekly to verify equipment operation. The satisfactory operation check of 4.3.1 shall be used to monitor proper operation of the equipment daily.

4.4.4.6 Reliability failure actions, Group F. In the event of failure(s) during reliability testing, failure action shall be taken as required in MIL-STD-781, para 5.5.

4.4.4.7 Corrective action, Group F. In the event the reliability test reaches a reject decision, corrective action shall be taken as required in MIL-STD-781, para 5.7.

4.4 5 Burn-in. Each deliverable equipment shall be subjected to a minimum 96 hours on-time burn-in period, prior to Group A testing. During the last 24 hours of burn-in, the equipment must operate failure free. Up until this time, equipment will be allowed to accumulate failures. Each equipment which fails during the final 24 hour period shall be repaired and returned to test until it successfully survives a 24 hour period without failure. Failures which occur during the burn-in test shall be noted and reported but shall not count toward the establishment of equipment MTBF.

4.4.5.1 Burn-in conditions. Test Level B of MIL-STD-781 shall constitute the required test conditions. The input voltage cycling procedure per para 5.2.4 shall be invoked.

4.4.5.2 Burn-in tests. Prior to burn-in, the satisfactory operation test of 4.3.1 shall be conducted. Daily satisfactory operation checks (4.3.1)

Source: https://assist.dla.mil -- Downloaded: 202610-22T19:53Z
Check the source to verify that this is the current version before use.

shall be conducted. For the last 24 hour failure free period, a complete satisfactory operation test shall be conducted prior to and after the period.

4.4.6 Optional quality assurance system. The procuring activity may substitute a Department of Defense (DOD) approved supplier's quality assurance system for any or all tests in TABLE I.

5. PACKAGING

5.1 Preservation, packaging. Packaging for delivery shall be in accordance with MIL-T-28800 and as specified by the procuring activity.

6. NOTES

6.1 Intended use. The equipment is intended for use in maintaining electronic communications equipment by equipment repair units in intermediate and depot maintenance.

6.1.1 Equipment replaced. Equipment procured according to this specification is intended to replace all versions of various older equipments now fielded. These equipments to be replaced are listed in the appropriate specification sheets.

6.2 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification and any amendment thereto, plus the title, number and date of the applicable specification sheet and the complete type designation (See 3.1).
b. Packaging requirements (see Section 5).
c. When rough handling and functional tests are required.
d. Place of final inspection.
e. Technical literature required.
f. Quantity of tools and running spare parts required.
g. Marking for shipment and shipping containers.
h. Test plans and test reports.
i. Classification of inspection and number of samples required.
j. Rack-mounting requirements.
k. Alternate power requirements.
l. Maintainability rationale.
m. Reliability rationale.

Source: https://assist.dla.mil -- Downloaded: 202610-22T19:53Z
Check the source to verify that this is the current version before use.

6.3 Contract data requirements.

a. Nomenclature assignment.
b. National stock number.
c. Equipment sample test plans.
d. Pretest performance records.
e. First production test data.
f. Identification plate drawing.
g. Accessory plate drawing.
h. Locator plate drawing.

6.4 Definitions.

6.4.1 Level A performance tests. Level A testing (TABLE II) is a reduced amount of testing which is performed on each equipment produced. Its purpose is to insure that all functions and modes of operation of the equipment are evaluated without extensively checking each parameter as required in the Level B test (TABLE III). The approved equipment test procedure shall specify the actual amount of testing to be performed.

6.4.2 Reliability rationale. Reliability rationale submitted with the bid should provide clear and concise rationale showing how the reliability of the equipment complies with minimum requirements of the solicitation. This may include data from previous reliability tests, reliability predictions and other data available to the offeror. In the absence of such data on the equipment being offered, such data on similar equipment of equal or greater complexity produced by the offeror may be submitted. However, such data must be clearly identified as comparative data, and accomplished with specifications and technical literature on the similar equipment. The acceptance of the submitted rationale does not relieve the successful bidder of performing and successfully completing the production reliability testing.

6.4.3 Maintainability rationale. Maintainability rationale submitted with the bid should provide clear and concise rationale showing how the maintainability of the equipment complies with requirements of the solicitation. This may include data from previous maintainability tests, records of repair and calibration data, and other data on the equipment available to the offeror. Data on selection of module size and other design features

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

relevant to maintainability may also be submitted. The acceptance of the submitted rationale does not relieve the successful bidder of performing and successfully completing the production maintainability testing.

Custodian:
Army - EL

Reviewers:
Army - AV, AR, MD

User:
Army - AT, SC

Preparing Activity:
Army - EL

(Project 6625-A710)

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEPARTMENT OF THE ARMY
DOD – 314

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
US Army Electronics Command
ATTN DRSEL-RD-TS
Fort Monmouth, NJ 07703

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22 R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document DoD contractors government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government Fold on lines on reverse side, staple in corner, and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE MIL-M-49158(EL) MEASURING SETS, RADIO INTERFERENCE GENERAL SPECIFICATION FOR

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes" in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T19:53Z
Check the source to verify that this is the current version before use.