MIL-T-55202(MU)
AMENDMENT 1
12 October 1964

MILITARY SPECIFICATION

TEST SET, TRANSISTOR: AN/USM-136

This amendment forms part of Military Specification MIL-T-55202(MU), 2 November 1962.

Paragraph 2.1, Under DRAWINGS, MUNITIONS COMMAND: Delete "F8625338 - Scaling Amplifier" and substitute "F8625596 - Scaling Amplifier." Add:

"C8222453 - ALS Projection Readout"

Paragraph 3.6.1, subassemblies, item (a) (3): Delete Drawing "F8625338" and substitute "F8625996."

Paragraph 3.6.4.3, in first sentence: Delete, "a projector" and substitute "projectors." Add Item (i):

"(i) An ALS projection readout conforming to Drawing C8222453 shall indicate lead orientation, polarity and material of diodes and transistors when found by the automatic lead selector."

Paragraph 3.6.5: Delete third sentence in last paragraph and substitute the following:

"When LEADS KNOWN is selected the OPEN and SHORT test feature and PROJECTION READ_ OUT of the automatic lead selector is not obtainable for the semiconductor under test."

Paragraph 3.6.6.2.1.1, Test time, Item (a); Change "3 seconds" to "6 seconds."

Paragraph 3.7, Item (d): Delete.

Paragraph 4.1: Delete paragraph and substitute the following paragraph.

"4.1 Responsibility for inspection.- Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements."

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T22:37Z
Check the source to verify that this is the current version before use.

Paragraph 4.9.1.1.7: Delete paragraph and substitute:

"Vibration.- The transistor test set shall be non-operating with the front cover removed and shall be vibrated to the following tests:

Paragraph 4.9.1.1.9: Delete the first sentence and substitute:

"The front and rear fiberglass covers attached to an empty test set center section shall be placed in a test chamber equal to MIL-C-9436."

Paragraph 4.9.2.5, Item (e): Delete the first sentence and substitute:

"(e) RESET the test set to READY, move switch to MANUAL position, remove the oscilloscope leads, remove the inspection test card and replace with the 2N404 test card, Drawing D8625444."

Delete the last sentence and substitute:

"Move switch to LEADS UNKNOWN and place a known good 2N404 transistor into the TRANSISTOR test socket."

Item (f): Add the following after "a GO light shall have lit.":

"The ALS projection readout shall indicate PNP, germanium and the correct lead orientation."

Item (l): Delete "3 seconds" and substitute "6 seconds."

Item (m): Delete and substitute the following:

"(m) RESET the test set to READY, remove and replace the 2N404 transistor with a known good NPN silicon transistor such as a 2N699, but continue to use the 2N404 test card."

"(n) Press the TEST button and within approximately 6 seconds the leads shall have been found and the GO light shall have lit. NPN silicon and correct lead orientation shall be indicated on the ALS projection readout."

"(o) RESET the test set to READY, remove the test card and test transistor."

Paragraph 4.9.2.6.2, Item (b): Change "62.5 Volts" to "50 Volts" in the second sentence.

Paragraph 4.9.2.7.1, Item (a): Delete last sentence.

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T22:37Z
Check the source to verify that this is the current version before use.

Item (b): Delete and substitute the following:

"MANUAL ADVANCE into test position 1. Observe the value of current on the milliammeter while pressing the RECHECK button. The value of emitter-to-base current should not exceed 0.25 milliamperes for this PNP polarity test."

Item (c): Delete and substitute the following:

"Place an oscilloscope input probe across the 100K ohm test resistor with the oscilloscope ground lead on the emitter lead."

Custodian:
Army - MU

Preparing activity:
Army - MU
Project No. 6625-A170

Source: https://assist.dla.mil -- Downloaded: 2026-10-22T22:37Z
Check the source to verify that this is the current version before use.