A-A-51089A
26 August 1988
SUPERSEDING
A-A-51089
15 December 1985

COMMERCIAL ITEM DESCRIPTION

TUBE, STOMACH, SURGICAL (Salem)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers sterile, disposable, radiopaque stomach tubes of the Salem design suitable for enteral feeding and aspiration of stomach contents.

Salient characteristics.

General. The disposable stomach tube shall be suitable for nasogastric enteral feeding and aspiration of stomach contents.

Sizes. The stomach tube shall be provided in the following sizes:

Size	Outside diameter (French)
18	18 ± 1
16	16 ± 1
14	14 ± 1
12	12 ± 1
10	10 ± 1

Characteristics. The stomach tube shall have a length of 47 ± 2 inches, and shall have a radiopaque stripe throughout its length.

The stomach tube shall be of the Salem design, with the lumen separated into two channels. The larger primary channel shall be the suction channel. The smaller secondary channel shall be the ventilation channel. The channels shall interconnect inside the distal tip of the stomach tube. The exterior of the distal tip shall be closed and rounded.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

The proximal end of the suction channel shall be flared, or, as an alternate, shall have a tapered female connector permanently attached to it. The flare or connector shall accommodate a removable adaptor. The removable adaptor shall be provided with the stomach tube, and shall be compatible for use with both catheter-tipped syringes and rubber tubing.

The ventilation channel shall have a short piece of tubing (pigtail) inserted near the proximal end. The pigtail shall be permanently attached to the ventilation channel. The free end of the pigtail shall terminate in a funnel connector. The funnel connector shall make a mechanically secure and airtight seal with the removable adaptor. The pigtail shall be of sufficient length to allow the funnel connector to be easily attached to the removable adapter when the adaptor is inserted into the primary channel.

The distal end of the stomach tube shall have eleven (11) oval eyes cut into the suction channel. Ten of the oval eyes shall be arranged symmetrically about the radiopaque stripe starting at the distal end. The eleventh oval eye shall bisect the radiopaque stripe 2-1/2 to 3 inches from the distal tip.

The suction tube shall be marked at 18, 22, 26, 30 $\pm$ 0.25 inches respectively inches from the distal tip.

Material. The stomach tube shall be manufactured from a thermosensitive, flexible plastic. The pigtail shall be manufactured from a flexible plastic. The connector shall be manufactured from a rigid or semi-rigid plastic.

Workmanship. The stomach tube shall be free from defects which detract from its appearance or impair its function.

Unit. Package (PG). One package containing 50 individually packaged sterile tubes of one size only, as specified, constitutes one unit. The individual package shall be a sealed, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Catheters shall be examined to determine compliance with specification requirements.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one stomach tube.

Sterility. Each stomach tube shall be sterile when tested in accordance with the USP.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Biocompatibility. The stomach tube shall meet the requirements of the USP for Class IV plastics. The connector and pigtail shall meet the requirements of the USP for Class II plastics. A sufficient amount of raw material shall be tested to represent the lot or batch.

Performance. Tubes shall be tested for performance as follows:

The pigtail shall first be tested for security of assembly by immobilizing the tube and applying a two pound tensile force to the pigtail for 15 seconds (minimum). Any pigtail which separates shall be cause for rejection. The pigtail shall then be tested for leakage. Crimp the tube six to twelve inches distal to the tube/pigtail juncture. Place the removable adaptor into the pigtail and connect low-pressure air to the adaptor. Submerge the tube and pigtail in water. Bring the air pressure to 5 psig. Any leakage evidenced by bubbles shall be cause for rejection, whether the leakage occurs from the connector/pigtail fit or from a defect in the pigtail. Connector/tube primary channel fit shall be tested in the same manner as above. Sampling shall be level S-4 with an AQL of 1.0.

Certificate of biocompatibility. A certificate of compliance, stipulating offered items meet the requirements specified herein, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years, and that there has been no changes to material, formulation and method of processing. The certificate shall include the current contract number and applicable NSN, the date of actual testing, the test method and results, the name and title of the person evaluating test results, and a statement to the effect that material formulation and processing have not changed since the testing was performed.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	SIZE
6515-00-149-0315	12
6515-00-149-0316	16
6515-00-851-0310	14
6515-00-917-1912	18
6515-01-110-9182	10

Ordering data . Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

A-A-51089A

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4388

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51089A	TUBE, STOMACH, SURGICAL (Salem)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:50Z
Check the source to verify that this is the current version before use.