INCH-POUND

A-A-51625C
17 June 1994
SUPERSEDING
A-A-51625B
14 February 1991

COMMERCIAL ITEM DESCRIPTION

DISPOSAL CONTAINER, HYPODERMIC NEEDLE AND SYRINGE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a plastic, autoclavable, hypodermic needle and syringe disposable container.

Salient characteristics:

The hypodermic needle and syringe disposal container shall be a disposable, autoclavable, red, puncture-resistant, polypropylene container designed for safe disposal of used hypodermic needles and syringes, as well as other types of sharps. The container shall be supplied in the following sizes:

Size I	-	3 quart to 4 quart capacity
Size II	-	7 quart to 8-1/2 quart capacity
Size III	-	4-1/2 gallon to 5-1/2 gallon capacity
Size IV	-	8 gallon to 10 gallon capacity

The shape of the container shall be round, rectangular, or square. A biohazard symbol and the legend "Infectious Waste", "Caution", or other appropriate warning legend, shall be clearly marked on the side or front of the container.

The container shall withstand sterilization temperatures up to 275°F without softening, melting, or degrading, and shall be disposable through incineration.

The container shall sit on a flat surface without rock or wobble, and shall be free of cracks, tears, sharp edges, dirt, and other foreign materials.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Size I and III containers only.

The container shall be constructed with a single orifice and a one-way valve to prevent retrieval of disposed items and spillage. The opening shall accommodate 60 milliliter syringes. The container shall have a pre-attached, vented, locking lid to promote safety and prevent retrieval of disposed items.

The container shall be translucent to permit easy viewing so overfilling and jamming can be avoided. As an alternative, the top only may be translucent to see the fill level.

Size II container only.

The container shall have a 7 quart to 8-1/2 quart capacity and a pre-attached vented lid that closes tightly to prevent recovery of disposed items. The lid shall have a needle remover hole to assist in removing phlebotomy needles.

Size IV container only.

The container lid shall have two openings for disposal of items. The lid shall not be able to be removed after the lid is secured to prevent retrieval of disposed items. The container shall have some type of carrying handle and shall be nestable for space convenience. The container may be opaque or translucent.

Workmanship. The disposal container shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG).

For Size I, one package containing 24 disposal containers, as specified, constitutes one unit.

For Size II, one package containing 12 disposal containers, as specified, constitutes one unit.

For Size III, one package containing 8 disposal containers, as specified, constitutes one unit.

For Size IV, one package containing 10 disposal containers, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one disposal container as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for test purposes shall be one disposal container. Sampling shall be inspection level S-3 with an AQL of 2.5.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Numbers, Sizes and respective Item Identifications are covered by this document:

NATIONAL STOCK NUMBER	SIZE	ITEM IDENTIFICATION
6530-01-188-5345	I	DISPOSAL CONTAINER, HYPODERMIC NEEDLE AND SYRINGE, Plastic, 24s
6530-01-183-2863	II	DISPOSAL CONTAINER, HYPODERMIC NEEDLE AND SYRINGE, Plastic, 12s
6530-01-185-3844	III	DISPOSAL CONTAINER, HYPODERMIC NEEDLE AND SYRINGE, Plastic, 8s
6530-01-172-6992	IV	DISPOSAL CONTAINER, HYPODERMIC NEEDLE AND SYRINGE, Plastic, 10s

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6530-2433
Location: c:\eric\dispocid\sg

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given.

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-[illegible]625C	2 DOCUMENT DATE (YYMMDD) [illegible] June [illegible]

3 DOCUMENT TITLE

DISPOSAL CONTAINER, HYPODERMIC NEEDLE AND SYRINGE

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial: (215)7372870 (2) AUTOVON: 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church, VA 22041 3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 *Previous editions are obsolete* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.