A-A-51627
25 August 1986

COMMERCIAL ITEM DESCRIPTION

PASTE, DENTAL LABORATORY (Silicone Insulating and Separating)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a silicone insulating and separating paste, with a catalyst, for use in dental laboratory

Salient characteristics:

The dental laboratory paste shall be a silicone insulating and separating paste, with a catalyst, for use in dental laboratory

The paste shall be capable of protecting both porcelain and plastic porsthetic teeth from the adherance of investment material so as to provide clean, well-finished gingival and interproximal areas in denture construction. It shall be easily mixed within 30 seconds, allowing a minimum of 3 minutes working time, and cure to a flexible coating so as to allow immediate investment application and subsequent boil-out within 30 minutes of pouring. The values for mixing, working time and boil-out shall be based on the use of the paste at 23 $\pm$ 2. 0 degrees Celsius (^{o}C) and 50 $\pm$ 10 percent relative humidity.

The shelf life of the material shall be not less than 24 months from date of manufacture when stored at normal room temperative (59°-86°F). The inspection test date shall be 24 months from date of manufacture. Not less than 20 months of the expiration dating period shall remain at time of delivery to the Government.

Workmanship. The paste shall be clean and free from defects which may detract from its appearance or impair its serviceability.

Unit. Each (EA). One 10 ounce tube of paste with sufficient catalyst, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the currentversion before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one roll of aprons, as specified. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

Stock Number	Item Identification
6520-00-275-0858	PASTE, DENTAL LABORATORY, Silicone Insulating and Separating, 10 Oz. (283.5 grams): with catalyst.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITY
VA-OSS
PHS

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA - DM

PROJECT NO. 6520 -1906

Location: NCR DISK 2

59040

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44946

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the currentversion before use.