COMMERCIAL ITEM DESCRIPTION

CUP, MICRO SAMPLE, AUTOMATED ANALYSIS
(Conical, Polystyrene, Disposable)

The General Services Administration has authorized the use of this commercial item description in preference to Military Specification MIL-B-37405, dated 24 May 1976.

This Commercial Item Description covers a disposable polystyrene micro sample cup for use with automatic analyzers.

Salient characteristics:

Shall be a disposable polystyrene micro sample cup with a conical interior, for use with automatic analyzers such as the Technicon Instruments A-1 and A-2 blood typing and two channel flame photometer auto-analyzers.

Shall be fabricated of optically clear polystyrene plastic conforming to ASTM D703 (Polystyrene Molding and Extrusion Materials).

Shall have a volumetric capacity of 2.0 milliliters (ml) $\pm$ 0.2 ml.

Shall be essentially cylindrical on the outside, and shall include a flange projection toward the top. The flange shall support the micro sample cup so that it does not fall through a cutout in a sampler plate of an analyzer.

Shall include a conical bottom design on the inside of the cylinder. The bottom design shall be leakproof.

Shall have a minimum outside flange diameter of 0.595 inches. The outside diameter under the flange shall not exceed 0.565 inches.

Shall not wobble when standing on a flat horizontal surface.

AMSC N/A FSC 6630

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the current version before use.

Workmanship. The micro sample cup shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing one thousand cups, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specifications where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Sampling for inspections and tests. Sampling for inspections and test shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6630-01-009-3762

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS
FHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD-MB

Agent

DLA-DM

Project No. 6630-0445

Location: NCR Disk #3

U.S. GOVERNMENT PRINTING OFFICE: 1987 - 704-034/70298

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.