INCH-POUND

A-A-51630A
12 March 1993
SUPERSEDING
A-A-51630
25 August 1986

COMMERCIAL ITEM DESCRIPTION

BLOOD SAMPLE, COLLECTING SET

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a disposable blood sample collecting set to be used in "prep" for, obtaining and storage of arterial blood samples.

Salient characteristics:

Shall be a disposable arterial blood sampling kit contained in a snap-open polyethylene bag consisting of the following:

a. One 5 ml polypropylene syringe with sufficient sodium heparin United States Pharmacopeia (USP) to anticoagulate 5 ml of whole blood. The plunger shall have positive stops for both compressed and filled positions with a rubber tip for sealing. A polypropylene fitting shall be provided to seal the syringe tip luer fitting.

b. One sterile 22 gauge by 1 inch clear or color coded plastic hubbed needle with a short bevel. The needle cannula shall be fabricated from Class 3, Type 304 corrosion-resisting steel. The hub shall be fabricated of polypropylene. The needle shall be contained in a polypropylene container to protect the point and maintain sterility.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

c. One sterile 20 gauge by 1-1/2 inch clear or color coded plastic hubbed needle with a short bevel. The needle cannula shall be fabricated from Class 3, Type 304 corrosion-resisting steel. The hub shall be fabricated from polypropylene. The needle shall be contained in a polypropylene container to protect the point and maintain sterility.

d. One self-venting rubber stopper tip for use either on syringe or on needle. The stopper tip shall be 3/4 inch long by 1/2 inch in diameter.

e. One sterile iodophor wipe with dimensions of at least 1-1/4 inches by 1-1/2 inches. The wipe shall contain 1 percent polyvamer iodine complex titratable iodine. The wipe shall be contained in a foil package.

f. Two sterile, 8 ply minimum gauze pads with dimensions of 2 inches by 2 inches. The pads shall conform to USP Type VII gauze.

g. One sterile alcohol wipe with dimensions of at least 1-1/4 inches by 1-1/2 inches containing 70 percent isopropyl alcohol. The wipe shall be contained in a foil package.

h. One patient label with dimensions of 1-1/2 inches by 2-1/2 inches indicating patient name, patient temperature, room number, date, test, time and ventilation. The label shall have a peel-off self-adhering adhesive back. As an alternate, label shall be 1-1/4 inches by 1-5/8 inches in dimensions.

i. Ice bag shall be a zip lock 0.003 inch polyethylene bag with dimensions of 6 inches wide by 8-1/2 inches high (minimum). The snap-open bag shall be 6 inches by 6-1/2 inches and shall contain the kit components less the ice bag. The filled snap-open bag shall be stored in the ice bag. As as alternate, ice bag shall be 5 inches by 9-5/8 inches and the snap open bag shall be 5-1/4 inches by 11-1/2 inches minimum.

Materials. Corrosion-resisting steel shall be in accordance with ASTM Designation F899.

Luer fittings. Luer fittings shall be in accordance with ANSI/HIMA MD70.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Sterility. The sterile components shall meet the sterility requirements when tested in accordance with the USP or AAMI guidelines. Sterility assurance level shall be 0.0001.

Pull test. The needle cannula shall not pull out of, or loosen from, the needle hub when subjected to an axial force of ten pounds for a period of 3 minutes. Both needles shall meet this test requirement

Workmanship. The set components shall be free from defects which detract from their appearance or impair their serviceability. Plastic components shall be free from dents, kinks, pits, blisters, flash, cuts, holes and scratches. Metal components shall be free from dents, pits, burrs, slivers, scale and corrosion. Needles shall not be bent, shall be sharp and shall be free from feathered edges.

Unit. Package (PG). One package containing one hundred collecting sets, each set contained in a snap-open polybag as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for examination purposes shall be one as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one blood sample collecting set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-360-2131	BLOOD SAMPLE, COLLECTING SET, Arterial Disposable, 100's

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

A-A-51630A

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MP

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-5535
Location: c:\word\doc\mish\blood2cd

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: A-A-54828	2. DOCUMENT DATE (YYMMDD): 11 March 1997

3. DOCUMENT TITLE: SYRINGE, HYPODERMIC, Angiographic Injector

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME Defense Personnel Support Center ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: (215)7372870 (2) AUTOVON: 444-2870
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete — 198-290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.