A-A-0051633(DM)
20 August 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

DISPENSER, SENSITIVITY DISCS

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This commercial item description covers a twelve (12) place dispenser for BBL brand sensitivity discs, for use in bacteriological sensitivity testing by the Kirby Bauer method.

Salient characteristics:

Shall be a twelve(12) place dispenser for dispensing BBL brand impregnated paper discs (contained in dispensing cartridges) onto a 150 millimeter (mm) diameter agar surface.

Shall be suitable for accomodating stock listed BBL dispensing cartridges containing the following types of paper susceptibility test discs:

National Stock Number	Susceptibility test disc impregnation (micrograms or units)
6550-01-031-6158	10 mcg of Ampicillin
6550-01-031-6160	100 mcg of Carbencillin
6550-01-031-6161	30 mcg of Cephalothin
6550-01-031-6162	30 mcg of Chloromycetin
6550-01-031-6163	15 mcg of Erythromycin
6550-01-031-6166	10 mcg of Gentamicin

AMSC/NA FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the current version before use.

National Stock Number	Susceptibility test disc impregnation (micrograms or units)
6550-01-031-6167	30 mcg of Kanamycin
6550-01-031-6169	5 mcg of Methicillin
6550-01-031-6170	10 units of Penicillin G
6550-01-031-6171	300 units of Polymyxin B
6550-01-031-6173	30 mcg of Tetracycline
6550-01-113-0111	5 mcg of Chloramphenicol
6550-01-113-0113	30 mcg of Nalidixic Acid

Shall have twelve (12) numbered cartridge holes that are spaced 33 mm apart between disc centers.

Shall be designed to simultaneously dispense 12 paper discs from a preset height, and with self-tamping (ie. without the need for follow-up manual tamping), onto an agar surface. The dispensing shall be the result of a plunger type dispensing action.

Shall be supplied with a moisture tight storage case and desiccant.

Shall be supplied with instructions.

Workmanship. The dispenser shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One dispenser, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-01-139-7478

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

PGC

Project No. 6640-1405

Location: CID/DIS,SEN,DISC
NCR/disk 7

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 605-036/44948

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the currentversion before use.