A-A-51635
26 AUGUST 1986

COMMERCIAL ITEM DESCRIPTION

TUBE, DUODENAL, SURGICAL, Levin, Rubber

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers three sizes of rubber surgical tubes of the Levin design suitable for use in suction and irrigation of the duodenum.

Salient characteristics:

Style and design. The tube shall have the dimensions shown in Table 1 for the size specified. The proximal end shall be flared to accommodate a suction adapter or an irrigation syringe. The distal end shall be closed and rounded. The tube shall have two sets, of two eyes each, staggered on diametrically opposite sides of the distal end. The distal-most eye shall be 1 cm + 0.3, - 0 cm from the distal tip. The distal-most eye of the opposite set shall be placed on the opposite side of the tube at the midpoint of the first set. The eyes shall have a cross-sectional area equal to or greater than that of the lumen of the tube, but shall not exceed that of the lumen by more than 20 percent. The edges of the eyes shall be chamfered or bevelled.

TABLE I. Tube Characteristics

TABLE 1.

Fr. size	Inside diameter (cm)	Length (cm)	Eye-to-eye separation of set (cm)	Breaking strength (pounds)	Ultimate elongation (percent)
12	0.15 ± 0.02	122 ± 2.0	3.0 ± 0.3	42	400
14	0.18 ± 0.02	122 ± 2.0	5.0 ± 0.5	56	400
16	0.25 ± 0.02	122 ± 2.0	5.8 ± 0.5	67	400

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Graduations. The tube shall have four bands at 10.2 ± 0.1 cm intervals. The first band shall be 45.7 ± 0.1 cm from the distal tip of the tube for the 12 French size and shall be 48.3 ± 0.1 cm from the distal tip for the 14 and 16 French sizes.

Material. The tube shall be manufactured from compounded rubber which contains a minimum of 50 percent natural rubber.

Markings. The tube shall be permanently marked with the manufacturer's name or registered trademark and the French size. The markings shall be near the proximal end.

Physical Properties. Tubes shall be tested for breaking strength and ultimate elongation using Test Method Numbers 4111 and 4121 respectively of Federal Test Method Standard 601. The value of breaking force, as defined in Test Method 4111, shall be used as the value for breaking strength. Breaking strength and ultimate elongation values shall be equal to or greater than the values listed in Table 1 for the size tested. In addition, specimens cut from the same tubes shall be subjected to accelerated aging as specified in Test Method 7221 of Federal Test Method Standard 601, and the tests for breaking strength and ultimate elongation performed upon the aged specimens. The values obtained from the aged specimens shall not be less than 75 percent of the values obtained from the unaged specimens cut from the same tube.

Biocompatibility. Rubber material shall be tested in accordance with the Intracutaneous Test and Intramuscular Implantation Test of the United States Pharmacopeia (USP). A sufficient amount of material shall be tested as to be representative of the lot or batch.

Workmanship. The exterior and lumen of the tube shall be free from cracks, blisters, embedded particles, and foreign matter. There shall be no defects which detract from the appearance of or impair the serviceability of the tube.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One tube, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for examination purposes shall be one tube. Sampling shall be Level II and an AQL of 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one tube. Sampling shall be S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-385-9800	TUBE, DUODENAL, SURGICAL, Levin, 12 French
6515-00-385-9840	TUBE, DUODENAL, SURGICAL, Levin, 14 French
6515-00-385-9880	TUBE, DUODENAL, SURGICAL, Levin, 16 French

Ordering data. (Intermediate/exterior package quantities and labeling and marking) must be specified in the contract and/or order).

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:58Z
Check the source to verify that this is the currentversion before use.

COMMERCIAL ITEM DESCRIPTION A-A-51635

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA-MPQAS

05322

Project No. 6515-3552

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44839

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:58Z
Check the source to verify that this is the currentversion before use.