A-A-0051636(DM)
22 August 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

SYRINGE, FLUID DISPENSING, 10 ML

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a spring loaded, 10 ml fluid dispensing syringe designed for dispensing of intravenous additives and unit dose fluids in the pharmacy.

Salient characteristics.

General. Shall be a sterile, disposable, non-pyrogenic, spring loaded 10 ml syringe designed to supply a preset amount of fluid.

Material. The plastic tubing shall be clear, colorless, transparent, flexible polyvinyl chloride. The syringe shall be fabricated from clear, rigid plastic. The plunger tip shall be fabricated from rubber. The spike, clamp, valve, Luer adapters and end protectors shall be made of a suitable rigid plastic. The plastic components coming in contact with the fluid shall meet the requirements of the Pharmacopeia of the United States for Class IV plastics (USP). Those rubber and plastic components coming in contact with the fluid and end protectors shall contain not more than 5 parts per million (ppm) total lead, barium and cadmium when tested in accordance with the physico-chemical tests in USP.

Characteristics. The fluid dispensing syringe shall be a spring loaded 10 ml syringe with subgraduations every 1/5 ml. Each whole number graduation shall be numbered and be darker and longer than the subgraduations. The scale shall be legible and readable with the syringe held in the vertical position. The syringe shall have a handle. The syringe shall have a Luer lock tip.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

The plunger tip shall be lubricated with pharmaceutical grade silicone lubricant. The plunger shall make a good seal between the barrel and plunger such that there shall be no leakage past the plunger in either direction.

The plunger tip shall have a minimum of two rings making contact with the barrel. The distal tip ring shall accurately indicate the syringe fluid volume on the barrel scale. The plunger tip shall be of a color which clearly contrasts with the syringe barrel for maximum visibility. The proximal end of the plunger shall have a thumb grip.

Each syringe shall be supplied with 36 inches $\pm$ 4 inches of tubing. A vented conical spike with a bacterial filter shall be permanently attached to the distal end of the tubing. The tubing shall have a hardness of 68 +6, -5 Shore A durometer. A two-way valve shall be permanently attached in a leakproof manner to the tubing. The two-way valve shall prevent liquid from entering the tubing when the syringe is being emptied and prevent air from entering through the needle (needle not provided) when the syringe is drawing liquid from a container via the tubing. The distal end of the two-way valve shall have a female Luer lock adapter. The proximal end of the two-way valve shall be a male Luer lock adapter. An on-off clamp shall be provided on the tubing.

End protectors shall be provided on the syringe outlet, two-way valve inlet and outlet and the spike. Each end protector shall be readily removable. When placed over each open end it shall, after sterilization, maintain sterility and integrity of the entire fluid path of the device until manually removed.

Marking. Syringes shall bear the manufacturer's name or registered trademark and the capacity in milliliters (ml) or cubic centimeters (cc).

Luer fittings. The Luer fittings shall meet the dimensional and performance requirements of ANSI/HIMA MD70.1.

Instructions. An instruction sheet detailing the proper use of the syringe shall be provided with each device.

Workmanship. The syringe set shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Package (PG). One package containing twelve syringes, each to be individually packaged sterile, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for visual examination shall be in accordance with MIL-STD-105. The unit of product for examination shall be one syringe. Sampling shall be Level II with an AQL of 6.5

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one fluid dispensing syringe. AQL shall be 1.0 percent defective. Inspection level shall be S-2.

Accuracy for graduations. The graduated scale shall be accurate to within ± 5 percent of the volume indicated on the scale. Syringes shall be tested for volumetric accuracy using a calibrated burette.

Sterility and pyrogenicity. Syringes shall be tested for sterility and freedom from pyrogens in accordance with the USP.

Biocompatibility. Plastic material shall be tested in accordance with the USP tests for Class IV plastics. Rubber material shall be tested in accordance with the Intramuscular and Intracutaneous sensitivity tests of the USP.

Certificates of biocompatibility. A certificate of compliance, stipulating that offered items meet the biocompatibility test, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years and that there have been no changes to material formulation and method of processing since actual testing was performed. The certificate shall include the date the testing was performed, the test method and results, the name and title of the person evaluating the results, the current contract number and NSN, a statement certifying that material processing and formulation have not changed since the time that biocompatibility was performed and the signature and title of the person making the certification.

Tubing hardness. Plastic tubing hardness shall be determined in accordance with ASTM D2240 Standard Test Method for Rubber Property-Durometer Hardness. The test shall be conducted at a room temperature of 27° plus or minus 5°C (80° plus or minus 9°F) on samples which had been conditioned at this temperature for not less then 3 hours prior to testing. The test sample shall be laminated to a thickness of not less than 1/4 inch. The gage reading shall be made 15 seconds after application of the instrument.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.

Accelerated aging test. Preheat an air oven to 158°F ± 1.8°F prior to placing the syringe into the oven. The syringe shall stay in the oven for seven days. The appearance of the syringe before and after the aging test shall be noted. Any cracking or other deterioration shall constitute failure of the test. Attach the syringe to an intravenous container filled with water and operate the syringe. If the syringe leaks or does not deliver fluid properly, it shall be rejected.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-140-5357	SYRINGE, FLUID DISPENSING, 10 ML

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 59195

Project No. 6515-3853

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44949

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.