A-A-51643
3 September 1986

COMMERCIAL ITEM DESCRIPTION

ADAPTER, RIGHT ANGLE ELBOW, TRACHEAL ANESTHESIA MASK

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers conductive, neoprene tracheal right-angle mask adapters suitable for use during standard anesthesia procedures.

Salient characteristics:

General. The right-angle elbow-adapter shall be conductive and suitable for use in standard anesthesia procedures involving medicinal gases.

Characteristics. The right-angle elbow-adapter shall not permanently deform or be adversely effected by anesthesia agents or gases used in standard anesthesia processes. The distal or patient end of the adapter shall possess a standard tapered 22 mm outside diameter and shoulder designed to provide a secure fit when attached to an anesthesia mask. The taper shall be in accordance with CGA Pamphlet M-1. The inside diameter of the patient end of the adapter shall be 15 mm female connection with a basic taper in accordance with ANSI Z79.2.

The proximal or machine end shall be a tapered 15 mm male in accordance with ANSI Z79.2. This connection shall be designed to connect to anesthesia machines in accordance with ANSI Z79.9.

The adapter shall have an adjustable spring-loaded relief valve on top of the elbow curve.

Material. Each adapter shall be fabricated from conductive neoprene. The relief valve components shall be fabricated from corrosion-resistant material. External components of each valve shall be fabricated from chromium plated brass.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.

Accelerated aging test. Preheat an air oven to 158°FR ± 3.6°F (70°C ± 2°C) prior to placing sample in the oven. The adapter shall be placed in the air oven for seven days. The appearance of the adapter after completion of the aging test shall be noted. The valve components shall not separate from the neoprene as a result of the aging test. Durometer hardness and tensile strength shall not decrease more than 10 percent due to the aging test. Both durometer hardness and tensile strength valves for the adapter shall be established prior to performing the accelerated aging test. Durometer hardness shall be determined in accordance with ASTM D2240. Tensile strength shall be determined in accordance with ASTM D412.

Sterilization. All adapters shall be designed to withhstand repeated steam sterilization at 270°F (136°C). The adapter shall be placed in a steam sterilizer and sterilized at 270°F for 3 minutes. The adapter shall then be removed from the sterilizer and allowed to cool to room temperature. This test shall be performed 10 times. The adapter shall show no signs of damage, separation, corrosion or permanent deformation.

Conductivity. The right-angle elbow-adapter shall have a maximum resistance of 1,000,000 ohms when tested in accordance with NFPA No. 99. Each adapter shall comply with the requirements of NFPA No. 99, Chapter 3, Use of Inhalation Anesthetics (Flammable and Non-Flammable).

Workmanship. The right-angle adapter shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (Ea). One adapter, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. AQL shall be 1.0 percent defective including tests specified in CGA and ANSI. Inspection level shall be S-2. The unit of product for test purposes shall be one right-angle elbow-adapter.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-153-5920	ADAPTER, RIGHT-ANGLE ELBOW, Tracheal Anesthesia Mask

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51025

Project No. 6515-3504

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/45083

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.