A-A-51645
10 September 1986

COMMERCIAL ITEM DESCRIPTION

DISH, MOISTURE DETERMINATION, (Aluminum)

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-D-365a.

This commercial item description covers a moisture determination dish with overlapping cover for use in making official moisture determinations.

Salient characteristics:

The moisture determination dish shall be formed of heavy gauge sheet aluminum 0.025 + 0.0025 inch thick conforming to H14 of Federal Specification QQ-A-250/1c.

Each dish shall be cylindrical in shape and shall consist of a bottom section and a close fitting overlapping cover.

The bottom section shall have straight sides with the same diameter top and bottom and shall set flat on a plane and level surface.

The cover shall be flat with straight sides dimensioned to fit closely over the sides of the bottom section with an overlap of 5 ± 2 millimeters (mm).

The overall dimensions of the complete dish including the cover shall be 90 ± 2 millimeters (mm) diameter and 51 ± 2 mm in height.

The dishes shall be given a nondyed anodic coating in accordance with finish No. 7.1.1 of MIL-STD-171 or an alkaline dip.

Reference specifications. Where Federal Specifications or Military Standards are referenced herein, the issue of that specification or standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Workmanship. The dish shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One dish with cover, as specified, constitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one dish with cover.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the currenversion before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6640-00-811-4431

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1400
Location: CID/DISH, MOISTURE DETERMINATION/NCR disk, 6

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 — 605-036/45084

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.