A-A-51646
10 September 1986

COMMERCIAL ITEM DESCRIPTION

SPLINT, UNIVERSAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a radiolucent, malleable lightweight aluminum, universal splint for immobilization of injured extremities.

Salient characteristics:

Shall be a malleable lightweight aluminum splint covered on both sides with foam, suitable for immobilization of injured extremities. The foam should overlap all edges of the aluminum by 1/8 inch, so that no edges of the aluminum are exposed.

The splint shall be approximately 4-1/4 inches (10.8 cm) wide by 36 inches (92.2 cm) long by 3/16 inch (0.48 cm) thick.

The aluminum shall be a dead soft alloy, a maximum of 0.020 inch (0.051 cm) in thickness, that is radiolucent to permit X-ray of the injured extremity without removing the splint.

The foam covering shall be approximately 1/8 inch (0.32 cm) in thickness on one side of the aluminum and 1/16 inch (0.16 cm) on the opposite side. Shall be of a closed cell material that is non-absorbent of blood or other liquids, and which can be easily cleaned with a variety of contact cleaners.

The splint shall be capable of being cut by ordinary scissors.

The color shall be gray for field use.

Individual splints shall be enclosed in a shrink wrap with directions for application enclosed on a sheet of paper rolled into the splint.

Workmanship. The splint shall be free from defects which detract from its appearance or impair its serviceability.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing twelve individually packaged splints, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for inspections and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one splint.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-225-4681	SPLINT, UNIVERSAL, Malleable Lightweight Aluminum

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3720
Location: CID/SPLINT, UNIVERSAL, Malleable Lightweight Aluminum/NCR disk, 6

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/45085

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.