A-A-51647
24 April 1987

COMMERCIAL ITEM DESCRIPTION

TRAY, ENDOSCOPIC INSTRUMENT STERILIZATION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a cold sterilization and disinfection tray for use with flexible fiberoptic endoscopes or other instruments.

Salient characteristics:

The plastic tray shall be suitable for use in the cold sterilization and disinfection of endoscopic instruments, surgical instruments, and other equipment.

The tray shall have an approximate diameter of 19 inches (48.32 cm) with a depth of approximately 6 inches (15.2 cm) with a suitable lid.

The tray and lid shall be constructed of Acrylonitrile Butadiene Styrene (ABS) which shall be non-reactive with disinfectants and detergents. The tray and lid shall each be a one piece construction capable of withstanding cold sterilization without any signs of crazing, cracking, distortion, or other visible deterioration.

The tray lid shall rest evenly, without rock, on the tray. The tray shall rest evenly on a flat surface without rock or wobble.

The tray and lid shall be free of cracks, tears, holes, or ripples. All surfaces and edges shall be smooth and free from burrs or sharp edges. The finish shall be free of scratches or mars. The tray and lid shall be free of grease, oil, dirt, chips, or other foreign materials.

Workmanship. The tray shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One tray, as specified, constitutes one unit.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

APPROVED: AJM/MPM/LS

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

TABLE I. Sampling for examination.

	Inspection Level	AQL (defects per hundred units
For visual examination		
Major defects	II	1.0
Total (major and minor defects combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The trays shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed on Table II. Examination shall not be restricted to the classified possible defects listed below.

TABLE II. Classification of defects.

Categories	Defects
Major	
101	Tray not free of cracks, tears, holes, ripples.
102	Edges not free of burrs or sharpness.
103	Tray rocks, wobbles, or does not rest evenly on a flat surface.
104	Cover does not rest evenly on the tray body.
Minor	
201	Finish not free of scratches or mars.
202	Tray not free of grease, oil, dirt, chips, or other foreign matter.

Dimensional examination. The trays shall be examined for defects in dimensions. Any dimension not within tolerances specified herein shall be classified as a defect.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Tray shall be filled with suitable disinfectant, used for cold sterilization for 30 minutes. At the end of this period the tray shall be drained and allowed to dry. The cycle shall be repeated five (5) times. The tray shall be examined for defects specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

First Article Inspection. First article inspection shall be performed on the sterilization tray when a first article is required. This inspection shall include the examination and tests.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended and as implemented by federal acquisition regulation subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-173-2040

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1882

PGC 51086

Location: CID/T-ENDOSCOPIC/H13-5

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51647	Tray, Endoscopic Instrument Sterilization

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil/ – Downloaded: 2026-Apr-26 08:08:59Z
Check the source to verify that this is the current version before use.