A-A-51648
24 April 1987

COMMERCIAL ITEM DESCRIPTION

TRAY SET, ENDOSCOPIC INSTRUMENT STERILIZATION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a cold sterilization and disinfection tray set for use with flexible fiberoptic endoscopes or other instruments.

Salient characteristics:

The plastic tray set shall be suitable for use in the cold sterilization and disinfection of endoscopic instruments, surgical instruments, and other equipment.

The tray set shall consist of two soaking and rinsing trays approximately 26 inches (66.1 cm) long by 5 inches (12.7 cm) wide by 5 inches (12.7 cm) deep, one perforated inner liner, and one lid.

The tray set shall be constructed of a suitable plastic material which shall be non-reactive with disinfectants and detergents. The trays, liner, and lid shall be a one piece construction capable of withstanding cold sterilization without any signs of crazing, cracking, distortion, or other visible deterioration.

The tray set shall have the capability of either steam or gas (ethylene oxide) sterilization between usage in order that a sterile field be created for each use.

The tray lid and liner shall rest evenly without rock in the body of the tray. The tray shall rest evenly on a flat surface without rock or wobble.

All components of the tray set shall be free of cracks, tears, holes, or ripples. All surfaces and edges shall be smooth and free from burrs or sharp edges. Finish shall be free of scratches or mars. All components shall be free of grease, oil, dirt, chips, or other foreign materials.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Workmanship. The tray set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One tray set, as specified, constitutes one unit.

TABLE I. Sampling for examination.

	Inspection Level	AQL (defects per hundred units
For visual examination		
Major defects	II	1.0
Total (major and minor defects combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The trays shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed on Table II. Examination shall not be restricted to the classified possible defects listed below.

TABLE II. Classification of defects.

Categories	Defects
Major	
101	Tray not free of cracks, tears, holes, ripples.
102	Edges not free of burrs or sharpness.
103	Tray rocks, wobbles, or does not rest evenly on a flat surface.
104	Cover does not rest evenly on the tray body.
Minor	
201	Finish not free of scratches or mars.
202	Tray not free of grease, oil, dirt, chips, or other foreign matter.

Dimensional examination. The trays shall be examined for defects in dimensions. Any dimension not within tolerances specified herein shall be classified as a defect.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Tray shall be filled with suitable disinfectant, used for cold sterilization for 1 hour. At the end of this period the tray shall be drained and allowed to dry. The cycle shall be repeated twenty (20) times. The tray shall be examined for defects specified herein.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

First article inspection. First article inspection shall be performed on the tray when a first article sample is required. This inspection shall include tray set the examination and tests.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one tray set.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-173-2041

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1883

PGC 51118

Location: CID/TRAY/H13-5

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A -51648	Tray Set, Endoscopic Intrument Sterilization

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for submitting comments and suggestions for improvements All users of military standardization documents are invited to pr suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)* mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretatio too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorizat to waive any portion of the referenced document*(s)* or to amend contractual requirements

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO PO
NECES
IF MA
IN T
UNITED

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.