INTERIM COMMERCIAL ITEM DESCRIPTION

TUBE, TRACHEOSTOMY, DISPOSABLE, SHILEY

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a sterile disposable, soft plastic, low pressure cuff tracheostomy tube, Shiley style, suitable for long term ventilation of tracheostomy patients who do not require high ventilation pressures.

Salient characteristics.

Construction. The tracheostomy tube shall consist of an outer cannula with tubing connecting to a pilot balloon, an inner cannula and an obturator.

Material. The outer cannula shall be fabricated of semi-rigid polyvinylchloride with the cuff, tubing and pilot balloon fabricated of polyvinylchloride. The inner cannula shall be fabricated from polypropylene and the obturator and flange from polycarbonate. All material shall be medical grade.

Outer cannula. The outer cannula shall have a soft thin walled, high volume, low pressure cuff that uniformly distributes pressure against the tracheal wall. Shall have a one-way Luer valve pressure retention system and a swivel neck plate for comfort and conformity to neck anatomy with the size clearly imprinted on the front outer surface of the neck plate. The outer cannula shall be radiopaque. The cuff shall be inflated through plastic tubing connected to a pilot balloon containing the one-way Luer valve. The cuff shall be designed to reach tracheal occlusion at internal cuff pressures of less than 25 mm Hg. The cannula shall be suitably curved to allow access to the trachea.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Inner cannula. The tracheostomy tube shall have a removable inner cannula with a twist lock, integral 15 mm connector. The inner cannula shall be easily removed and cleaned to maintain a patent airway without removing the outer tube. Inner cannula proximal end design shall conform to American National Standard Specification ANSI Z79.2.

Obturator. Item shall be provided with a smooth tapered obturator to simplify insertion and minimize trauma.

Item shall meet sterility requirements when tested in accordance with the USP. Shall be non-toxic and meet the requirements for Class II plastics as stated in the current USP.

Workmanship. The tracheostomy tube shall be free from defects which detract from its appearance or impair its serviceability. Components shall be free from cuts, tears, weak sections, flash, dust, grease or other foreign matter.

Unit. Each (EA). Each tube packaged sterile in a sealed peel-open type container, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one tracheostomy tube.

Pressure test. The cuff, tubing, one-way valve and pilot balloon shall not burst, sustain any damage or leak when subjected to a pressure of 1.5 psi for 5 minutes. Inspection test level shall be S-3 with an AQL of 0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-147-0204	TUBE, TRACHEOSTOMY, Disposable, Shiley

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 50068

Project No. 6515-3858

U.S. GOVERNMENT PRINTING OFFICE: 1986 - 704-036/45596

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.