A-A-51656
19 September 1986

COMMERCIAL ITEM DESCRIPTION

ELECTROSURGICAL APPARATUS, (PORTABLE)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a portable electrosurgical apparatus.

Salient characteristics:

The system shall be a portable electrosurgical apparatus suitable for fulguration, and coagulation with cutting during minor surgical procedure treatments in the room and clinics. The system shall have the capacity for graduated power increments, monopolar and/or bipolar, and shall be actuated by a footswitch or handswitch handle.

The electrical unit shall be approximately 10 inches by 5-1/2 inches by 4 inches (25.4 cm by 13.97 cm by 10.16 cm) with a weight of approximately 6 pounds (2.7 kg). The package shall consist of the following:

a. One (1) each solid state electrosurgical apparatus.
b. One (1) each switching handle and cord.
c. One (1) each needle electrode, 1/2 inch (1.27 cm) long.
d. One (1) each straight tissue desiccation needle, 4 inches (10.16 cm) long.
e. One (1) each mounting plate with screws.

The system shall be drip proof.

The pencil for the system shall be watertight easy to grip and lightweight, suitable for close work, and slip-resistant even when wet. The pencil shall be a seamless one piece handle and have no burrs or sharp edges to puncture surgical gloves.

The system shall have a power on/off switch to control all primary input power to the unit. The power output for bipolar shall be 30 +/- 5 watts (125 - 500 ohm load), high (23 +/- 5 watts/500 ohm load), and low (20 +/- 5 watts/ 500 ohm load). The power output for monopolar applications shall peak at 1000 ohm load impedances, and operate between 100 and 5000 ohms.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release., distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

The electrosurgical apparatus shall operate on 115 volts, 50/60 Hz, AC 125 watts. It shall have a 6-foot (1.8 m) three wire grounded hospital grade line cord permanently attached with molded plug. The unit shall be a high frequency generator with solid state circuitry to produce a damped sinusoidal output wave. The housing of the unit shall be fabricated of a high impact strength, high di-electric, self-extinguishing thermoplastic or other suitable material. The electrosurgical apparatus shall have two (2) panel lights: one to indicate when input power is activated and one to indicate output circuit is activated. It shall have a dial which serves as the on-off switch and that is also the output selection dial from 0-100 in graduated current intensity increments of five. The unit shall have a power output selection of high, low, and bipolar current for the switching handle. The electrosurgical apparatus shall have an adjustable, audible signal which activates during use of the switching handle.

The switching handle shall be fabricated of high impact strength thermoplastic, approximately 6 to 8 inches (15.24 to 20.32 cm) in length, have fingertip on-off control, and be capable of accepting all electrodes with a shank diameter of 3/32 length. The handle and cord shall be completely drip proof when electrode is installed and be capable of withstanding repeated sterilization processes with radiation (ETO).

Safety. All items shall conform to the applicable requirements of a nationally recognized standard for the safety of medical equipment, of the issue in effect on date of solicitation. All items shall be approved as conforming to such a standard by an independent nationally recognized testing laboratory, as applicable.

Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts and accessories available for the apparatus. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 2 copies of the approved technical manual to the Contracting Officer for distribution. Should the

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

supplier have furnished acceptable manuals for the identical item within two years preceding tne date of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of the drafts and submission of 2 additional copies may be waived upon requests to the Contracting Officer. Waiver of submission of drafts and submission of 2 additional copies shall be at the option of the Government.

Workmanship. The portable electrosurgical apparatus shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One electrosurgical apparatus, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this document:

6515-01-147-4667

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS.

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3459

Location CID/APPARATUS/HlD-10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51656	ELECTROSURGICAL APPARATUS, (PORTABLE)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)*: _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.