A-A-51657A
25 October 1989
SUPERSEDING
A-A-51657
19 September 1986

COMMERCIAL ITEM DESCRIPTION

TRAY, INSTRUMENT,

(Corrosion-Resisting Steel)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a perforated corrosion-resisting steel instrument tray with handles suitable for use in sterilizing instruments in standard sterilizers.

Salient characteristics:

The instrument tray shall be suitable for use in sterilizing instruments in standard sterilizers. The tray shall be fabricated of American Iron and Steel Institute (AISI) type 304 corrosion-resisting steel, a minimum of 0.031 inch (No. 22 United States Standard Gage) thick, properly tempered and passivated. The tray shall be furnished in the following sizes:

- Size 1 - 15 inches plus or minus 1/4 inch long by 10-1/2 inches plus or minus 1/4 inch wide by 3-1/2 inches plus or minus 1/4 inch deep.

  Size 2 - 20 inches plus or minus 1/4 inch long by 10-1/2 inches plus or minus 1/4 inch wide by 3-1/2 inches plus or minus 1/4 inch deep.

  Size 3 - 10 inches plus or minus 1/4 inch long by 6-1/2 inches plus or minus 1/4 inch wide by 2-1/2 inches plus or minus 1/4 inch deep.

  Size 4 - 10 inches plus or minus 1/4 inch long by 10-1/2 inches plus or minus 1/4 inch wide by 3-1/2 inches plus or minus 1/4 inch deep.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

The bottom surface of the tray shall be made of perforated metal to assure steam penetration and circulation with minimal condensation. The metal shall have perforations approximately 1/8 inch in diameter, a minimum of 20 holes per square inch. The sides of the tray need not have holes. The tray bottom may be recessed to aid in manufacturing if necessary.

The top of the tray shall be either hemmed or beaded to prevent injury to users. The radius of the tray corners shall be 1/2 inch or less. The corners shall be smooth without cracks or deformations.

The tray shall have 2 handles made of approximately 3/16 inch diameter AISI Series 300 corrosion-resisting steel rod. The handles shall be approximately 3-3/4 inches wide, suitable for lifting the tray with 4 fingers, and shall be located opposite each other on the sides of the tray with the smaller dimensions.

The handles shall swivel from a fully down to a fully up position, and shall swivel inside a retainer securely attached to the tray body by means of a minimum of four spot welds. The handles shall be capable of being used with extension handles. In the down position, the bottom of the handles shall be slightly bent outward to facilitate grasping.

The tray shall rest evenly on a flat surface without rock or wobble.

The tray shall be free of cracks, tears, holes, or ripples. The edges shall be free of burrs or sharpness. The tray finish shall be a No. 4 commercial finish, or better, and be reasonably free from scratches or mars.

The tray shall be free of dirt, grease, graphite, metal particles, or other foreign material.

Workmanship. The instrument tray shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One tray, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests (dimensional), shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one instrument tray.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-01-166-9036 (Size 1)
6530-01-167-6681 (Size 2)
6530-01-168-0111 (Size 3)
6530-01-173-2085 (Size 4)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-2095

Location:
CID/TRAY/INSTRU/COR/
RES/STL/H13-35

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
-51657A	TRAY, INSTRUMENT (Corrosion-Resisting Steel)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street, City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4956 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC- RSTE
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.