NNN-B-001636 (DSA-DM)
AMENDMENT -1
October 21, 1970

AMENDMENT

TO

INTERIM FEDERAL SPECIFICATION

BURNER, ALCOHOL, DENTAL

This amendment was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this amendment as a valid exception to Interim Federal Specification NNN-B-001636 (DSA-DM).

Page 1:

2.1 Specifications, standards and drawings. Delete "invitation for bids or request for proposal" and substitute "invitations for bid or requests for proposal".

Under Federal Specifications, Add:

"QQ-C-320 Chromium plating (Electro deposited)".

Delete "QQ-B-663" and substitute "QQ-B-613".

Page 2:

Under "STANDARDS" insert:

"FED-STD-123 Marking for Domestic Shipment (Civilian Agencies)".

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

Page 3:

3.2 Construction. At the end of paragraph, add: "All metal components shall be coated with a uniform thickness of chromium over nickel. Outer surfaces of the burners shall be polished to a bright finish. Finished thickness (inches) of the following component brass parts shall be as follows:

Bottom of reservoir	-	0.020 Min.
Spun sides of reservoir	-	0.015 Min.
Threaded portion of wick tube section	-	0.010 Min.
Inner tube	-	0.010 Min.
Outer tube	-	0.015 Min."

3.2.1 Tapered wick tube section. After the last sentence add: "The tapered wick tube section shall be constructed so that the vertical axis is perpendicular to a horizontal plane."

Page 3:

Add new paragraph:

"3.2.4 Reservoir. The reservoir shall be fabricated of brass (see 3.1.1) and shall conform to DPSC Drawing NO. 23210 for style, design and dimensions. The inside of the opening shall be threaded to accommodate the tapered wick section."

Page 4:

3.6 Finish. Delete entirely and substitute:

"3.6 Finish. Burners shall be finished in accordance with Class 1, Type 1 of QQ-C-320."

4.1 and 4.1.1 Delete and substitute:

"4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate."

"4.1.3 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item."

5.1.3.1 Unit packages. Second sentence, line 4 following "Box" insert "shall be".

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6640-0674

Review information is current as of date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.