NNN-B-001636 (DSA-DM)
April 15, 1970
Interim Revision of Fed.
Spec. GG-L-101
August 10, 1948

INTERIM FEDERAL SPECIFICATION

BURNER, ALCOHOL, DENTAL

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim Federal specification as a valid exception to Federal Specification GG-L-101.

1. SCOPE

1.1 Scope. This specification covers one type of metal alcohol burner for use in dental facilities.

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawings. The following specifications and standards, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

QQ-B-663	Brass, Leaded and Nonleaded, Flat Products (Plate, Bar, Sheet and Strip)
JJJ-C-561	Cotton, Sterile (Purified and Nonsterile)
NNN-A-250	Adhesive, Water Resistant (For Sealing Fiberboard Boxes)
PPP-B-566	Boxes, Folding, Paperboard.
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Setup.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water Resistant (For Carton Sealing)

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547 Liners, Case, and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105 Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 Marking for Shipment and Storage

DRAWINGS

Defense Personnel Support Center (DPSC).
#23210 Burner, Alcohol

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D. C. 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston; New York; Atlanta; Chicago; Kansas City, Mo.; Forth Worth; Denver; San Francisco; Auburn, Washington; and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

(Copies of Military specifications, standards and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Materials.

3.1.1 Brass. Brass used in the fabrication of burners shall be in accordance with QQ-B-613 for Copper-alloy number 268 for Quarter-hard strip.

3.1.2 Cotton. Cotton used in the fabrication of the wick shall be in accordance with JJJ-C-561 for Grade B cotton.

3.2 Construction. Burner shall be fabricated from brass (see 3.1.1). Burner shall consist of a reservoir, a tapered wick tube section, a cap and a chain. Burner shall be constructed so that the base end of the wick tube section can be screwed into the top end of the reservoir.

3.2.1 Tapered wick tube section. Wick tube section shall be tapered and shall consist of a hollow base, an outer tube and an inner tube suitable for holding the wick, and a serrated wick-feed thumb screw. The base of the wick tube section shall be threaded 24 threads per inch. An air-vent hole shall be provided on one side of the assembly.

3.2.2 Cap and chain. A dome-shaped cap fabricated from brass (see 3.1.1) shall be constructed so that the inside diameter makes a snug fit with the top end of the wick tube section. The cap shall be attached to the wick tube section by a chain.

3.2.3 Wick. The core of the wick shall be fabricated from cotton (see 3.1.2) and loosely twisted into yarns and the covering shall be a loosely woven cotton fabric in the form of a seamless, hollow tube. Wick shall be 7 ± 1/4 inches long and shall have a diameter of 1/4 ± 1/8 inch.

3.3 Capacity. The capacity of the burner shall be not less than 3 and not greater than 3-1/4 ounces.

3.4 Style and design. Burner shall conform to DPSC Drawing number 23210 for style, design, and dimensions.

3.5 Leakage. Burner shall be constructed so that there is no evidence of leakage when tested as specified in 4.4.1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

3.6 Finish. All metal components shall be coated with a uniform thickness of chromium over nickel. Outer surfaces of burners shall be polished to a bright finish. Finished thickness (inches) of the following component brass parts shall be as follows:

Bottom of reservoir	-	0.20 min.
Spun sides of reservoir	-	0.015 min.
Threaded portion of wick tube section	-	0.010 min.
Inner wick tube	-	0.010 min.
Outer wick tube	-	0.015 min.

3.7 Identification marking. Burners shall be permanently and legibly marked with the supplier's name, manufacturer's name or registered trademark.

3.8 Workmanship. Workmanship shall be first class throughout. The finished burners shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

4.1.2 Certificates of quality. Certificates of quality, supplied by manufacturers of the brass and cotton, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the FSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	2.5
Minor defects	II	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 except that the Acceptance Number shall be zero.

Component (unit of product)	Characteristic	Requirement	procedure	Inspection level
Burner	Leakage	3.4	4.4.1	S-1

4.3 Examination. The burner shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

Categories and defects*

Major

101	Component part missing.
102	Component not secured.
103	Burner not free of dents.
104	Burner not free of sharp edges.
105	Wick not capable of being racked up and down when turning the wide-feed thumb-screw.
106	Chain broken.
107	Tapered wick tube section not perpendicular to the base.
108	Threads not free of deformation.
109	Burner not free of unplated areas.

Minor

201	Plating not smooth.
202	Plating not free of flaking and peeling.
203	Manufacturer's identification not complete, correct, legible, or permanent.
204	Exterior surfaces not entirely polished.

* Examination shall not be restricted to the classified possible defects listed above.

4.3.2 Dimension examination. The burner shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Leakage. The reservoir shall be filled with methyl or ethyl alcohol and allowed to stand for 1/2 hour on white blotter type paper. Alcohol shall contain an alcohol soluble dye. Blotter shall show no sign of stain due to leakage of burner.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Unit of issue. One burner, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in the following table shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

Table	Packaging Quantities	
Unit package	Intermediate package	Exterior container
1 unit	6 units	96 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be packaged in a box of appropriate size and design constructed in accordance with PPP-B-566 or PPP-B-676, except commercial colors will be acceptable, or PPP-B-636, type CF, class domestic. Box designed with suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of, or damage to, the contents. Closure shall be adequate to prevent accidental opening.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B, or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3. Closure and strapping shall be as specified in the appendix of the applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636 class domestic. Closure shall be as specified in the appendix of the applicable box specification.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with Fed. Std. No. 123.

5.3.2 Military activities.

5.3.2.1 Interior packages. Each unit package and each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. Alcohol burners covered by this specification are intended to be used in dental laboratories.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) FSN
(c) Selection of applicable levels of packing (see 5.1 and 5.2).

6.3 This specification does not cover all types of the commodity indicated by the title of this specification, or these which are commercially available, but is intended to cover the type which is normally procured to meet requirements.

6.4 This specification covers the following items shown in the Federal Supply Catalog:

FSN	Item Identification
6640-537-5010	Burner, Alcohol

Custodians

Army - MD
Navy - MS
Air Force - 03

Review activities

Army - MD
Navy - MS
Air Force - 03

Review information is current as of the date of this document. For future coordination of change to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

CAUTION

NOTICE TO BIDDERS/OFFERORS

DO NOT CONDITION OR BASE YOUR BID/OFFER ON ANY CURRENT PROCUREMENT ON THE INFORMATION SUBMITTED ON THIS FORM SINCE ANY CHANGES OR DELETIONS IN THE SPECIFICATIONS MAY RENDER YOUR BID/OFFER NON-RESPONSIVE IN WHICH CASE IT CANNOT BE CONSIDERED FOR AWARD.

Fold

DEFENSE PERSONNEL SUPPORT CENTER
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

DEFENSE SUPPLY AGENCY
OFFICIAL BUSINESS

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

Fold

DPSC FORM 2418
MAR 68

EDITION AUG 66, WILL BE USED UNTIL EXHAUSTED

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

SPECIFICATION ANALYSIS SHEET

Form Approved
Budget Bureau No. 22-R255

INSTRUCTIONS: This sheet is to be filled out by personnel, either Government or contractor, involved in the use of the specification in procurement of products for ultimate use by the Department of Defense. This sheet is provided for obtaining information on the use of this specification which will insure that suitable products can be procured with a minimum amount of delay and at the least cost. Comments and the return of this form will be appreciated. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments and suggestions submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or serve to amend contractual requirements.

SPECIFICATION NNN-B-001636(DSA-DM) FSN:

ORGANIZATION

CITY AND STATE | CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE SPECIFICATION CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY SPECIFICATION REQUIREMENT CONSIDERED TOO RIGID

3. IS THE SPECIFICATION RESTRICTIVE?

☐ YES ☐ NO *(If "yes", in what way?)*

4. REMARKS *(Attach any pertinent data which may be of use in improving this specification. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity)*

SUBMITTED BY *(Printed or typed name and activity - Optional)* | DATE

DD FORM 1 JAN 68 1426 REPLACES EDITION OF 1 OCT 64, WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.