FEDERAL SPECIFICATION

BURET, STRAIGHT, PRECISION

This amendment, which forms a part of Federal Specification NNN-B-789, dated March 25, 1963, was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

Page 1, paragraph 1.2.1.1: Delete "2.1.1" and substitute "1.2.1.1".

Page 2, paragraph 2.1, under Federal Standards, delete:

"Fed. Test Method Std. No. 626—Glassware; Methods for Sampling and Testing."

Page 2, paragraph 2.2: After reference to Commercial Standard, add:

"American Society for Testing and Materials:

C 148—Polariscopic Examination of Glass Containers.

(Copies of ASTM publications may be obtained from the American Society for Testing and Materials, 1916 Race Street, Philadelphia 3, Pa.)"

Page 2, paragraph 3.2, line 2: Delete "that of the standard reference disc No. 1" and substitute "Temper No. 1".

Page 3, paragraph 3.4.1: To the end of the paragraph, add:

"Style II and III burets shall have a filling tube positioned under the stopcock which will permit filling the buret through the stopcock. The filling tube shall be in the form of an elbow, extending downward from the stopcock. The open end of the elbow shall extend toward the rear from the buret and be slightly enlarged and bulbous. At the top, style III burets shall have an ungraduated tapered interior extention terminating in an overflow hole. The exterior of the buret shall extend above the point of overflow of the interior extention and shall have an outlet tube positioned below the overflow point of the interior extension at a point along the junction of the exterior to the interior extention. The end of the outlet tube shall be slightly enlarged and bulbous. The maximum capacity of the interior extention to the initial buret graduation shall not exceed 1.5 milliliters."

Page 3, paragraph 3.4.2: Delete in its entirety and substitute:

"3.4.2 *Stopcocks.* All burets shall have permanently attached stopcocks. Stopcocks shall not leak when tested as specified in 4.6.3. Stopcock plugs shall be provided with a retaining device and may be of glass or polytetrafluoroethylene (see 6.2). Glass stopcocks shall conform to Commercial Standard CS21, size 2. Polytetrafluoroethylene plugs may be of conventional or so-called reverse design."

Page 6, paragraph 4.6.1: Delete "No. 420 of Fed. Test Method Std. No 626." and substitute "A or B of ASTM C 148."

Page 7, paragraph 6.2, subparagraph (d): Delete in its entirety and substitute:

"(d) Whether glass or polytetrafluoroethylene plugs are required."

Page 8, change custodian symbols as below.

MILITARY CUSTODIANS:

Army—MU
Navy—BuMed
Air Force—MAA

Preparing Activity

Army—MU

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.