NNN-C-1494A
April 27, 1973
SUPERSEDING
Int. Fed. Spec. NNN-C-001494 (GSA-FSS)
September 28, 1968

FEDERAL SPECIFICATION

CUP, BLOOD SAMPLE

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 Scope. This specification covers plastic sample cups for use in an Auto-Analyzer (see 6.1).

1.2 Classification.

1.2.1 Types. Plastic blood sample cups covered by this specification shall be of the following types and capacities, as specified (see 6.2).

Type I - Conical bottom, 0.5 ml. capacity.
Type II - Conical bottom, 2.0 ml. capacity.
Type III - Conical bottom, 4.0 ml. capacity.
Type IV - Flat bottom, 2.5 ml. capacity.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposal, form a part of this specification to the extent specified herein.

Federal Specifications:

L-P-396 - Plastic Molding, Polystyrene.
PPP-B-636 - Boxes, Shipping, Fiberboard.

Federal Standard:

Fed. Std. No. 123 - Marking for Domestic Shipment (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthy supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Services Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Standards.

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless a specific issue is identified, the issue in effect on date of invitation for bids or request for proposal shall apply.

National Motor Freight Traffic Association, Inc., Agent:

National Motor Freight Classification.

(Application for copies should be addressed to the American Trucking Associations, Inc., Tariff Order Section, 1616 P Street, N.W., Washington, DC 20036.)

Uniform Classification Committee, Agent:

Uniform Freight Classification.

(Application for copies should be addressed to the Uniform Classification Committee, Room 1106, 222 South Riverside Plaza, Chicago, IL 60606.)

3. REQUIREMENTS

3.1 Material. The material shall be a clear polystyrene, meeting the requirements of type I, class I of L-P-396 (see 4.1.2).

3.2 Capacities and dimensions. The capacities and dimensions of the cups shall conform to the requirements in table I.

TABLE I.

Characteristic	Type I		Type II		Type III		Type IV	
	Min	Max	Min	Max	Min	Max	Min	Max
Capacity, ml.	0.5		2.0		4.0		2.5	
Height, Overall, mm.	22.9	25.4	22.2	28.4	35.6	40.6	22.2	25.4
Inside diameter, mm.	6.3	8.9	11.1	12.7	12.7	16.5	11.1	12.7
Angle between conical bottom and skirt, degrees.	14.0	18.0	25.0	35.0	40.0	50.0	NO CONE	NO CONE
Cup wall width, mm.	0.63	0.89	0.55	0.89	0.89	1.3	0.55	0.89
Weight, Overall, grams.	0.80	1.75	0.80	1.65	1.9	2.3	0.80	1.65
Depth, Overall, inside, mm.	21.6	24.1	21.6	24.1	33.0	38.1	21.6	24.1

3.3 Construction. The cups (all types) shall be one piece construction, and have an outer ridge approximately 3.0 mm below the lip. The cups shall stand unsupported in an upright position on a level surface.

3.3.1 Types I, II, and III. The cups shall have a circular cross-section, with parallel or slightly tapered walls which change angle to form a conical bottom. The line of the wall shall be continued to form a skirt which will permit the cups to stand unsupported as specified in 3.3. The angle between the skirt and the conical bottom shall be as specified in table I.

3.3.2 Type IV. The cups shall have a circular cross-section with parallel walls, cylindrical in shape, and a flat bottom.

3.4 Workmanship. The sampling cups shall be free from cracks, burrs, dirt or blemishes which detract from their appearance of may impair their utility.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.2 Certificate of analysis. In lieu of contractor performing tests, certified test data furnished by the manufacturer of the polystyrene shall be acceptable as evidence of compliance with 3.1.

4.2 Sampling for inspection and acceptance. Sampling for inspection and acception shall be performed in accordance with the provisions set forth in MIL-STD-105, except where otherwise indicated. For purposes of sampling, one inspection lot for examination shall consist of cups of the same type and capacity, made by one manufacturer and submitted for delivery at one time.

4.2.1 Visual and dimensional examination (end item). The inspection levels and acceptable quality levels (AQL's) shall be as indicated below. The sample unit shall be one sampling cup unless otherwise specified.

	Inspection level	AQL (percent defective)
Visual examination		
Major defects	II	1.0
Minor defects	II	4.0
Dimensional examination	S-3	2.5

4.2.2 Defects. Major defects shall consist of cups being cracked, burred, dirty or deformed so that cross-section is out of round. Minor defects shall consist of haziness in the plastic or scratches.

4.2.3 Dimensional examination. Cups shall be examined for defects in capacity and dimensions. Any deviation from the requirements for capacity and dimensions set forth in table I shall be considered a defect.

4.3 Examination for preparation for delivery. An examination shall be made to determine that packing, packing and marking comply with the requirements of section 5. Defects shall be scored as specified in table II. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 with an AQL of 4.0 defects, expressed in terms of defects per hundred units.

TABLE II. Examination of preparation for delivery

Examine	Defect
Containers	Not as specified.
Contents	More or less than required.
Marking	Omitted; incorrect; illegible; improper size, location, sequence or method of application.
Material	Component missing or damaged.
Workmanship	Bulging or distortion of containers, cushioning inadequate, improper or missing.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A, B, or C, as specified (see 6.2).

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:49Z
Check the source to verify that this is the current version before use.

5.1.1 Level A. One thousand cups, of one type only, shall be packaged in a polyethylene bag. The filled sealed bag shall be packaged in a box conforming to PPP-B-636, class weather-resistant, grade W5c.

5.1.2 Level B. One thousand cups, of one type only, shall be packaged in a polyethylene bag. The filled sealed bag shall be packaged in a box conforming to PPP-B-636, class domestic.

5.1.3 Level C. Cups, in quantities as specified (see 6.2), of one type only shall be packaged to afford adequate protection against damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.2).

5.2.1 Level A. Ten boxes (10,000 cups) of type I, II, or IV and five boxes (5,000 cups) of type III, packages as specified in 5.1.1, shall be packed in a box conforming to PPP-B-636, class weather-resistant, grade V3c.

5.2.2 Level B. Ten boxes (10,000 cups) of type I, II, or IV and five boxes (5,000 cups) of type III, packages as specified in 5.1.2, shall be packed in a box conforming to PPP-B-636, class domestic.

5.2.3 Level C. Cups in quantities as specified (see 6.2), packaged as specified in 5.1.3, shall be packed in containers to assure carrier acceptance and safe arrival at destination in compliance with Uniform Freight Classification or National Motor Freight Classification, as applicable.

5.3 Marking. In addition to any special marking required by the procurement documents, marking of the interior packages and exterior shipping containers, shall be in accordance with Fed. Std. No. 123 for civil agencies or MIL-STD-129 for military agencies, as applicable (see 6.2).

6. NOTES

6.1 Intended use. The plastic sample cups are intended for use with automatic analyzers.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

(a) Title, number, and date of this specification.
(b) Type and capacity required (see 1.2.1).
(c) Selection of the applicable level of packaging and packing (see 5.1 and 5.2).
(d) Quantities required (see 5.1.3 and 5.2.3).
(e) Applicable marking document (see 5.3).

6.3 Type II covers FSN 6640-710-1775.

6.4 Samples. It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples are necessary with bids, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

Preparing Activity:

GSA-FSS

CIVIL AGENCY INTEREST:

HEW-NIH
VA-DMS

U. S. GOVERNMENT PRINTING OFFICE : 1973 O - 514-180 (576)

Orders for this publication are to be placed with General Services Administration acting as an agent for the Superintendent of Documents. See Section 2 of this specification to obtain extra copies and other documents referenced herein. Price 10 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.