NNN-G-698a
February 14, 1963
SUPERSEDING
Int. Fed. Spec. NNN-G-00698 (ARMY-Cm1C)
June 27, 1961 and
Fed. Spec. RR-G-675
November 23, 1959

FEDERAL SPECIFICATION

GRID, WIRE GAUZE, LABORATORY

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 **Scope.** This specification covers two types of wire gauze grids suitable for use in laboratories.

1.2 **Classification.**

1.2.1 *Types, grades, and dimensions.* Wire gauze grids shall be of the following types, grades, and dimensions, as specified (see 6.1).

Type I—Grid, wire gauze, plain.

Grade A—Brass.
Grade B—Nickel-chromium alloy, no iron.

Type II—Grid, wire gauze, asbestos center patch.

Grade C—Corrosion-resisting steel.
Grade D—Carbon steel-zinc coated.

Dimensions:

Size 1—4 inches square.
Size 2—5 inches square.
Size 3—6 inches square.

2. APPLICABLE SPECIFICATIONS, STANDARDS, AND OTHER PUBLICATIONS

2.1 **Specifications and standards.** The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification to the extent specified herein.

Federal Specifications:

QQ-W-321—Wire, Brass.
QQ-W-423—Wire, Steel, Corrosion-Resisting.
QQ-W-461—Wire, Steel, Carbon (Round, Bare and Coated).
UU-P-268—Paper, Kraft, Untreated, Wrapping.
UU-T-111—Tape; Paper, Gummed (Sealing and Securing).
PPP-B-566—Boxes, Folding, Paperboard.
PPP-B-585—Boxes, Wood, Wirebound.
PPP-B-601—Boxes, Wood, Cleated-Plywood.
PPP-B-621—Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636—Box, Fiberboard.
PPP-B-676—Boxes, Set-up, Paperboard.

Federal Standards:

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.
Fed. Std. No. 123—Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Informa-

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

tion in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Washington, D. C., Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, and Auburn, Wash.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Specification:

MIL-L-10547—Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

Military Standards:

MIL-STD-105—Sampling Procedures and Tables for Inspection by Attributes.

MIL-STD-129—Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 **Other publications.** The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

Uniform Freight Classification Rules.

(Application for copies should be addressed to Uniform Classification Committee, 202 Union Station, Chicago 6, Illinois.)

3. REQUIREMENTS

3.1 Material. Wire gauze grids covered by this specification shall be brass, nickel-chromium alloy (no iron), corrosion-resisting steel, or carbon steel, zinc-coated.

3.2 Chemical composition.

3.2.1 *Grade A.* Brass wire shall comply with chemical composition A or B in accordance with QQ-W-321.

3.2.2 *Grade B.* Nickel-chromium alloy wire shall be free from iron and shall be 80 percent nickel and 20 percent chromium.

3.2.3 *Grade C.* Corrosion-resisting steel wire shall comply with the chemical compo-

TABLE I

Type	Dimensions* inches, square	Grade of wire	Wire diameter, inch	Asbestos patch diameter, inches	Mesh per linear inch
I	4	B	0.020 ± 0.001	---	16
II	4	D	0.0159 thru 0.023	3 ± 1/8	10 to 16
II	5	D	0.0159 thru 0.023	3 ± 1/8	10 to 16
I	6	B	0.020 ± 0.001	---	16
I	6	A	0.018 ± 0.001	---	16
II	6	D	0.0159 thru 0.023	3-1/2 ± 1/8	10 to 16
II	6	C	0.025 ± 0.001	3-1/2 ± 1/8	10 to 16

* All edges shall be turned under 1/4 ± 1/16 inch on each edge. Side dimensions shall have a tolerance of ± 1/8 inch.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

sitions FS 302 or 304, in accordance with QQ-W-423.

3.2.4 *Grade D.* Carbon steel wire shall comply with chemical compositions FS 1010 or 1015 in accordance with QQ-W-461.

3.3 **Dimensions and mesh.** The grid wire gauzes shall conform to the requirements of table I for dimensions and mesh and wire diameter.

3.4 **Asbestos center.** A sufficient amount and thickness of asbestos shall be pressed into the wire mesh to a degree which will insure its retention for the entire life of the wire grid. Dimensions shall conform to the requirements of table I.

3.5 **Finish.**

3.5.1 *Carbon steel, zinc coated.* The zinc coating on the individual wires which make up the wire gauze shall be not less than 0.10 ounce per square foot and shall be uniformly and completely coated with zinc applied by the electroplating process.

3.6 **Workmanship.** The wire gauze shall be free of damage, such as broken wires, holes, dents, or distortion which impairs its serviceability or use.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 The supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own or any other inspection facilities and services acceptable to the Government. Inspection records of the examination and tests shall be kept complete and available to the Government as specified in the contract or order. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.2 **Special provisions.**

4.2.1 *Alternative inspection (including testing) procedures.* The supplier may utilize any alternative inspection procedure which will insure equal or better quality by submitting a written proposal with justification and obtaining written approval from the Government prior to instituting the procedure. In case of dispute, the procedures of this specification will govern.

4.2.2 *Supplemental evidence.* The supplier shall provide evidence acceptable to the contracting officer that the requirements of 3.1, 3.2, 3.5.1, and section 5 have been satisfied.

4.3 **Lotting.** A lot shall consist of the wire gauze grids of the same type, grade, and dimensions from one manufacturer and offered for acceptance at one time.

4.4 **Sampling**

4.4.1 *Nondestructive examination.* Sampling shall be conducted in accordance with MIL-STD-105.

4.5 **Examination.** Sample wire gauze grids and level A and level B preparation for delivery requirements shall be examined in accordance with the classification of defects and with MIL-STD-105.

4.5.1 *Classification of defects.*

4.5.1.1 *Grid, wire gauze.*

Categories	Defects
Critical:	None defined
Major:	*AQL 4.0 percent defective*
101	Wire diameter or mesh incorrect
102	Coating incorrect where specified
103	Edges improperly turned
104	Asbestos center incorrect or incorrectly incorporated
105	Damage (broken wire, holes, dents, distortion)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

4.5.1.2 *Preparation for delivery (levels A and B).*

Categories	Defects
Critical:	None defined
Major:	*AQL 1.5 percent defective*
101	Wrapping, intermediate, or shipping container not as specified
102	Unlike grids packaged per wrap or intermediate package
103	Closure of wrap, intermediate package, or shipping container not as specified
104	Wrap, intermediate package, or shipping container damaged
105	Marking of wrap, intermediate package, or shipping container illegible, incorrect, or incomplete

5. PREPARATION FOR DELIVERY

For civil agency procurement, the definitions and applications of the levels of packaging and packing shall be in accordance with Fed. Std. No. 102.

5.1 **Packaging.** Packaging shall be level A or C, as specified (see 6.1).

5.1.1 *Level A.*

5.1.1.1 *Unit packaging.* The number of wire gauze grids specified (see 6.1), of the same type, grade, and dimensions, shall be completely wrapped in minimum 40 pound basis weight kraft paper conforming to UU-P-268. The wrap shall be sealed with tape conforming to UU-T-111.

5.1.1.1.1 *Intermediate packaging.* The number of unit packages specified (see 6.1) shall be intermediately packaged in a snug-fitting box conforming to PPP-B-566, style 1 or II, type and class optional; PPP-B-636, type I, class 1; or PPP-B-676, type, class, and style optional. Box closure shall be secured with tape conforming to UU-T-111. All wire gauze grids in an intermediate package shall be of the same type, grade, and size.

5.1.2 *Level C.* Wire gauze grids shall be packaged in accordance with the manufacturer's commercial practice suitable for shipment from the supply source to the first receiving activity for immediate use.

5.2 **Packing.** Packing shall be level A, B, or C, as specified (see 6.1).

5.2.1 *Level A.* Wire gauze grids shall be packed in a box conforming to PPP-B-621, class 2; PPP-B-601, overseas type; or PPP-B-585, style 3, class 3. The container shall be provided with a liner conforming to MIL-L-10547 which shall be closed and sealed in accordance with the appendix thereto. As far as practical, exterior containers shall be of uniform size and shape, be of minimum cube and tare, and contain identical quantities of like items. Gross weight shall not exceed 200 pounds. Containers shall be closed and strapped in accordance with the applicable box specification or appendix thereto.

5.2.2 *Level B.* Wire gauze grids shall be packed in fiberboard boxes conforming to PPP-B-636, type I or II, class 1, style RSC, minimum bursting strength 275 pounds per square inch.

5.2.3 *Level C.* Wire gauze grids shall be packed to insure carrier acceptance and safe delivery at first domestic destination. Containers shall be in accordance with Uniform Freight Classification Rules or regulations of other carriers applicable to the mode of transportation.

5.3 **Marking.**

5.3.1 *Civil agencies.* In addition to any special marking required by the contract or order, marking shall be in accordance with Fed. Std. No. 123.

5.3.2 *Military agencies.* In addition to any special marking required by the contract or order, unit packages, intermediate packages, and shipping containers shall be marked in accordance with MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Ordering data. Purchasers should exercise any desired options offered herein and procurement documents should specify the following:

(*a*) Title, number, and date of this specification.

(*b*) Level of packaging and packing required.

(*c*) Type, grade, and dimensions of grid required.

(*d*) Number of grids per unit, intermediate, or shipping container required.

6.2 Military use. Items procured under this specification for military use are to be limited to the variety shown on the applicable military standard or military supply standards. Personnel of the military departments are requested to refer to these documents for guidance.

6.3 Transportation description. Transportation description applicable to this item is:

Wire goods, not otherwise indexed by name.
Carload minimum weight 20,000 pounds, subject to Rule 34, Uniform Freight Classification.

Wire goods, not otherwise indexed.
Motor volume minimum weight 20,000 pounds.

Notice. When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY CUSTODIANS:

Army—MU

Navy—Md

Air Force—DAY

OTHER INTERESTS:

Army—EMO

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.