INTERIM AMENDMENT TO FEDERAL SPECIFICATION

PIPET, SEROLOGICAL, Sahli, Diluting, Hemoglobinometer, 0.02 ml.

This interim amendment was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above. The General Services Administration has authorized the use of this interim amendment as a valid exception to Federal specification NNN-P-001482B (DSA-DM).

Page 4

Table I. Delete and substitute:

"Table I For examination

	Inspection level	AQL (percent defective)
'For visual examination		
Major defects	II	2.5
Minor defects	II	4.0
For dimensional examination	S-1	2.5 "

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6640-0905

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.