MIN-P-1482A
March 19, 1970
SUPERSEDING
Int. Fed. Spec. MIN-P-001482(GSA-FSS)
October 9, 1968

FEDERAL SPECIFICATION

PIPET, SEROLOGICAL, SAHLI HEMOGLOBIN

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE

1.1 This specification covers Sahli diluting pipet for use in the estimation of hemoglobin.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein:

Federal Specifications:

PPP-B-566 - Boxes, Folding, Paperboard.
PPP-B-636 - Box, Fiberboard.
PPP-B-665 - Boxes, Paperboard, Metal Stayed (Including Stay Material).
PPP-B-676 - Boxes, Setup.

Federal Standards:

Fed. Std. No. 123 - Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington, D. C. 20402.

(Single copies of this specification and other Federal Specifications required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, D.C., Atlanta, Chicago, Kansas City, Mo., Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, Wash.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

3. REQUIREMENTS

3.1 Material. The pipet shall be made from high quality milk backing tubing glass.

3.2 Design and dimensions. The pipet shall be of the conventional Sahli design and shall not be less than 5 inches in overall length. The cross section of the pipet at any point shall be circular. The distal end shall be drawn down and the end carefully ground and polished to a shorter taper joint. The pipet shall be calibrated to contain 20 cu. mm. (0.02 milliliter), held to a tolerance of plus or minus two percent. The graduation mark shall completely encircle the transparent portion of the pipet at a right angle to the long axis. The graduations may be applied by either of the following methods: etched and filled with a permanent pigment or by application of a stain fired into the glass without etching and shall pass the test for permanency in 4.3.3.

3.3 Serial number. An identifying serial number and the name or registered trade-mark of the manufacturer shall be placed on each pipet.

3.4 Thermal resistance.

3.4.1 Boiling. There shall be no evidence of corrosion, scumming, chipping or cracking when tested as described in 4.3.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

3.4.2 Shock. There shall be no evidence of cracking when tested as described in 4.3.2.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.2 Sampling, inspecting, and testing. One (1) out of every 144 pipets shall be inspected and tested to determine compliance with the requirements of this specification.

4.3 Test methods.

4.3.1 Boiling test. The pipets shall be completely immersed in distilled water in a closed vessel having a steam vent and boiled continuously for 6 hours, boiling distilled water being added when necessary to compensate for evaporation losses with the pipets suspended in water so they do not come in contact with the containing vessel.

4.3.2 Shock. The pipets shall be immersed in tap water at 65° plus or minus 2.5° F., allowed to remain 10 minutes and then transferred into boiling water. Repeat the operation in 5 minutes with no breakage occurring.

4.3.3 Permanency of pigment. Immerse the pipets in a Chromic Acid Cleaning Mixture* at room temperature. Let stand for 15 minutes, then remove samples and rinse thoroughly with a cloth or cleaning tissue. The appearance of lines and number is to be the same as before the test, as judged by the eye.

*Chromic Acid Mixture formula:

Sodium Dichromate	200 gm.
Water	1000 ml.
Sulfuric Acid	1000 ml.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level B.

5.1.1 Unit packaging. Twelve pipets shall be packaged in a snug-fitting paperboard box conforming to PPP-B-566, PPP-B-665, or PPP-B-676. The box shall be fitted with dividers or cushioning to protect the pipets from coming in contact with each other.

5.2 Packing. Packing shall be level B.

5.2.1 Shipping container packing. Forty-eight unit packages shall be packed in a snug-fitting fiberboard box conforming to PPP-B-636, class domestic, grade 275.

5.3 Marking. Each unit packages shall be marked with the Federal stock number, nomenclature, and quantity. Shipping containers shall be marked in accordance with Fed. Std. No. 123. The quantity shall be shown as 48 pg.

6. NOTES

6.1 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

(a) Title, number, and date of this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

MILITARY CUSTODIANS:

Army - MD
Navy - MS
Air Force - 03
DSA - DM

Review activities:

Army - MD
Navy - MS
Air Force - 03
DSA - DM

User activities:

Army - MD
Navy - MS
Air Force - 03
DSA - DM

Preparing activity:

GSA-FSS

CIVIL AGENCIES INTEREST:

VA
GSA-FSS
HHW

Orders for this publications are to be placed with General Services Administration, acting as an agent for the Superintendent of Documents. See section 2 of this specification to obtain extra copies and other documents referenced herein. Price 10 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:49Z
Check the source to verify that this is the current version before use.