A-A-0051670(DM)
26 September 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

STOCKINGS, ANTI-EMBOLISM

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers one type, four sizes of a thigh length anti-embolism stocking which is worn by surgical patients as an aid in promoting venal return while bedridden.

Salient characteristics:

The stockings shall be anti-embolism type, thigh length, contour knitted with a combination of stretch nylon and spandex elastic yarns in the following sizes:

Size

1. Small, Regular length, for use on calfs up to 12 inches (12-3/4 in max.), 29 to 33 inch length (31 in. min.)

2. Medium, short length, for use on 12 to 15 inch calf, up to 29 inch length.

3. Medium, regular length, for use on 12 to 15 inch calf, 29 to 33 inch length (31 in min).

4. Medium long length, for use on 12 to 15 inch calf, over 33 inch length (31 in min).

The leg length shall be measured form the gluteal furrow to the bottom of the heel.

The stockings shall be designed to provide uniformly graduated compression decreasing from foot to knee and from knee to upper thigh. The stockings shall exert proper compression to the legs of hospital patients as a prophylaxis for thrombo-embolism.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

The stockings shall be seamless with a full foot and shall contain an inspection toe opening. As an alternate, a clear-view inspection window may be substituted for the inspection toe opening.

The stockings shall be knit with stretch type nylon yarns using a combination of "S" and "Z" twist torque yarns to produce a balanced fabric. The spandex elastic yarn shall be at least single wrapped with continuous multi-filament nylon stretch yarn.

The color of the stockings shall be white. The top of the stocking shall be color coded to indicate size: white for medium sizes and yellow for size small.

Workmanship. The anti-embolism stockings shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing six pair of anti-embolism stockings of one size only, along with laundry instructions, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:59Z
Check the source to verify that this is the currentversion before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one unit of issue.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

SIZE 1: 6515-01-229-2640
SIZE 2: 6515-01-229-2639
SIZE 3: 6515-01-229-2638
SIZE 4: 6515-01-229-7709

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the currentversion before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

YQP 70015

Project No. 6515-3874

Location: CID/STOCKING/H15-10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.