A-A-0051703 (DM)
13 October 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

CONTAINER AND PUMP, DENTAL (Mouthrinse)

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by DLA, Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers a container and pump suitable for use in dispensing 10 milliliters of sodium fluoride mouthrinse.

Class 1 - Container (Jug)
Class 2 - Pump

Salient characteristics:

The container shall be suitable for use with the pump to form a manually operated dispenser capable of dispensing 10 milliliters of (0.2%) sodium fluoride (NaF) solution upon each stroke.

Class 1, Container. The container shall be a one piece molded construction, including formed handle, of extra heavy gauge, natural colored, food grade polyethylene. The handle shall be so spaced from the body of the container to provide ease of handling. The container shall be unbreakable, leakproof and inert to a 0.2% NaF solution. The container shall be 10 ± 1/4 inches in height. The capacity of the container shall be 1600-2200 milliliters and shall feature graduations at the 500, 1000 and 1500 milliliter level. The container shall have an inner diameter opening of 1-1/4 ± 1/16 inches. The outer diameter of the opening shall be threaded to accommodate a standard 38 mm cap.

A standard 38 mm safety (childproof) cap shall be provided with each container and shall fit and seal without leakage, for storage of the solution. The cap and its liner shall be inert to the 0.2% NaF solution.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

The container shall be permanently labeled to read, "Sodium Fluoride Mouthrinse (0.2%). CAUTION: Do not swallow." Complete printed instructions shall be affixed to the container and the label shall be a waterproof plastic. The printing shall be waterproof and shall not run or wash out. Label and printing shall not be damaged by continuous washing.

Class 2, Pump. The pump shall screw onto the container and be manually operated. The pump with one full stroke shall be capable of ejecting 10 ml ± 1 ml of 0.2% NaF solution. The pump shall consist of a delivery tube with corrosion resistant steel fitments, a standard 38 mm screw cap, auxiliary cap and culminating in a molded plastic spout. The delivery tube shall be made of a suitable plastic, except for fitments. The fitments (the spring and two balls) shall be fabricated of 18-8 corrosion resistant steel. The fitments shall show no sign of corrosion when tested in accordance with the boil test as specified herein. The pickup end of the delivery tube shall be diagonally cut to assist in uptake. The uptake tip shall extend to within 3/4 inch of the bottom of the container to minimize waste. The standard 38 mm screw cap shall be a ring through which the delivery tube passes. The standard 38 mm screw cap together with the auxiliary screw cap shall securely attach the pump to the container without leakage. The opening of the spout shall extend outward nominally 3-1/4 inches from the center axis of the pump with the ejection tip angled downward to facilitate cup filling. All parts of the pump shall be inert to NaF solutions. There shall be no leakage in the entire system.

Workmanship. The container and pump shall be clean and free of corrosion, pitting, warping, deformation, cuts, holes, binding of moving parts or any defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One container (class 1) or one pump (class 2), as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Boil test. The fitments, before testing, shall be scrubbed with soap and water, rinsed thoroughly in distilled water, and dried. After cleaning, the fitments shall be boiled in distilled water in a glass beaker for 30 minutes and allowed to remain submerged for an additional 24 hours.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for tests purposes shall be one unit of issue.

Contractor certification. The contractor shall certify that the product offered meets the the salient characteristics of this description and conform to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

Stock Number	Item Identification
6520-01-223-1526	CONTAINER, MOUTHRINSE, DENTAL, 1500 ml
6520-01-225-4680	PUMP, DENTAL, Manually Operated, 10 ml Dispensing

Ordering data. (Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

Civil Agency Coordinating Activities:
VA-OSS
PHS

60104

PROJECT NO. 6520-1977

Location: NCR - Disk 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

DEFENSE LOGISTICS AGENCY
DEFENSE PERSONNEL SUPPORT CENTER
ATTN DPSC RSTS
2800 SOUTH 20TH STREET
PO BOX 8419
PHILADELPHIA PA 19101 8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-0051703 (DM)	CONTAINER AND PUMP, DENTAL (Mouthrinse)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:00Z
Check the source to verify that this is the current version before use.