A-A-51686
10 October 1986

COMMERCIAL ITEM DESCRIPTION

CRUCIBLE, FILTERING AND IGNITION, LABORATORY

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification NNN-C-00680(DLA-DM)

This commercial item description covers two types of laboratory crucibles for use in laboratory procedures.

Salient characteristics:

Laboratory crucibles shall have the following dimensional characteristics as listed in Table I.

Table I - Crucible Dimensions

Type	Style	Size	Capacity	Height*	Outside Diameter*
I-A	High Form	1	5 ml	20 mm	24 mm
I-A	High Form	2	10 ml	26 mm	30 mm
I-A	High Form	3	15 ml	29 mm	35 mm
I-A	High Form	4	30 ml	37 mm	43 mm
I-B	Wide Form	1	30 ml	31 mm	50 mm
I-B	Wide Form	2	50 ml	39 mm	61 mm
I-B	Wide Form	3	150 ml	52 mm	84 mm
II	Filtering	1	25 ml	41 mm	36 mm
II	Filtering	2	40 ml	47 mm	40 mm

*Tolerance, ± 1mm

Type I-A. Shall be a laboratory ignition crucible of the high form and shall have the nominal dimensions as listed in Table I.

Type I-B. Shall be a laboratory ignition crucible of the low-wide form and shall have the nominal dimensions as listed in Table I.

AMSC N/A FSC 6640
DISTRIBUTION STATE A. Approved for public release; distribution is unlimited

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.

Type II. Shall be a laboratory filtering crucible of the Gooch type and shall have the nominal dimensions as listed in Table I.

Crucibles - Shall be fabricated of porcelain glazed inside and out except for the outside bottom surface. Shall be designed for use with standard size crucible covers. Crucibles shall be permanently and legibly marked with the manufacturer name or registered trademark. Markings shall be legible after tests have been performed. Type I crucibles shall have a flat non-perforated base suitable for resting firmly on a flat surface. The side wall shall be formed by a sloping convex surface extending upward from the base to the rim. Type II crucibles shall have the same style and design as Type I except that the base shall be perforated.

Laboratory crucibles shall withstand normal laboratory procedures and show no sign of deterioration or deformation of materials. Lot sizes shall be uniform in all characteristics and materials of construction for each type of crucible. Crucibles shall have none of the following which shall be considered defects.

Incorrect capacities and dimensions
Nonuniform glazing
Cracks, chipping, pinholes, blistering, orange peel
Nonuniform appearance
Deformation of materials
Missing, incorrect or illegible identification markings
Grease, oil or other foreign matter not easily cleaned during normal laboratory procedures

Workmanship. The laboratory crucibles shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One crucible of the type and size as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspections, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. Unless otherwise specified, sampling for examination shall be conducted at inspection level S-2, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Tests.

Thermal Shock Resistance - The crucibles shall be heated to a temperature of 160° ± 2°C in an oven for 30 minutes, removed and immediately plunged into water at a temperature of 9° ± 1°C. Crucibles shall show no sign of cracking or crazing.

Glaze Adhesion - Permanent adhesion of one glazed surface to another glazed procelain surface shall not occur below 1150°C.

Resistance to Chemical Attack - Tared crucibles shall be boiled in a 10 percent solution of sodium hydroxide for 8 hours, rinsed, dried and reweighed. Weight loss shall be calculated as follows and shall not exceed 0.144 percent.

$$L = \frac{100\ (B-A)}{B}$$

where: L=percent weight loss
B=weight in grams before boiling
A=weight in grams after boiling

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

For information only.

The following National Stock Numbers are covered by this document:

6640-00-893-6915	Crucible, Ignition, Lab	Size 1, Type IA
6640-00-290-6780	Crucible, Ignition, Lab	Size 2, Type IA
6640-00-290-6781	Crucible, Ignition, Lab	Size 3, Type IA
6640-00-419-4000	Crucible, Ignition, Lab	Size 4, Type IA
6640-00-290-6790	Crucible, Ignition, Lab	Size 1, Type IB
6640-00-290-6791	Crucible, Ignition, Lab	Size 2, Type IB
6640-00-290-6792	Crucible, Ignition, Lab	Size 3, Type IB
6640-00-419-2000	Crucible, Filtering, Lab	Size 1, Type II
6640-00-290-6549	Crucible, Filtering, Lab	Size 2, Type II

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

COMMERCIAL ITEM DESCRIPTION

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6640-1377

CID-1-CRU-FIL-IGN-LAB

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.