A-A-51691A
21 April 1987
SUPERSEDING
A-A-0051691(DM)
25 September 1986

COMMERCIAL ITEM DESCRIPTION

DIRECTOR AND TONGUE TIE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers two types of director and tongue ties.

Salient characteristics.

The instrument shall be a straight, grooved director and tongue tie suitable for use in surgical procedures. The type of the director shall be as specified below:

Type	Figure
Plain Tip	Figure 1
Olivary Tip	Figure 2

Plain tip directors shall have a blunt, round open end. Olivary tip directors shall have a probe at the end suitable for exploration and guidance.

The director and tongue tie shall be fabricated of Class 3 or Class 4 corrosion-resistant steel in accordance with ASTM F899. Instruments fabricated of Class 3 material shall have a hardness of Rockwell B-75 to B-100 when tested as specified in ASTM E18, and meet the requirements of the boil test specified in GG-I-526. Instruments fabricated of Class 4 material shall have a hardness of Rockwell C-40 to C-50 when tested as specified in ASTM E18, and meet the requirements of the boil test and copper sulfate test specified in GG-I-526.

Instruments shall have a mirror finish, except the groove which may have a satin finish.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:59Z
Check the source to verify that this is the current version before use.

Workmanship. The director and tongue tie shall be free from loose metal, abrasions, sharp surfaces, cuts, cracks and other defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One director and tongue tie of the same type, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 , inspection level S-2 and AQL 2.5%. The unit of product for tests purposes shall be one director and tongue tie.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-00-325-9400	DIRECTOR AND TONGUE TIE, Olivary Point, 5-1/2 Inch
6515-01-164-8889	DIRECTOR AND TONGUE TIE, Straight, 6 Inch

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4004

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.




FIGURE 1

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.




FIGURE 2.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:59Z
Check the source to verify that this is the current version before use.