A-A-51694
29 September 1986

COMMERCIAL ITEM DESCRIPTION

STOPCOCK, ARTERIAL-VENOUS PRESSURE MONITORING (ONE-WAY)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a sterile disposable plastic stopcock suitable for use in arterial and venous pressure monitoring.

Salient characteristics:

Material. Stopcocks shall be manufactured from a rigid plastic material.

Style and design. The stopcock shall have two ports located 180 degrees apart. One port shall be a male Luer lock connector and the other shall be a female Luer lock connector. The Luer lock connectors shall meet the dimensional requirements of American National Standards Institute Healtr Industry Manufacturer's Association (ANSI/HIMA) Standard MD70.1. The handle of the rotating member of the stopcock shall be an outwardly projecting arm. The arm shall be parallel to the fluid path when the stopcock is in the open position, and shall be perpendicular to the fluid path when the stopcock is closed.

Sterility and freedom from pyrogens. The stopcock shall be tested for sterility in accordance with the United States Pharmacopiea (USP). Freedom from pyrogens shall be tested in accordance with the USP Pyrogen Test or Bacterial Endotoxins Test.

Safety and heavy metals. Material used in the manufacture of the stopcock shall meet the requirements of the USP for Class II plastics. In addition, the materials shall contain less than 5 parts per million of heavy metals when tested in accordance with the USP.

Workmanship. The item shall be processed in such a manner as to be uniform in quality and free from defects that will affect life, serviceability, or appearance.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One stopcock, individually packaged sterile in a suitable tear resistant, peel-open type container capable of maintaining sterility of contents unless package is damaged or opened constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Leakage. Luer connectors shall be tested in accordance with the Test for Water Leakage of ANSI/HIMA MD70.1 except that the 45 psi pressure shall be maintained for three minutes. The outlet valve described in ANSI/HIMA MD70.1 shall be the rotating member of the stopcock. In addition to the requirements of ANSI/HIMA MD70.1, there shall be no leakage sufficient to form a falling drop from the rotating member-stopcock body juncture.

Security of assembly. Luer connectors shall be tested in accordance with ANSI/HIMA MD70.1 Test for Resistance to Separation and Unscrewing. In addition, the stopcock body shall be secured and an 8 pound (minimum) tensile force applied to the rotating member. The rotating member and the body shall show no signs of breakage, nor shall the rotating member separate from the stopcock-body.

Clear passage of fluids. The test apparatus used in the leakage test shall be connected to the stopcock. The stopcock shall be closed and a pressure of 3.5 psi applied. The stopcock shall then be opened. There shall be clear passage of water within 2 seconds of opening the stopcock.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:59Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for performance tests shall be conducted in accordance with MIL-STD-105 for sampling level S-1, AQL 1.0. The unit of product for test purposes shall be one stopcock.

Certificates of compliance for performance, safety, and heavy metals. A certificate of compliance, stipulating that offered items meet the performance, safety and heavy metals tests, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years, and that there have been no changes to material formulation and method of processing since actual testing was performed. The certificate shall include the date of actual testing, the test method and results, the name and title of the person evaluating results, the current contract number and NSN, a statement to the effect that material processing and formulation have not changed since actual testing was performed, and the signature and title of the person making the certification.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:59Z
Check the source to verify that this is the currentversion before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-9126	STOPCOCK, ARTERIAL-VENOUS PRESSURE MONITORING

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 51407

Project No. 6515-3607

U S GOVERNMENT PRINTING OFFICE: 1986 - 704-036/45606

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use i submitting comments and suggestions for improvements. All users of military standardization documents are invited to provid suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, an mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, w too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATE

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:59Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51694	STOPCOCK, ARTERIAL-VENOUS PRESSURE MONITORING(ONE WAY)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T08:59Z
Check the source to verify that this is the current version before use.