A-A-51985
26 June 1987

COMMERCIAL ITEM DESCRIPTION

APRON, X-RAY PROTECTIVE

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a lead-impregnated vinyl plastic apron suitable for protection of medical personnel during fluoroscopic or radiological procedures.

Salient characteristics:

The apron shall consist of 4 or more layers of lead-impregnated vinyl plastic, or of a mixture of lead-vinyl and antimony-vinyl layers. The front and back facing shall be made of fiberglass, nylon-reinforced vinyl, or flexible urethane plastic. The shoulder (or back) pieces shall be made of the same material and color as used to make the front and back facings of the apron body.

The edges of the apron and shoulder pieces shall be bound with cloth tape, or otherwise suitably sealed.

The apron shall have a nominal dimension of 38 inches from the bottom of the neck cutout to the bottom of the apron. The width of the lower portion of the apron shall be a nominal 24 inches. The height of the neck cutout shall be 4 to 6-1/2 inches.

Two ties shall be furnished to secure the shoulder pieces. One strong tie, a minimum of 36 inches in length, shall be securely attached to each shoulder piece.

The apron shall provide a minimum of 0.5 mm lead equivalent protection at 85 KVP. Each apron shall be identified to show the manufacturer's name, lead equivalent protection at 85 KVP, month and year of manufacture or other control data to identify the apron by lot number, and a cautionary statement for users to advise them to periodically radiograph the apron to detect any damage to lead protective layers. The marking shall be by means of tabs secured to the inside of the shoulder pieces or edge binding, or by imprinting of the inside of the shoulder pieces.

AMSC N/A FSC 6532

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

The apron stitching shall be at least 6 stitches per inch for each line of stitching used. The stitching shall not damage the material. All seams shall be tightly secured, with no openings.

The finished apron shall be supple, washable, resistant to cracking after repeated flexing, shall not soil user's clothing when worn, and shall not exceed 10 pounds in total weight.

The apron shall have no cuts, holes, blisters, thin or hard spots, or abraded areas. The apron shall have no repairs or patches.

The apron shall be clean and free of grease or oil stains, or embedded foreign material.

Workmanship. The x-ray protective apron shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One x-ray protective apron, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for test purposes shall be one x-ray protective apron as specified in this description.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional defects	S-1	2.5

Examination. The x-ray protective apron shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

TABLE II. Categories and Defects.

Categories	Defects
Major	
101	Apron front and back facing not made of correct material.
102	Shoulder piece ties not supplied.
103	Apron does not have edge binding.
104	Apron has a cut, hole, or thin spot.
Minor	
201	Apron stitching is not at least 6 stitches per inch for each line of stitching used.
202	Apron has a repair or a patch.
203	Apron is not supple, washable, and resistant to cracking after repeated flexing.
204	Apron is not clean and free of grease or oil stains, or embedded foreign material.
205	Apron not properly identified, as per the required information.

Dimensional Examination. The x-ray protective apron shall be examined for defects in dimensions. Any dimension not within the specifications herein shall be classified as a defect.

Tests.

Lead Equivalent test.

Contractor shall radiograph two apron samples at 85 KVP with intensifying screens at an appropriate milliampere-seconds and target-film distance using a lead step wedge. The lead equivalent of the two samples shall then be determined. Failure to achieve an equivalency of 0.50 mm on either of the two samples shall constitute a major defect.

Weight test.

The apron shall be weighed. Failure to weigh 10 pounds or less shall constitute a major defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6532-00-935-9765

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6532-1167

PGC 50923

Location: CID/APRON/H13-24

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTC
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A A 51985	Apron, X-Ray Protective

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.