A-A-0051990(DM)
29 May 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

SYRINGE, CARTRIDGE (Medical, Plastic)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a plastic medical syringe to be used with 2 ml cartridge-needle units.

Salient characteristics:

The syringes shall be made of polycarbonate or ABS plastic or combinations of the two plastics. The tip of the plunger shall be made of austenitic stainless steel Type 303 or 304 in accordance with ASTM F899 or brass which has been nickel plated or chromium-nickel plated.

The syringe shall have a side opening of about 3/8 inch by 3-1/2 inches on one side and on the opposite side, in order to facililate gripping the cartridge-needle unit, a smaller opening of 5/16 by 9/16 up to 5/16 by 1-1/16 inches. At the distal end of the syringe there shall be a 0.290 + .005,-0 inch diameter opening to accommodate the needle hub of the cartridge-needle unit. The syringe barrel shall be provided with a winged finger grip. The finger grip shall be threaded to allow the cartridge-needle unit to be clamped within the barrel or, if the finger grip is molded integral with the barrel, have within the finger grip a threaded sleeve which can be screwed securely against the top of the cartridge-needle unit. The plunger shall have a female threaded tip with a 3-48 UNC-3B thread. The plunger tip shall be attachable to the screw of the rubber piston or plug in the cartridge-needle unit. There shall be a circular thumb rest at the top of the plunger. Serrations shall be present on the periphery of the 3/4 inch thumb rest and, in the case of adjustable sleeve, if present, its circular adjustment wheel. The over-all length of the syringe shall be 4-1/2 to 5 inches. The barrel shall be approximately 1/2 inch in outside diameter.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the currentversion before use.

The syringe shall accommodate 2 ml cartridge-needle units having an 11/32 inch outside diameter and measure 3 inches on the cylindrical portion. The syringe shall be designed so that the plunger can not be withdrawn completely from the barrel. The syringe assembly shall have adequate mechanical strength to move the cartridge piston or plug without undergoing any distortion.

The syringe shall be provided with instructions for use.

Workmanship. Workmanship shall be first class throughout. The syringe shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Each (EA). One syringe, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one syringe.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection level shall be as indicated in Table I. The unit for test purposes shall be one syringe.

Sampling for examination

Type of Defect	Inspection level	AQL (% defective)
101 Incorrect thread size in plunger tip	S-2	*
102 Plunger able to be withdrawn from barrel	S-2	*
103 Crazing or cracking of plastic	S-2	*
104 Dimensions not as specified	S-2	*
105 Construction not rigid during pressure or aspiration stroke on 2 ml cartridge	S-1	*
106 Presence of rough or sharp edges	S-2	*
107 Presence of foreign matter	S-2	*
108 Instruction not present	S-1	*

*Acceptance number shall be zero.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the currentversion before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exteri package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-245-1871	SYRINGE, CARTRIDGE, Medical, Plastic, Side Loading, For 2 ml Cartridge-Needle Units

Ordering data (Intermediate/exterior packaging quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:03Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4132

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:03Z
Check the source to verify that this is the current version before use.