A-A-0051994 (DM)
15 June 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

TESTER, PULP, DENTAL

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a battery operated dental pulp tester which shall have a digital display of condition of pulp tissue. The tester shall be suitable for use in determining the vitality of dental pulps. The dental pulp tester shall be an Analytic Technology Vitality Scanner, Model 2006, or equal, for performance having the following salient characteristics:

Salient characteristics:

Shall be a battery-powered transistorized tester which shall have an LED (Light Emitting Diode) digital display, with removable autoclavable probe tips, and an automatic turn-on, reset, turn-off feature. The tester shall be suitable for use in determining the vitality of pulps of teeth. The tester shall be portable and shall consist of a base module, a portable pen-sized probe which shall be connected to the base module by a flexible cord, two removable autoclavable probe tips, grounding lead with lip clip, and 1/2 fluid ounce topical fluoride gel.

The tips shall be epoxy coated stainless steel. The tips shall be furnished in two sizes, one short tip (5/8 inch long) and one long tip (1-1/8 inch long). The tips shall fit the probe securely so as to prevent loosening or falling out from the handle.

The housing of the tester shall be fabricated from aluminum, and exposed surfaces shall have a protective coating. The tester shall be 4-1/2 inches wide by 2-1/2 inches high by 7-3/4 inches deep. The probe handle shall be stainless steel. The probe handle shall be 6 inches long. Exposed surfaces of the tester housing and probe handle shall withstand repeated wipe down with ADA (American Dental Association) acceptable disinfectants without deterioration of function or esthetics.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

The length of the cord shall be a minimum of 5 feet long. Tolerances on dimensions shall be within plus or minus 10 percent of specified value.

The digital display shall provide instant, easy, readable data. The readout shall indicate the stimulus level on a relative scale. The stimulus scale shall indicate stimulus values from 0 to 80 in proportion to output voltage and shall be readily readable. A light in the probe tip shall indicate that electric contact has been established. The light in the probe tip shall start flashing when full scale on the display is reached.

The tester shall automatically turn on when the probe tip makes contact with a tooth. The automatic turn-on shall prevent testing unless a good electrical contact has been established (as indicated by the light in the probe tip). The stimulus level as sensed by the patient shall be indicated by the digital display. The display shall stop incrementing when the probe is disconnected from the tooth. The output voltage shall reset to its lowest value after a delay of 0.5 to 1.0 second. The output voltage shall be maintained without resetting if contact is reestablished within the delay interval. The tester shall turn off automatically following a delay of 10 to 15 seconds after breaking the electrical contact with a tooth. The tester shall indicate, via digital display, the stimulus level at which the tooth responded. The display shall automatically reset to zero after removal from the tooth.

The output stimulus shall be in bursts of ten high frequency pulses of negative polarity. The peak output shall go from a 15 volt peak to a 300 volt peak with 2 megohms load impedance. The spacing between pulses shall be determined by the rate control.

The rates of stimuli shall be easily controlled by a rate control wheel. The rate control wheel shall have a marking to indicate increasing and decreasing direction. The scale on the rate control wheel shall start at 1 and stop at 9. The tester output shall start at a low level and shall have a slow rate of increase in the bottom range of the scale. The rate of increase shall increase toward the middle and upper ranges so as to provide a more accurate diagnosis while minimizing the risk of jolting the patient.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

The dental pulp tester shall be powered from four, commercially available, size AA, 1.5 volt alkaline batteries. The dental pulp tester shall have a light on the module to indicate when the batteries are low. The dental pulp tester shall be furnished with batteries. The batteries shall be new. Not more than seventy-five (75) days shall have elapsed from date of manufacture of the batteries to the date of delivery of the tester to the Government. The batteries shall fit securely in the battery holder so as to prevent the batteries from falling out. The tester shall be suitable for use with rubber gloved hands. The tester shall be furnished with a grounding lead having a lip clip on one end and a clip suitable for attachment to the probe handle on the other end. The lip clip shall connect to the patient's lip. The clips shall be stainless steel. The grounding lead shall be an insulated covered copper wire. The lead shall safely permit current to flow through the tooth via the hand of the operator and then to the probe handle. The lip clip shall be removable for autoclaving. The tester shall be furnished with a 1/2 fluid ounce bottle of topical fluoride gel.

Technical manuals with complete instructions for use shall be furnished along with the clear warning that the dental pulp tester shall not be used on patients with pacemakers and any other safety precautions. The unit shall be safe for use on the patient.

Workmanship. The Dental Pulp Tester shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One tester, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one dental pulp tester. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:04Z
Check the source to verify that this is the currentversion before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6520-01-125-6618	TESTER, PULP, DENTAL, Battery Operated, Digital Display.

For replacement alkaline batteries - order NSN 6135-00-935-2587, Battery, Non-rechargeable, 1.5 volts, Unit of Issue Package (PG).

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Project No. 6520-2016

49134

Location: NCR CID DISK 3

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:04Z
Check the source to verify that this is the current version before use.