A-A-51994A
12 June 1989
SUPERSEDING
A-A-0051994 (DM)
15 June 1987

COMMERCIAL ITEM DESCRIPTION

TESTER, PULP, CANAL

The General Services Administration has authorized the use of this Commericial Item Description.

This commercial item description item covers a battery operated dental pulp tester used to determine the condition of pulp tissue.

Salient Characteristics:

Description. Shall be a battery-powered transistorized tester which shall have a LED (Light Emitting Diode) digital display, with removable, autoclavable probe tips, and an automatic turn-on, reset, turn-off feature. The tester shall be suitable for use in determining the vitality of a tooth's pulp. The tester shall be portable and shall consist of a base module, a portable pen-sized probe connected to the base module by a flexible cord, two removable, autoclavable probe tips, a grounding lead with lip clip, and 1/2 fluid ounce of topical flouride gel.

Tips. The tips shall be epoxy coated stainless steel. The tips shall be furnished in two sizes; one short tip (5/8 inch long) and one long tip (1-1/8 inch long). The tips shall fit the probe securely so as to prevent loosening or falling out from the handle.

Housing. The housing of the tester shall be fabricated from aluminum, and exposed surfaces shall have a protective coating. The tester shall be 4-1/2 inches wide by 2-1/2 inches high by 7-3/4 inches deep. The probe handle shall be fabricated from stainless steel and shall be 6 inches long. Exposed surfaces of the tester housing and probe handle shall withstand repeated wipe down with ADA (American Dental Association) acceptable disinfectants without deterioration of function or esthetics.

Cord. The length of the cord shall be a minimum of 5 feet long.

AMSC N/A FSC 6520
DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Digital display. The digital display shall provide instant, easy, readable data. The readout shall indicate the stimulus level on a relative scale. The stimulus scale shall indicate stimulus values from 0 to 80 in proportion to output voltage and shall be readily readable. A light in the probe tip shall start flashing when full scale on the display is reached.

Performance. The tester shall automatically turn on when the probe tip makes contact with a tooth. The automatic turn-on shall prevent testing unless a good electrical contact has been established (as indicated by the light in the probe tip). The stimulus level as sensed by the patient shall be indicated by the digital display. The display shall stop incrementing when the probe is disconnected from the tooth. The output voltage shall reset to its lowest value after a delay of 1.0 to 2.0 seconds. The display shall automatically reset to zero when again making contact with a tooth after the delay specified above. The output voltage shall be maintained without resetting if contact is reestablished within the delay interval. The tester shall turn off automatically following a delay of 10 to 15 seconds after breaking the electrical contact with a tooth. The tester shall indicate, via digital display, the stimulus level at which the tooth responded.

The output stimulus shall be in bursts of ten high frequency pulses of negative polarity. The peak outputs shall go from a 15 volt peak to a 300 volt peak with 2 megohms load impedance. The spacing between pulses shall be determined by the rate control.

The rates of stimuli shall be easily controlled by a rate control wheel. The rate control wheel shall have a marking to indicate increasing and decreasing direction. The scale on the rate control wheel shall start at 1 and stop at 9. The tester output shall start at a low level and shall have a slow rate of increase in the bottom range of the scale. The rate of increase shall increase toward the middle and upper ranges so as to provide a more accurate diagnosis while minimizing the risk of jolting the patient.

Batteries. The dental pulp tester shall be powered from four, commercially available, size AA, 1.5 volt alkaline batteries. The dental pulp tester shall have a light on the module to indicate when the batteries are low. The dental pulp tester shall be furnished with new batteries. The batteries shall fit securely in the battery holder so as to prevent the batteries from falling out.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Ground lead and lip clip. The tester shall be suitable for use with rubber gloved hands. The tester shall be furnished with a grounding lead having a lip clip on one end and a clip suitable for attachment to the probe handle on the other hand. The lip clip shall connect to the patient's lip. The clips shall be stainless steel. The grounding lead shall be an insulated covered copper wire. The lead shall safely permit current to flow through the tooth, the lip clip and then via the lead to the probe handle. The lip clip shall be removable for autoclaving. The tester shall be furnished with a 1/2 fluid ounce bottle of topical fluoride gel.

Tolerances. Tolerances on dimensions shall be within plus or minus 10 percent of specified value.

Technical manuals. User/Operator, user maintenance and service manuals along with a parts price list and wiring schematics shall accompany each tester. A clear warning that the dental pulp tester shall not be used on patients with pacemakers and any other precautions shall also be included. The unit shall be safe for use on the patient.

Workmanship. The dental pulp tester shall be free from defects which detract from its appearance or may impair its serviceability.

Unit. Each (EA). One tester, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection, purposes shall be one dental pulp tester. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5 (percent defective) and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0 (percent defective).

Examination. The pulp tester shall be examined to determine compliance with the requirements of this specification. Examination shall include, but shall not be limited to the following defects:

Defects

Grounding lead with clip missing.
Probe tips do not fit probe securely.
Probe tips missing or not as specified.
Probe tip light doesn't flash upon reaching full scale.
Topical flouride gel missing.
Tester doesn't turn on when tooth contact is made.
Tester doesn't turn off following the specified contact delay.
Display doesn't automatically reset to zero after the specified contact delay.
Housing not as specified.
Cord length not as specified.
Batteries not included.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Tests. Test shall be conducted to determine compliance with specifications requirements. Where feasible, the same sample shall be used for the determination of two or more characteristics.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pounds units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-125-6618	TESTER, PULP, CANAL, DENTAL, Battery Operated, Digital Display

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

CUSTODIANS:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No.: 6520-2267
Location: NCR Disk 1

49134

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in writing comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-BSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51994A	TESTER, PULP, CANAL

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)* _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.