A-A-51996
4 June 1987

COMMERCIAL ITEM DESCRIPTION

ESOPHAGOSCOPE (Fiber Optic Holinger)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a rigid, Holinger, fiber optic Esophagoscope suitable for performing diagnostic esophageal procedures.

Salient characteristics:

Shall be a rigid, metal, fiber optic esophagoscope, Holinger style, 7 mm by 45 cm long, with an aspiration channel, light carrier and window plug, suitable for performing diagnostic esophageal procedures. The metal parts of the esophagoscope shall be fabricated of a suitable corrosion-resisting steel in accordance with ASTM F899 and shall be of the proper hardness and passivation.

The distal end of the esophagoscope shall have a smooth, blemish-free tip to prevent damage to the tissue. The esophagoscope shall have an aspiration channel permanently attached to the posterior-exterior portion of the instrument.

The channel shall have a smooth, distal tip, shortened flare, and a 15 mm proximal end with a finger hole plate for suction control. Design of the proximal end shall be in accordance with Figure 3 of American National Standard ANSI Z79.2.

The oval fiber optic light carrier shall allow almost full lumen for use of instrumentation and viewing. The light carrier shall have a fiber light bundle compatible with 7 mm by 45 cm long esophagoscope. The carrier shall allow for constant, cool illumination during use and shall be capable of repeated sterilization by ethylene oxide or cold sterilization solutions without damage to the light bundle.

The window plug shall fit a 15 mm proximal extension and shall provide a closed system for inflation during esophagoscopy.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

The esophagoscope shall be compatible with a high refractory optical glass cable sheathed in a soft, pliable silicone elastomer. Cable ends shall be protected by stainless steel ferrules and have "Pilling Style" adapters. The esophagoscope shall also be compatible with a portable, dual, illumination light source with two multiple port faceplates for use with fiber optic equipment/instrumentation, 120/240 Volt, 50/60 Hz.

Corrosion resistance. The esphogoscope corrosion-resisting steel parts shall successfully pass the copper sulfate and boil tests specified in Federal Specification GG-I-526.

Marking. In addition to marking requirements stated in GG-I-526, the esophagoscope of foreign manufacture shall be marked stainless.

Workmanship. The esophagoscope shall be free from defects which detract from its appearance or impair its serviceability.

PRODUCT QUALITY LEVEL.

Visual and dimensional examination. Shall be examined to verify that the design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One esophagoscope, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for Tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one esophagoscope. Inspection level shall be S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-204-2672	ESOPHAGOSCOPE, Holinger, Fiber Optic, Rigid, 7 mm by 45 cm

The esophagoscope shall be compatable with NSN 6515-01-153-9649 Light, Endoscopic Instrument Fiber Optic 120/240 Volt, 50/60 HZ, AC, Portable and NSN 6515-01-139-8769, Cable, Illuminator, Fiber Optic, 8 Feet Long.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES

VA-OSS
PHS
FDA-MPQAS

PGC 58417

Project No. 6515-3476

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use ubmitting comments and suggestions for improvements All users of military standardization documents are invited to provi suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, ar mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, w too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements

(Fold along this line)

Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATE

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN DPSC-RSTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-51996	ESOPHAGOSCOPE (Fiber Optic Holinger)

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b ADDRESS *(Street City State ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*

5 PROBLEM AREAS

a Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Incl Code)* – Optional
c MAILING ADDRESS *(Street City State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.