A-A-51997
22 June 1987

COMMERCIAL ITEM DESCRIPTION

MANDREL, DENTAL HANDPIECE, No. 303

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a No. 303 dental handpiece mandrel for holding abrasive wheels and disks.

Classifications.

Class AL. Latch angle handpiece
Class S. Straight handpiece

Salient characteristics:

The mandrel shall be a No. 303 regular head size mandrel suitable for securely holding abrasive wheels and disks in dental polishing and grinding procedures.

The mandrel shall be fabricated from Class 3, austenitic, or Class 4, martensitic, stainless steel in conformance with Federal Specification GG-I-526. Martensitic and austenitic steel mandrels shall show no corrosion, when subjected to the boil test of Federal Specification GG-I-526. Martensitic steel mandrel shall also pass the copper sulfate test of Federal Specification GG-I-526. Mandrel shall have a smooth, bright polished finish, without defects such as cracks or burs.


The mandrel shall consist of a one-piece threaded mandrel body, a corrosion resisting washer, and a matching threaded screw with a slotted head. The mandrel shall be furnished in the applicable Class, as specified. Class AL mandrel shall be suitable for use in latch type angle dental handpieces. Class S mandrel shall be suitable for use in straight dental handpieces and chucks of dental laboratory lathes. Mandrel shall be in accordance with Figure 1.

The total indicated runout of the mandrel shall be no greater than 0.003 inch, for both Class AL and S mandrels.

AMSC N/A FSC 6520


DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.



CLASS S - STRAIGHT HANDPIECE
NSN 6520-00-926-8846

NOTE.
SAME SCREW AND WASHER USED FOR BOTH ANGLE AND STRAIGHT HANDPIECES.



CLASS AL - LATCH ANGLE HANDPIECE
NSN 6520-00-926-8845

DIMENSIONS ARE IN INCHES

FIGURE 1. Mandrel Dental Handpiece, No. 303

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Workmanship. The mandrel shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing six mandrels, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Inspection shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one mandrel. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6520-00-926-8845	MANDREL, DENTAL HANDPIECE, Latch Angle Handpiece, Screw Type, 6s: No. 303
6520-00-926-8846	MANDREL, DENTAL HANDPIECE, Straight Handpiece, Screw Type, 6s: No. 303

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:04Z
Check the source to verify that this is the currentversion before use.

Ordering data. Class of mandrel, intermediate/exterior package quantities, labeling, and marking must be specified in the contract an/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

05384

Project No. 6520-1997

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

c. Reason/Rationale for Recommendation:

8. REMARKS

7a. NAME OF SUBMITTER (Last, First, MI) – Optional	b. WORK TELEPHONE NUMBER (Include Area Code) – Optional
c. MAILING ADDRESS (Street, City, State, ZIP Code) – Optional	8. DATE OF SUBMISSION (YYMMDD)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51997	MANDREL, DENTAL HANDPIECE, #303

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION (Mark one)
	☐ VENDOR
b. ADDRESS (Street, City, State, ZIP Code)	☐ USER
	☐ MANUFACTURER
	☐ OTHER (Specify)

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:04Z
Check the source to verify that this is the current version before use.