COMMERCIAL ITEM DESCRIPTION

ALTIMETER/BAROMETER, POCKET SIZE, WITH CASE

The General Services Administration has authorized the use of this commercial item description.

This description covers pocket-size altimeter/barometers that measure both barometric pressure and altitude. The altimeter/barometer is intended for use in general engineering and surveying.

Salient characteristics:

Design: The altimeter/barometer shall be of the aneroid type, pocket size and shall be sturdy, light-weight, compact, adaptable to rapid work, and entirely suitable for severe use. They shall be designed and constructed to withstand extended periods of transportation by hand, pack, air, land, or water. The weight shall not exceed four ounces (excluding carrying case). The pocket size altimeter/barometer shall be temperature compensated to a minimum of -4 °F (-20 °C) to +86 °F (+30 °C), and shall be no more than 3 X 3 X 1 (7.62 X 7.62 X 2.54 cm) inch in size. The altitude range shall be from 0 (sea level) feet to 15,000 feet (4572 meters), and be accurate to within ±30 feet (9.14 meters), with a minimum graduation of the dial to 20 feet (6.096 meters), and be estimated to 10 feet (3.048 meters). The barometer dial shall be 31 to 22.5 inches (78.74 to 57.15 cm) of mercury. The altimeter/barometer shall be provided with a carrying case, and an operator's manual explaining the use of the altimeter/barometer. Provisions for zeroing the altimeter/barometer should be provided.

Material: The altimeter/barometer shall be fabricated of hardened plastic.

Regulatory requirements. The offeror/contractor is encouraged to use recovered materials in accordance with Public Law 94-580 to the maximum extent practicable.

Packaging and marking. Military packaging shall be specified when shipments are to overseas Department of Defense (DoD) consignees. Commercial packaging may be specified for shipments to DoD consignees within the Continental United States and to non-DoD consignees worldwide. When a supplier's normal packaging will satisfy the requirements for military packaging, it may be used.

Military packaging. Each altimeter/barometer with carrying case shall be preserved and packed to protect against physical damage and deterioration in a logistic environment characterized by the following:

1. Multiple handling during transportation and intransit storage.

AMSC N/A FSC 6675

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

2. Shock, vibration, and static loading forces during worldwide shipment by all transport modes.
3. Uncontrolled warehouse storage for extended periods.
4. Weather exposure during shipment and intransit transfers.
5. Stacking loads during shipment and storage.

Quantities of altimeter/barometers destined to a single addressee may be unitized into one or more unit loads.

Commercial. Unless otherwise specified in the contract or order, each altimeter/barometer shall be packaged in accordance with the supplier's normal method for retail distribution. Altimeter/barometers shipped to a single destination may be unitized to permit handling as a single unit. Shipping containers shall comply with Uniform Freight Classification or National Freight Classification rules.

Marking. Unit, intermediate (when provided), and exterior packs shall be marked with the item nomenclature, stock or part number, quality per pack and contract number as a minimum; additional marking shall be as specified in the contract or order. Exterior packs shall have in addition the complete consignee's address. Markings shall be applied in a legible manner, in a contrasting color, of an appropriate size, and of waterproof materials.

Quality assurance. The altimeter/barometer shall be examined for the following defects. Presence of one or more defects shall be cause for rejection.

1. Design not as specified.
2. Altitude range, barometer scale and accuracy not as specified.
3. Dimensions not as specified.
4. Weight not as specified.
5. Dial and graduations not as specified.
6. Carrying case and operator's manual not provided.
7. Provisions for zeroing the altimater/barometer not provided.
8. Material not as specified.
9. Packaging and marking not as specified.

Material safety data sheet. A Material Safety Data sheet shall be prepared in accordance with FED-STD-313. One copy shall be submitted to the contracting officer. In addition, a copy shall be provided to the addresses in FED-STD-313, appendix B.

Certification of commercial item. The contractor shall certify that the product offered to the Government meets the salient characteristics of this description and that the product conforms to the producer's own drawings, specifications, standards, and quality assurance practices and is the same product as the manufacturer's current product offered by the contractor in the commercial marketplace for at least one year preceding the solicitation. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.

AS APPLICABLE.

Small business. A small business that cannot meet the certification-of-commercial-item requirement of this document, may make an offer under this solicitation; and, if such small business is the apparent successful offeror, the Government may require the small business to submit a sample altimeter/barometer. Any such altimeter/barometer may be evaluated for this solicitation if time to contract award permits. Such evaluation would be to determine conformance to all requirements of this commercial item description and to determine equivalency to altimeter/barometer available in the commercial marketplace in the areas of durability, human factors engineering, and safety. It should be noted that samples are not required unless specifically requested by the contracting officer. Note. The Government may elect not to evaluate a sample from an apparent low bidder for award for this solicitation if a determination of urgency is made by the contracting officer. The sample would then be evaluated prior to future solicitation for comparable items.

Notes:

1. When this commercial item description is used for procurement, the commercial item certification clause must appear in the solicitation.
2. Contracting officers will identify those activities requiring copies of completed Material Safety Data Sheets prepared in accordance with FED-STD-313. The pertinent government mailing addresses for submission of data are listed in appendix B of FED-STD-313.

MILITARY INTEREST:

Custodians
Army - ME
Navy - YD
Air Force - 99

Review Activities
Air Force - 84
DLA - GS

User Activity
Navy - MC

CIVIL AGENCY COORDINATING ACTIVITY:

GSA-FGC

PREPARING ACTIVITY:
Army - ME

Project 6675-0804



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:04Z
Check the source to verify that this is the current version before use.