P-L-790
DECEMBER 18, 1956
SUPERSEDING
Int. Fed. Spec.
P-L-00790(GSA-FSS)
April 19, 1956, and
Fed. Spec. P-C-582
October 8, 1945

FEDERAL SPECIFICATION

LUBRICANT, SILICATE CEMENT, DENTAL

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 *Scope.* — This specification covers a dental lubricant prepared to protect freshly inserted silicate restorations against moisture, and to lubricate strips and disks during finishing.

1.2 Classification.—This specification covers two compositions of lubricant.

2. APPLICABLE SPECIFICATIONS, STANDARDS, AND OTHER PUBLICATIONS

2.1 Specifications and standards.—The following specifications and standard, of the issue in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

PPP-B-566 — Boxes, Folding, Paperboard.

PPP-B-601—Boxes, Wood, Cleated-Plywood.

PPP-B-621—Boxes, Wood, Nailed and Lock-Corner.

PPP-B-676 — Boxes, Set-Up, Paperboard.

PPP-M-186 — Medicinal Products and Clinical Laboratory Reagents, General Specification for Containers (Packaging and Packing).

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

MIL-B-4229—Boxes; Paperboard, Metal-Stayed.

MIL-L-10547 — Liners, Case, Waterproof.

Military Standard:

MIL-STD-129—Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by the contractor in connection with specific procurement functions, should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:18Z
Check the source to verify that this is the current version before use.

2.2 Other publications. — The following documents forms a part of this specification. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply:

U. S. Pharmacopoeial Convention Publication:

Pharmacopoeia of the United States.

(Application for copies should be addressed to the Mack Printing Company, Easton, Pa.)

3. REQUIREMENTS

3.1 Material.—The lubricating compound shall have one of the following percentage compositions by weight as specified in the invitation for bids:

Composition 1:	*Percent*
White Petrolatum, USP	40
Hydrogenated Vegetable Oil	50
Theobroma Oil, USP	5
White Wax USP	5

Composition 2:	*Percent*
Paraffin	8.58
Paraffin Oil	36.4
Theobroma Oil, USP	36.4
Japan Wax	18.2
Boric Acid	0.2
Sodium Benzoate	0.1
Vanillin	0.02
Alcohol	0.1

3.2 Consistency.

3.2.1 The compound shall be of such consistency that when applied to the teeth in the usual manner it will not liquefy and run on the teeth.

3.2.2 The compound shall be of such consistency at 60°F. (15.6°C.) that it may be squeezed from the metal tube without excessive pressure.

3.2.3 The compound shall be of such consistency at 95°F. (35°C.) that it will not flow from the metal tube in a liquid state.

3.3 The material shall not separate nor become rancid.

3.4 The compound shall be contained in collapsible metal tubes as specified in Federal Specification PPP–M–186, for group B, class 1, closure A. Each tube shall contain a minimum of 1 ounce net weight.

3.5 Workmanship. — The workmanship shall be first class. The ingredients shall be so incorporated as to insure a uniformly smooth homogeneous finished product.

4. SAMPLING, INSPECTION AND TEST PROCEDURES

4.1 Sampling.—Unless otherwise specified, at least one tube of lubricating compound shall be taken at random from each delivery for inspection and tests.

4.2 Inspection.—The samples shall be inspected as necessary to determine compliance with this specification.

4.3 Tests. — Samples shall be tested as necessary to determine compliance with the applicable requirements of this specification.

5. PREPARATION FOR DELIVERY

(For definitions and levels, see 6.1.)

5.1 Packaging.

5.1.1 *Unit of issue.*—One (1) tube as specified, constitutes one unit.

5.1.2 *Quantities.* — Packaging quantities shall be as specified in table I.

TABLE I.—*Packaging quantities*

Unit pack	Intermediate pack	Shipping container
1 unit	12 units	288 units

5.1.2.1 *Packing variation permitted.* — If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:18Z
Check the source to verify that this is the current version before use.

of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

5.1.3 *Level A.*

5.1.3.1 *Unit package.*—Each tube shall be packaged in a carton of appropriate size constructed in accordance with Federal Specification PPP-B-566. Closure shall be effected in such manner as to prevent spillage of contents under normal handling.

5.1.3.2 *Intermediate package.*—Intermediate package shall be a carton of appropriate size constructed in accordance with Federal Specification PPP-B-566, PPP-B-676, or Military Specification MIL-B-4229. Closure shall be effected in such manner as to prevent spillage of contents under normal handling.

5.1.4 *Level C.*—Each unit shall be packaged in standard commercial containers of the size and kind commonly used for the purpose, which will afford the degree of protection required for shipment and use of the product for its intended purpose within the continental limits of the United States of America.

5.2 Packing.

5.2.1 *Level A.*

5.2.1.1 *Exterior container.*—Exterior container shall be designed for type 3 load and constructed in accordance with Federal Specification PPP-B-621.

5.2.1.2 *Case liner.* — Each box shall be lined with a waterproof case liner conforming to Military Specification MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto.

5.2.1.3 *Strapping.*—Strapping (when required) shall be in accordance with applicable box specification.

Note.—Strapping will not be required for shipment forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 *Level B.*

5.2.2.1 *Exterior container.*—Exterior container shall be designed for type 3 load, and constructed in accordance with Federal Specification PPP-B-621.

5.2.3 *Level C.*—Unless otherwise specified, the subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers at the lowest rate to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 *Tube.*—Each tube shall be marked with stock number, item identification, unit, quantity, lot (control) number, and name of contractor.

5.3.2 *Unit package.* — Each unit package shall be marked with stock number, item identification, unit, quantity, lot (control) number, office identifying number, and name of contractor.

5.3.3 *Intermediate packages and shipping container.*—Each intermediate package and each shipping container shall be marked in accordance with Military Standard MIL-STD-129. Service color marking and symbol are not required on containers for level A packs when addressed to domestic establishments for the Medical Services of the Armed Forces. Lot (control) number shall be shown.

6. NOTES

6.1 Federal Standard 102 shall be referred to for definitions and applications of the various levels of packaging protection for supplies and equipment.

6.2 Purchasers should exercise any desired options offered herein and should specify composition required. (See 3.1, 4.1, 5.1 and 5.2).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:18Z
Check the source to verify that this is the current version before use.

6.3 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.4 Federal specifications do not include all types, classes, grades, sizes, etc., of the commodities indicated by the title of the specifications, or which are commercially available, but are intended to cover the types, etc., which are suitable for Federal Government requirements.

6.5 Transportation description.—Transportation description applicable to this item is:

Cocoa butter (cocoa grease).

Carload minimum weight 36,000 pounds.

Truckload minimum weight 24,000 pounds.

6.6 This specification covers the following item appearing in the Armed Services Medical Stock List:

6520–529–2150—Lubricant, Silicate, Cement, Dental 1 oz.

Patent notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise, as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

MILITARY INTERESTS:

Army—M
Navy—Md
Air Force.

Copies of this specification may be purchased for 10 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:18Z
Check the source to verify that this is the current version before use.