FEDERAL SPECIFICATION

BANDAGES, SELF-ADHERENT

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE AND CLASSIFICATION

1.1 **Scope.**—This specification covers self-adherent cotton bandages.

1.2 **Type and sizes.**—Self-adherent bandages shall be of one type and of the following sizes, as specified in the invitation for bids:

Size 1.—2 inches by 5 yards
Size 2.—3 inches by 5 yards
Size 3.—as specified in the invitation for bids.

2. APPLICABLE SPECIFICATIONS AND OTHER PUBLICATIONS

2.1 **Specification.**—The following Federal Specification, of the issue in effect on date of invitation for bids, forms a part of this specification:

CCC–G–101—Gauze; Plain.

(Copies of Federal specifications and the Federal Specifications Index may be obtained upon application, accompanied by money order, coupon, or cash, to the Superintendent of Documents, Government Printing Office, Washington 25, D. C., or to the GSA Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, and Seattle. Prices may be obtained from the Superintendent of Documents or from the GSA Regional Offices.)

2.2 **Other publication.**—The following publication, of the issue in effect on date of invitation for bids, forms a part of this specification:

United States Pharmacopoeial Convention Publication
United States Pharmacopoeia (latest revision).

(Copies of United States Pharmacopoeial Convention publication may be obtained upon application to the Mack Publishing Company, Easton, Pa.)

2.3 Specifications and standard applicable only to individual departments are listed in section 7.

3. REQUIREMENTS

3.1 **Material.**

3.1.1 *Gauze.*—The gauze shall be in accordance with the requirements of Federal Specification CCC–G–101 for type II plain gauze. In addition, the breaking strength shall be not less than 14 pounds per inch of width.

3.1.2 *Adhesive mass.*—The adhesive mass shall be nontoxic and of such composition as to meet the requirements for adhesion described in 3.3.

3.2 **Fabrication.**—The bandage shall consist of gauze (see 3.1.1) impregnated with an adhesive mass. The bandage shall contain sufficient adhesive to enable the bandage to meet the requirements for adhesion in 3.3, but the weight of the impregnated bandage shall not exceed 90 gm. per square yard.

3.3 **Adhesion.**—Bandage shall adhere to itself and shall not stick to the skin or hair. When tested for adhesion in the manner described in 4.3.1, the bandage test strip shall pull a minimum of 6.5 pounds before the folded layers separate.

3.4 **Sterile wrapping.**—Each individual bandage shall be securely wrapped in suitable paper and sealed in a manner to maintain sterility of the bandage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:56Z
Check the source to verify that this is the current version before use.

3.5 **Sterilization.**—The bandage shall be sterile when tested in accordance with 4.3.2. The wrapped bandage shall be sterilized by high pressure steam and shall be dried after sterilization. Chemical sterilization is not permissible.

3.6 **Certificate of sterilization.**—When specified, the actual accomplishment of sterilization shall be certified in writing by sworn affidavit as having been accomplished by the manufacturer; such certificate shall accompany the test sample for the lot of which the sample is representative.

3.7 **Workmanship.**—Workmanship shall be first class throughout. Bandages shall be free from defects which detract from their appearance or which may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 **Sampling.**

4.1.1 *Bandage.*—Unless otherwise specified, at least 1 package of each lot of 500 or fraction thereof shall be taken at random from each delivery.

4.1.2 *Gauze.*—Unless otherwise specified, the contractor shall forward under separate cover, a test sample of 2 square yards of nonimpregnated gauze with each delivery.

4.2 **Inspection.**—Samples shall be inspected for size, material, wrapping, workmanship, packaging, packing, and marking.

4.3 **Tests.**

4.3.1 *Adhesion.*—Test strips shall be 1 inch in width and of sufficient length to fulfill the requirements of this test. The central portion of a strip shall be folded back on itself for a distance of 1 inch and then folded forward again. A 1-pound weight shall be rolled, without additional pressure, the length of the fold 10 times. The free ends of the test strip (the portions not folded) shall be clamped within the jaws of a tensile strength machine and the strip pulled until the folded layers separate. Adhesion shall be calculated from an average of pulls on five strips.

4.3.2 *Sterility.*—The bandage shall meet the requirements of the Sterility Tests for Solids as described in the latest revision of the U. S. Pharmacopoeia.

4.3.3 *Gauze.*—The gauze shall be tested as described in Federal Specification CCC-G-101.

5. PREPARATION FOR DELIVERY

5.1 **Packaging.**—Unless otherwise specified, each bandage shall be packaged in accordance with accepted commercial practice.

5.2 **Packing.**—The subject commodity shall be prepared for shipment to permit acceptance by carrier for transportation at the lowest applicable rate, and to afford maximum protection from normal hazards of transportation.

5.3 **Marking.**

5.3.1 *Packages.*—Unless otherwise specified, packages shall be marked with the name and size of the commodity, the quantity contained therein, and the names of the manufacturer and contractor.

5.3.2 *Shipping containers.*—Unless otherwise specified, shipping containers shall be marked with the name and size of the commodity as defined by the contract or order under which the shipment is made, the quantity contained therein, the name of the contractor, the number of the contract or order, the name and address of the consignee, and the gross weight.

6. NOTES

6.1 **Ordering data.**—Purchasers should specify the size required, certificate of sterilization if required, and should exercise any

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:56Z
Check the source to verify that this is the current version before use.

desired options offered herein. (See 1.2, 3.6, 4.1, 5.1, 5.3, 6.2, and 7.2.6).

6.2 It is believed that this specification adequately describes the characteristics necessary to secure the desired material, and that normally no samples will be necessary prior to award to determine compliance with this specification. If, for any particular purpose, samples with bids are necessary, they should be specifically asked for in the invitation for bids, and the particular purpose to be served by the bid sample should be definitely stated, the specification to apply in all other respects.

6.3 This specification covers the following items appearing in the Armed Services Catalog of Medical Matériel:

2-012-300 Bandage, Self-Adherent, 2 Inches by 5 Yards, 12s.

2-012-305 Bandage, Self-Adherent, 3 Inches by 5 Yards, 12s.

6.4 Other sizes of self-adherent bandages commercially available are as follows:

½	inch by	7½	yards
¾	inch by	2¾	yards
¾	inch by	7½	yards
1	inch by	7½	yards
1	inch by	5	yards
1	inch by	6	yards
1	inch by	10	yards
1	inch by	15	yards
1½	inches by	10	yards
2	inches by	10	yards

6.5 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of the specification, or which are commercially available, but is intended to cover the type, etc., normally procured to meet Federal Government requirements.

6.6 **Transportation description.**

6.6.1 *Transportation description applicable to these items is:*

Bandages or dressings, dentists', physicians' or surgeons'.

In boxes.

Carload weight 30,000 pounds.

Truckload weight 30,000 pounds.

Patent notice.—When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

7. DEPARTMENTAL REQUIREMENTS

7.1 The following specifications and standard, of the issues in effect on date of invitation for bids, and special requirements form a part of this specification for purchases made under this specification by the respective departments.

7.2 **Army, Navy, and Air Force.**

7.2.1 *Applicable specifications and standard.*

Federal Specifications:

NN-B-601—Boxes; Wood-Cleated-Plywood, for Domestic Shipment.

NN-B-621—Boxes; Wood, Nailed and Lock-Corner.

NN-B-631—Boxes; Wood, Wirebound (for Domestic Shipment).

Military Specifications:

JAN-P-105—Packaging and Packing for Overseas Shipment—Boxes, Wood, Cleated, Plywood.

JAN-P-106—Packaging and Packing for Overseas Shipment—Boxes, Wood, Nailed.

MIL-B-107—Boxes, Wood, Wirebound (Overseas Type).

JAN-P-120—Packaging and Packing for Overseas Shipment—Cartons, Folding, Paperboard.

JAN-P-125—Packaging and Packing for Overseas Shipment—Barrier-Materials, Waterproof, Flexible.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:56Z
Check the source to verify that this is the current version before use.

JAN-P-133—Packaging and Packing for Overseas Shipment—Boxes, Set-Up, Paperboard.
MIL-A-140—Adhesive, Water-Resistant, Waterproof Barrier-Material.

Navy Department Specification:
General Specifications for Inspection of Material.[1]

Military Standard:
MIL-STD-129—Marking of Shipments.

(Copies of specifications, standards, and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring agency or as directed by the contracting officer.)

7.2.2 *Packaging.*—Twelve bandages of the size specified constitute one unit. Each bandage shall be wrapped as specified in 3.4 and shall be individually packaged in a carton of suitable size and design. Twelve individual cartons (1 unit) shall be further packaged in a carton of appropriate size conforming to the requirements of Military Specification JAN-P-120 or JAN-P-138. Each carton shall be adequately secured to prevent accidental opening.

7.2.3 *Packing.*

7.2.3.1 *Domestic.*—Seventy-two unit cartons of size 1 bandage or forty-eight unit cartons of size 2 bandages shall be packed in a shipping container conforming to the requirements of Federal Specification NN-B-601, NN-B-621, or NN-B-631.

7.2.3.2 *Overseas.*—Seventy-two unit cartons of size 1 bandage or forty-eight unit cartons of size 2 bandage shall be packed in a shipping container designed for type 2 load, conforming to the requirements of Military Specification JAN-P-105, JAN-P-106, or MIL-B-107. Closure and strapping shall conform to the requirements of the appendix to applicable box specification. Each shipping container shall be lined with a sealed waterproof bag, or its equivalent, made from material conforming to the requirements of Military Specification JAN-P-125 for case liners; seams and closures shall be completely sealed by a continuous seam at least ¾ inch wide of water-resistant adhesive conforming to the requirements of Military Specification MIL-A-140.

7.2.3.3 *Packing variation.*—If the number of units in any shipment is not the exact multiple of the number of units specified to be contained in each shipping container, the units not constituting the nearest such multiple shall be packed in a shipping container of suitable size and design and so constructed as to insure safe delivery to specified destination.

7.2.4 *Marking.*

7.2.4.1 *Interior packages.*—Each individual carton shall be commercially identified. Each unit carton shall be marked with stock number, nomenclature, unit, quantity, office identifying number, date manufactured, and name of contractor.

7.2.4.2 *Shipping container.*—Each shipping container shall be marked in accordance with the requirements of Military Standard MIL-STD-129.

7.2.5 *Inspection procedures (Navy only).*—The general inspection procedures shall be in accordance with Navy Department General Specifications for Inspection of Material.

7.2.6 *Ordering data.*—Procurement documents should specify whether the subject commodity is to be packed for domestic or overseas shipment. (See 7.2.3.)

NAVY INTEREST: Md

[1] Applicable to Navy purchase only.

This specification is a part of Section IV, Part 5, of the Federal Standard Stock Catalog.

For sale by the Superintendent of Documents, Washington 25, D. C. Price 5 cents.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:56Z
Check the source to verify that this is the current version before use.