METRIC

A-A-54612
20 March 1992

COMMERCIAL ITEM DESCRIPTION

TEST KIT, HUMAN CHORIONIC GONADOTROPIN DETERMINATION

The General Services Administration has authorized the use of this Commercial Item Description as a replacement for Military Specification MIL-T-37805 which is cancelled.

Salient characteristics.

Shall be a suitable in vitro qualitative latex inhibition test, to detect Human Chorionic Gonadotropin in urine.

Each kit shall contain the following components:

(a) One hundred disposable slides each containing two dried reagents; latex coated with human Chorionic Gonadotropin and Anti-human Chorionic Gonadotropic Serum (Rabbit) dried as individual spots on the disposable slide.

(b) One vial of freeze-dried positive urine control and one vial of freeze-dried negative urine control.

(c) One hundred dispenser/stirrers capable of dispensing 0.03 ml per drop.

(d) One dropper vial for water.

AMSC/NA FSC 6550

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Sensitivity test. Shall be capable of detecting between 1 to 2 I.U. Human Chorionic Gonadotropin (HCG) per ml when tested as specified below.

The following method shall be employed in the detection of sensitivity of the pregnancy kit. Prepare, using a 0.9 percent NaCl/0.1 percent Bovine Serum Albumin solvent, dilutions of HCG to contain 0, 500, 1000, 1500, 2000, 2500 and 3000 I.U. per liter. Agglutination must be readily discernible with the standards made to contain 0 and 500 I.U. HCG per liter. Agglutination must be present with 1000 I.U. HCG per liter and Agglutination may be present with 1500 I.U. HCG per liter. The standard containing 2000 I.U. HCG per liter will inhibit the agglutination so that at the termination of the two minute rocking period only a very slight agglutination (if any) is visible. The standards containing greater than 2000 I.U. HCG per liter must show macroscopic inhibition of agglutination.

Reaction test. Shall be capable of detecting a positive response (no agglutination) and negative response (agglutination) when tested as specified below.

The test kit shall be tested against ten (10) known positive and ten (10) known negative pregnant female urines (Controls) using the method specified in the labeling. Positive responses (no agglutination) shall result when the positive control is used and negative responses (macro-agglutination, clumping) shall result when the negative control is used. A negative test shall be visible within two minutes and a positive test shall occur at the end of two minnutes.

Workmanship. The product shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One test kit, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered material. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to the degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Expiration dating required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6550-01-037-4877	TEST KIT, HUMAN CHORIONIC GONADOTROPIN DETERMINATION, 100 tests.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

A-A-54612

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

PGC 43315

Project No. 6550-0093

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-54612	20 March 1992

3. DOCUMENT TITLE

TEST KIT, HUMAN CHORIONIC GONADOTROPIN DETERMINATION

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code)	
Defense Personnel Support Center ATTN: MSE	1) Commercial (215)737-2870	(2) AUTOVON 444-2870
c. ADDRESS (Include Zip Code) 2800 S. 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.