METRIC

A-A-54620
21 August 1992

COMMERCIAL ITEM DESCRIPTION

ARTHROSCOPIC SYSTEM, SURGICAL, 4 mm

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers an arthroscopic surgical system suitable for use in athroscopic surgical procedures.

Salient characteristics:

The system shall be a 4 millimeter (mm) arthroscopic surgical system, suitable for use during joint visualization or during arthroscopic surgery procedures. The system shall consist of the following components:

Component	Quantity
Telescope	1
Pyromidal Trocar	1
Conical Obturator	1
Trocar Sleeve, Self-Locking with Stopcock	1
Cable, Fiberoptic	1

Fiberoptic Telescope. The system shall contain a 4 mm Fiberoptic Arthroscopic telescope with a 25° viewing angle, providing for a panoramic view without loss of contrast. The telescope shall be fabricated of a suitable corrosion resisting steel, properly hardened and passivated and in accordance with ASTM F899. The proximal eyepiece shall be of bakelite material and shall not crack during ethylene oxide or steam sterilization.

The fiber bundle's width shall be a minimum of 3.8 mm. Illumination and brightness through the lens shall produce true to life reproduction of color in both video and still photography. The telescope shall have an overall nominal length of 23.5 centimeter (cm) and a working length of 17.7 cm. Shall fit smoothly and securely lock into a 5 mm self-locking trocar sleeve.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

A fiberoptic light cord adapter shall be an integral part of the proximal portion of the telescope and shall be located immediately distal to the bakelite eyepiece.

The proximal end of the telescope shall be color coded to match other component parts of the arthroscopic system in use. Either the millimeter size or the company stock number shall be clearly imprinted somewhere on the proximal end of the telescope.

The telescope shall be approximately 23.5 cm and shall have a working length of approximately 17.7 cm.

Pyromidal trocar. The system shall contain a 4 mm corrosion-resisting steel trocar. The trocar shall fit smoothly into the 5 mm trocar sleeve with no gaping between the sleeve and the trocar at the distal end. When inserted into the 5 mm trocar sleeve the trocar tip shall protrude from the sleeve. The distal tip shall be beveled on two aspects to form a sharp trocar.

The trocar shall have an overall length of 21 cm and a working length of 18.6 cm. The trocar shall have an integral collar with lateral metal projections for a secure fit into the collar of the 5 mm self-locking arthroscopic trocar sleeve.

The extreme proximal end of the trocar shall be color coded to match other component parts of the arthoscopic system in use. Either the millimeter size or the company stock number shall be clearly imprinted on the proximal end of the trocar.

Conical obturator. The system shall contain a 4 mm corrosion-resisting steel obturator properly hardened and passivated. Shall be fabricated from stainless steel Type 303 or 304 in accordance with ASTM F899. When inserted into the 5 mm trocar sleeve the obturator shall fit smoothly without gaping between the sleeve and the obturator at the distal end. When inserted into the 5 mm trocar sleeve the conical end shall protrude. The distal tip shall be blunt and conical in shape.

The obturator shall have an overall length of 21 cm and a working length of 18.6 cm. The obturator shall have an integral collar with lateral metal projections for a secure fit into the collar of a 5 mm self locking arthroscopic trocar sleeve.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:36Z
Check the source to verify that this is the current version before use.

The extreme proximal end shall be color coded to match other component parts of the arthroscopic system in use. Either the millimeter size or the company stock number shall be clearly imprinted on the proximal end of the obturator.

Trocar sleeve, self locking with stopcock. The system shall contain a corrosion-resisting steel trocar fabricated from stainless Type 303 or 304. The sleeve diameter shall be 5 mm and be specifically designed for accommodating a 4 mm arthroscopic trocar and telescope.

The overall length of the trocar sleeve shall be a nominal 17.7 cm and the working length shall be 13.1 cm. The proximal end shall have a self locking collar which prevents the telescope from disengaging. The distal end shall be slightly angulated with smooth edges for decreasing the possibility of tissue damage. Several smooth edged holes shall be located on the lateral aspect of the distal tip. The trocar sleeve shall have a single integral luer-lock stopcock on the proximal portion of the sleeve shaft.

The self locking collar shall be color coded to match other component parts of the arthroscopic lens system. The millimeter size or the company stock number shall be clearly imprinted on the locking collar.

Fiberoptic cable. The system shall contain a fiberoptic cable having minimum length of 1.8 meters with a 4 mm bundle diameter. The cable ends shall be specifically designed for adapting into a 4 mm arthroscope and a fiberoptic light source without an adapter. Both ends shall have a strain relief sleeve for protection against sharp bending.

The fiber light cables shall be covered with medical grade silastic rubber tubing. The metal portions of the cable shall be fabricated from a corrosion resistant stainless steel Class 3, Type 302 or equal in accordance with ASTM F899.

Compatibility. All components of the arthroscope shall be compatible, one with the other.

Sterilization. All components shall be capable of withstanding repeated sterilization via steam at 275°F or ethylene oxide or long term 10 hour soaking in cold sterilization solution procedures.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Finish. The system shall be fabricated from materials which are inherently resistant to corrosion or which are treated to render them corrosion-resistant. Where painting or plating is used, appropriate cleaning and priming shall be performed prior to finishing. Unless otherwise specified, finish shall be equivalent or superior to that of comparable commercial equipment and shall be free of gaps, scratches, chips, and other imperfections.

Workmanship. The arthroscopic surgical system shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One arthroscopic surgical system, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. Unit of product for test purposes shall be one arthroscopic surgical system. Sampling shall be inspection level S-2 with an AQL of 1.0.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE 1: The following National Stock Number and its respective Item Identification is covered by this document:

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:36Z
Check the source to verify that this is the current version before use.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-206-5971	ARTHROSCOPIC SYSTEM, SURGICAL, 4 mm

NOTE 2: The following National Stock Numbers, and Item Identifications are applicable for this Commercial Item Description:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-171-6050	TELESCOPE, ARTHROSCOPIC, PANORAMIC, 25° Viewing Angle, 4 mm Outside Diameter
6515-01-166-3528	TROCAR, PYROMIDAL TIP, 1 mm, Outside Diameter
6515-01-166-3504	OBTURATOR, ARTHROSCOPE, 4 mm Diameter, Color Coded, Conical Tip
6515-01-173-2452	ARTHROSCOPIC, TROCAR SLEEVE, Self-Locking Collar with Stopcock, 5 mm Diameter
6515-01-139-8567	CABLE, Assembly, Fiberoptic, 4 mm, 6 feet (1.83 m) Long
6515-01-181-8711	LIGHT, Endoscope Instrument Fiber Optic, 120/240 Volt, 50/60 Hz, AC, Portable

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-4600
Location: c:\word\eric\arthrocd

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-54620	2 DOCUMENT DATE (YYMMDD) 21 August 1992

3 DOCUMENT TITLE
ARTHROSCOPIC SYSTEM, SURGICAL, 4 mm

4 NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible Attach extra sheets as needed)

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a [illegible]	b ORGANIZATION
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable) 7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial 215-737-2870 (2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 Previous editions are obsolete 198/29(

Source: https://assist.dla.mil - Downloaded: 2026-10-26T09:36Z
Check the source to verify that this is the current version before use.